

1265 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Time of Inspection: 04:50

Date of Inspection: 08/28/2019

Serial Number: 80-001265

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A4 Exp: 10/05/2019
0.000	0.048	0.076	0.196	0.081
0.000	0.049	0.077	0.196	0.080
0.000	0.049	0.076	0.196	0.081

Number of Simulators Used: 5 ⁴

ONE SIMULATOR FAILED DURING INSPECTION

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

08/28/2019

Date

1265
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Time of Inspection: 04:50

Date of Inspection: 08/28/2019

Serial Number: 80-001265

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22817080A4 Exp: 10/05/2019
0.000	0.048	0.076	0.196	0.081
0.000	0.049	0.077	0.196	0.080
0.000	0.049	0.076	0.196	0.081

Number of Simulators Used: 5 ⁴

ONE SIMULATOR FAILED DURING INSPECTION

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

08/28/2019
Date

1265
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001265

Time of Inspection: 02:57

Date of Inspection: 11/26/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) 201802B 201902B Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817080A4 Exp: 10/05/2019
0.000	0.048	0.079	0.198	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.048	0.079	0.198	0.081

Number of Simulators Used: 5

Remarks:

AMENDED LOT NUMBER FOR .05 DUE TO KEYSTROKE ERROR.
(201802B)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

11/26/2018
Date