

1266
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001266

Time of Inspection: 11:22

Date of Inspection: 05/22/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A2 Exp: 10/05/2020
0.000	0.047	0.077	0.187 / 0.195	0.078
0.000	0.048	0.078	0.190 / 0.195	0.078
0.000	0.048	0.079	0.190 / 0.195	0.078

Number of Simulators Used: 4

Remarks:

20: Control Outside Tolerance.

SUSPECTED AIR LEAK OR NEEDED ADDITIONAL TIME TO WARM AFTER SIMULATOR CHANGE.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

05/22/2019
Date

1266 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Time of Inspection: 10:06

Date of Inspection: 05/22/2019

Serial Number: 80-001266

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A2 Exp: 10/05/2020
0.000	0.048	0.078	0.205	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.079	0.203	0.078

Number of Simulators Used: 5

Remarks:

.20 SIMULATOR EXCEEDED .02 VARIANCE DUE TO UNKNOWN ISSUE WITH SIMULATOR. DUE TO EXCEEDING 34.2 C, REPEATED INSPECTION. NOTIFIED DR. KIRKLAND VIA PHONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ
Signature and Printed Name

05/22/2019
Date