

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED  
FDLE

FEB 11 2019

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 22:56

Date of Inspection: 01/31/2019

Serial Number: 80-001285  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A3 Exp: 10/05/2020
0.000	0.049	0.078	0.199	0.081
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.202	0.081

Number of Simulators Used: 5

Remarks:


RECEIVED

FEB 01 2019

Records Dept.  
Baker County Sheriffs Office

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name COREY S WILSON

01/31/2019  
 Date