

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO
Time of Inspection: 14:05

Date of Inspection: 09/30/2019

Serial Number: 80-001308
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG816701 Exp: 06/16/2020
0.000	0.046	0.075	0.195	0.081
0.000	0.046	0.076	0.197	0.081
0.000	0.046	0.076	0.197	0.081

Number of Simulators Used: 5

Remarks:

There were (2) inspections for the month of September 2019 due to the .08 solution was expired at the time of this inspection.

B. Miller
9-30-19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

B. Miller

BRIAN D MILLER

Signature and Printed Name

09/30/2019
Date