

Amended

1320 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKECHOBEE CO SO  
Time of Inspection: 14:43

Date of Inspection: 04/15/2019

Serial Number: 80-001320  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 507129 Exp: 08/04/2019
0.000	0.047	0.078	0.189 / 0.198	0.081
0.000	0.047	0.078	0.189 / 0.198	0.081
0.000	0.047	0.078	0.188 / 0.199	0.081

Number of Simulators Used: 4

Remarks:

20: Control Outside Tolerance.

The first .20 alcohol reference solution bottle I used had a leak and was not a full bottle.

I repeated the 0.20g/210L test using a full bottle of the .20 alcohol reference solution and those test results were within tolerance.

4/15/2019 (M)

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MARK W GARLAND

Signature and Printed Name

04/15/2019  
Date

1320  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE CO SO  
Time of Inspection: 14:43

Date of Inspection: 04/15/2019

Serial Number: 80-001320  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:507129 Exp: 08/04/2019
0.000	0.047	0.078	0.189 / 0.198	0.081
0.000	0.047	0.078	0.189 / 0.198	0.081
0.000	0.047	0.078	0.188 / 0.199	0.081

Number of Simulators Used: 4

Remarks:

20: Control Outside Tolerance.

The first .20 alcohol reference solution bottle I used had a leak and was not a full bottle. I repeated the 0.20g/210L test using a full bottle of the .20 alcohol reference solution and those test results were within tolerance.

4/15/2019 (M)

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Mark W. Garland*

MARK W GARLAND

Signature and Printed Name

04/15/2019  
Date