

- AMENDED -

1322  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE P.D.

Serial Number: 80-001322

Time of Inspection: 18:31

Date of Inspection: 04/16/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) <i>05/28/19</i> Lot#: <del>201820-201820</del> Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 840875 Exp: 09/19/2020
0.000	0.048	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.202	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RAUL E MARRERO

Signature and Printed Name

04/16/2019  
Date