

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CO. SO
Time of Inspection: 15:17

Date of Inspection: 02/27/2019

Serial Number: 80-001328
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:980558 Exp: 07/09/2021
0.000	0.049	0.071 / 0.079	0.198	0.080
0.000	0.049	0.077 / 0.080	0.200	0.080
0.000	0.049	0.079 / 0.080	0.201	0.079

Number of Simulators Used: 4 *Amended*


Remarks:

08: Control Outside Tolerance.

Solution utilized for .08 test was changed due to issue with simulator and allowed to sit at proper temperature for longer duration resulting in exceeding within compliance standards.
MA 3-28-19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

MARK A KANTORSKI

02/27/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CO. SO
Time of Inspection: 14:10

Date of Inspection: 02/27/2019

Serial Number: 80-001328
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.048	0.078	0.000 / 0.188	
0.000	0.049	0.078	0.000 / 0.197	
0.000	0.049	0.079	0.000 / 0.199	

Number of Simulators Used: 4 Amended


Remarks:

20: Control Outside Tolerance, Control Outside Tolerance.

Simulator utilized for 20 solution was leaking air and unable to hold a proper seal. Simulator was removed and test was completed with all new simulators. Simulator Broken was sent for repair/removal to Tech source M 3-28-19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

MARK A KANTORSKI

02/27/2019
Date