

1416
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Orange County Sheriff's Office

Serial Number: 80-001416

Time of Inspection: N/A

Date of Inspection: 09/11/2019

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted	N/A	N/A
Diagnostic Check (Pre-Inspection): OK	N/A	N/A
Alcohol Free Subject Test: 0.000	N/A	N/A
Mouth Alcohol Test: Slope Not Met	N/A	N/A
Interferent Detect Test: Interferent Detect	N/A	N/A
Diagnostic Check (Post-Inspection): OK	N/A	N/A

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.20g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: N/A Exp: N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A


Number of Simulators Used: N/A

Remarks: Not able to perform an Agency Inspection due to the instrument having a broken piece on the end of the Dry Gas Standard's supply hose. I notified Department Inspector Shayla Platt via telephone of this issue. Compliance to Agency Inspection requirement in Rule 11D-8 not determined.

(All of the information entered on this Agency Inspection Report was hand-typed by Agency Inspector K. Melville)

The above instrument complies (N/A) does not comply (N/A) with Chapter 11D-8, FAC. ok

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit ~~and that I performed this inspection~~ in accordance with the provisions of Chapter 11D-8, FAC.

ok


Kelly M. Melville

Signature and Printed Name

09/11/2019
Date