

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

3/4/19
JD

Agency: CLEARWATER PD

Time of Inspection: 20:27

Date of Inspection: 02/20/2019

Serial Number: 80-001740

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G <i>02/28/19</i> Exp: 03/20/2020 <i>02/22/2020</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.078	0.197	0.079

Number of Simulators Used: 5

Remarks:

ACETONE LOT2018-A MOUTH ALC SOL LOT2017-B

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Daniel Negersmith

DANIEL M NEGERSMITH

Signature and Printed Name

02/20/2019
Date