

# Florida Department of Law Enforcement Alcohol Testing Program

80-003347  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO

Time of Inspection: 14:53

Date of Inspection: 10/22/2019

Serial Number: 80-003347

Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201811C<br>Exp: 11/13/2020 | 0.08g/210L Test (g/210L)<br>Lot#:201808E<br>Exp: 08/22/2020 | 0.20g/210L Test (g/210L)<br>Lot#:201802G<br>Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:14219080A2<br>Exp: 08/05/2021 |
|----------------------------|---|---|---|---|
| RFI / 0.000                | 0.048   | 0.077   | 0.193   | 0.082   |
| / 0.000                    | 0.047   | 0.076   | 0.195   | 0.081   |
| / 0.000                    | 0.047   | 0.077   | 0.195   | 0.081   |


Number of Simulators Used: 4

**Remarks:**

00: RFI Detect. - Body scanner was running at this time. I wait till they were done.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 Signature and Printed Name BARRY NIXON

10/22/2019  
 Date

# Florida Department of Law Enforcement Alcohol Testing Program

20-003347  
BKH

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 16:39

Date of Inspection: 08/20/2019

Serial Number: 80-003347  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201811C<br>Exp: 11/13/2020 | 0.08g/210L Test (g/210L)<br>Lot#:201808E<br>Exp: 08/22/2020 | 0.20g/210L Test (g/210L)<br>Lot#:201802G<br>Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:02618080A3<br>Exp: 03/05/2020 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.047   | 0.077   | 0.196   | 0.081   |
| 0.000                      | 0.047   | 0.078   | 0.196   | 0.081   |
| 0.000                      | 0.047   | 0.078   | 0.197   | 0.080   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BARRY NIXON

Signature and Printed Name

08/20/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-003347  
BLK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 15:28

Date of Inspection: 08/20/2019

Serial Number: 80-003347  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            |     | No |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect |     | No |
| Diagnostic Check (Post-Inspection): OK      |     | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: _____<br>Exp: _____ | 0.08g/210L Test (g/210L)<br>Lot#: _____<br>Exp: _____ | 0.20g/210L Test (g/210L)<br>Lot#: _____<br>Exp: _____ | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: _____<br>Exp: _____ |
|----------------------------|---|---|---|--|
|                            |   |   |   |  |
|                            |   |   |   |  |
|                            |   |   |   |  |

Number of Simulators Used: 5

**Remarks:**

STOPPED TO RESET AND RESTART Non-compliance: SEE WRITTEN NOTES.  
 I did the Alcohol Free test after doing the mouth alcohol test on other instrument.  
 The test read 0.00 so I checked the injection valve and started the inspection  
 from the beginning

The above instrument complies (     ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name BARRY NIXON

08/20/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

*2/1/19*  
*JO*

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 05:28

Date of Inspection: 01/22/2019

Serial Number: 80-003347  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201802E<br>Exp: 02/06/2020 | 0.08g/210L Test (g/210L) <i>2.391g</i><br><del>201802E</del> <i>1/22/19</i><br>Lot#: <del>201802E</del><br>Exp: <del>02/06/2020</del><br><i>08/22/2020</i> | 0.20g/210L Test (g/210L)<br>Lot#: 201802G<br>Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 17817080A2<br>Exp: 08/05/2019 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.048  | 0.076  | 0.196  | 0.081  |
| 0.000                      | 0.048  | 0.076  | 0.196  | 0.082  |
| 0.000                      | 0.049  | 0.077  | 0.197  | 0.082  |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
BARRY NIXON  
Signature and Printed Name

01/22/2019  
Date