

# Florida Department of Law Enforcement Alcohol Testing Program

80-003945  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 08:05

Date of Inspection: 12/09/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2021
0.000	0.046	0.076	0.195	0.081
0.000	0.047	0.076	0.196	0.081
0.000	0.047	0.077	0.197	0.081

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ

Signature and Printed Name

12/09/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-003945  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 09:06

Date of Inspection: 11/18/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17919080A1 Exp: 08/05/2021
0.000	0.046	0.077	0.191	0.081
0.000	0.047	0.077	0.193	0.081
0.000	0.047	0.077	0.194	0.081

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ

Signature and Printed Name

11/18/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 08:13

Date of Inspection: 10/21/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2021
0.000	0.047	0.075	0.195	0.080
0.000	0.046	0.076	0.197	0.080
0.000	0.046	0.077	0.199	0.080

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ

Signature and Printed Name

10/21/2019

Date

RECEIVED

SEP 25 2019

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 09:36

Date of Inspection: 09/23/2019

Serial Number: 80-003945

Software: 8i00.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201806E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201.02G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1791908A1 Exp: 08/09/2020
0.000	0.045	0.075	0.195	0.081
0.000	0.045	0.077	0.197	0.081
0.000	0.047	0.077	0.199	0.081

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-9, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-9, FAC.

*Gordon L. Franz*

Signature and Printed Name

GORDON L. FRANZ

09/23/2019

Date

RECEIVED

# Florida Department of Law Enforcement

## Alcohol Testing Program

SEP 03 2019

FDLE

Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 10:04

Date of Inspection: 08/15/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2021
0.000	0.046	0.076	0.196	0.080
0.000	0.047	0.076	0.199	0.080
0.000	0.047	0.077	0.199	0.080

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ

Signature and Printed Name

08/15/2019

Date



RECEIVED

AUG 21 2019

# Florida Department of Law Enforcement

## Alcohol Testing Program

FDLE  
Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE LAND PD

Time of Inspection: 10:04

Date of Inspection: 08/15/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2011
0.000	0.046	0.076	0.196	0.080
0.000	0.047	0.076	0.199	0.080
0.000	0.047	0.077	0.199	0.080

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ  
Signature and Printed Name

08/15/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

7/25/19  
JO

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 10:03

Date of Inspection: 07/25/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17217030A3 Exp: 08/05/2019
0.000	0.046	0.076	0.194	0.079
0.000	0.047	0.077	0.196	0.079
0.000	0.047	0.077	0.196	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz

Signature and Printed Name

GORDON L. FRANZ

07/25/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

6/13/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 10:19

Date of Inspection: 06/13/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802C Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17817080A3 Exp: 08/05/2019
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.200	0.079
0.000	0.048	0.078	0.201	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

Signature and Printed Name

06/13/2019  
Date





Florida Department of  
Law Enforcement

## AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Lakeland PD

Instrument Serial Number: 80-003945

AGENCY INSPECTION DATA REVIEW	
Agency Inspector: Gordon L. Franz	Date of Inspection: 5/9/2019 Time of Inspection: 07:31
Agency Inspection Discrepancy: <input type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input checked="" type="checkbox"/> Erroneous Information <input type="checkbox"/> Procedural <input type="checkbox"/> Other _____	
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.	
<input checked="" type="checkbox"/> Lot Number <input checked="" type="checkbox"/> Expiration Date for <u>0.08g/ 210L</u> <input checked="" type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input checked="" type="checkbox"/> Incorrect <input type="checkbox"/> Expired.	
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input type="checkbox"/> REASON for repeating the following test(s); OR the <input type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input type="checkbox"/> 0.05 g/210L Test <input type="checkbox"/> 0.08 g/210L Test <input type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test	
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.	
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.	
<input type="checkbox"/> Other: _____	

OTHER ELECTRONIC DATA REVIEW	
<input type="checkbox"/> Login Records Date: _____	Comments: _____
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

CORRECTIVE ACTION
<input checked="" type="checkbox"/> Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by <u>5/30/2019</u> (Date).
<input type="checkbox"/> Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
<input type="checkbox"/> Upload the Agency Inspection(s).
<input type="checkbox"/> Remove the instrument from evidentiary use until otherwise directed by the Department.
<input type="checkbox"/> No action required
<input type="checkbox"/> Other: _____

  
Signature of Alcohol Testing Program Staff Member

5/9/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 07:31

Date of Inspection: 05/09/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2012	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.046	0.076	0.197	0.079
0.000	0.047	0.077	0.197	0.079
0.000	0.047	0.077	0.198	0.079

↑  
Wrong  
Date

Number of Simulators Used: 5

Remarks:



Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ

Signature and Printed Name

05/09/2019  
Date

## Graham, Thomas

---

**From:** Graham, Thomas  
**Sent:** Thursday, May 09, 2019 1:53 PM  
**To:** 'Franz, Gordon'  
**Subject:** RE: LPD's Intoxilyzer Inspections for May 2019  
**Attachments:** 80-003945 AI Review Form 5-9-2019.pdf

Gordon,

Please review the Agency Inspection for Intoxilyzer 8000 #80-003945 performed on 5/9/2019. Your 0.08 ARS Expiration date is incorrect.

Please see the attached file.

Thank you in advance.

TJ

---

**From:** Franz, Gordon [<mailto:Gordon.Franz@lakelandgov.net>]  
**Sent:** Thursday, May 09, 2019 12:33 PM  
**To:** Trulson, Tom; Ivette Miranda; Mumbauer, Chad; Suttle, Cory; Franz, Gordon; Walker, Steve; Addison, Brent; [eross@sao10.com](mailto:eross@sao10.com)  
**Cc:** Lehman, Hans; Graham, Thomas  
**Subject:** LPD's Intoxilyzer Inspections for May 2019

To whom it may concern,

Please see the attached files for the Lakeland Police Department's monthly Intoxilyzer Inspections for May 2019.

Please contact me with any questions or concerns.

Respectfully,

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497  
[Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)

## Graham, Thomas

---

**From:** Franz, Gordon <Gordon.Franz@lakelandgov.net>  
**Sent:** Thursday, May 09, 2019 2:44 PM  
**To:** Graham, Thomas  
**Subject:** Amended Form 40 for Intoxilyzer 80-003945  
**Attachments:** Amended Form 40.pdf

**Follow Up Flag:** FollowUp  
**Flag Status:** Flagged

TJ,

Attached you will find the amended Form 40 for Intoxilyzer 80-003945. I will watch my typing closer in the future! Thanks for all your assistance.

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497  
[Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)



---

**PUBLIC RECORDS NOTICE:**

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

5/9/19  
22

Agency: LAKELAND PD

Time of Inspection: 07:31

Date of Inspection: 05/09/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2012	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.046	0.076	0.197	0.079
0.000	0.047	0.077	0.197	0.079
0.000	0.047	0.077	0.198	0.079

Number of Simulators Used: 5

Remarks:

AMENDED:

THE CORRECT EXPIRATION DATE IS: 09/12/2019. THE EXPIRATION DATE OF 09/12/2012 WAS TYPED IN ERROR.

*Gordon L. Franz 5/9/2019*

Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ

Signature and Printed Name

05/09/2019  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

Copy

4/24/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 11:29

Date of Inspection: 04/24/2019

Serial Number: 80-003945  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.048	0.078	0.202	0.079
0.000	0.048	0.079	0.202	0.078
0.000	0.048	0.079	0.203	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ  
Signature and Printed Name

04/24/2019  
Date

# Florida Department of Law Enforcement

## Alcohol Testing Program

Copy

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

3/26/19  
JO

Agency: LAKELAND PD

Time of Inspection: 10:52

Date of Inspection: 03/25/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17817080A3 Exp: 08/01/2018
0.000	0.047	0.078	0.198	0.079
0.000	0.047	0.079	0.199	0.079
0.000	0.047	0.078	0.199	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz

GORDON L. FRANZ

Signature and Printed Name

03/25/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

2/20/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 12:56

Date of Inspection: 02/18/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17817080A3 Exp: 08/05/2019
0.000	0.047	0.077	0.198	0.078
0.000	0.047	0.078	0.200	0.079
0.000	0.048	0.078	0.200	0.079

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

Signature and Printed Name

GORDON L FRANZ

02/18/2019  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

1/24/19  
JO

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE LAND PD

Time of Inspection: 11:05

Date of Inspection: 01/14/2019

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17817080A3 Exp: 08/05/2019
0.000	0.048	0.078	0.196	0.079
0.000	0.050	0.078	0.197	0.080
0.000	0.050	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

Signature and Printed Name

GORDON L. FRANZ

01/14/2019  
Date