

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST PETERSBURG PD
Time of Inspection: 09:29

Date of Inspection: 12/23/2019

Serial Number: 80-005338
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2018 <i>09/22/2020</i>	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2019 <i>02/20/2020</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A3 Exp: 10/05/2020
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.080	0.200	0.080

Number of Simulators Used: 5

Remarks:
IN COMPLIANCE

Amended exp date for .08 and .20 Solution. Entered incorrectly

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael D Weiskopf

MICHAEL D WEISKOPF

Signature and Printed Name

12/23/2019
Date