

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD  
Time of Inspection: 14:37

Date of Inspection: 10/29/2019

005550  
Serial Number: 80-~~000000~~  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34917080A4 Exp: 02/05/2020
0.000	0.048	0.079	0.194	0.079
0.000	0.049	0.079	0.195	0.079
0.000	0.049	0.079	0.195	0.078

Number of Simulators Used: 5

Remarks:  
AI MYRTIL PRESENT

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

02/141

KEVIN MILLAN

Signature and Printed Name

10/29/2019  
Date