

5255
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT MYERS PD
Time of Inspection: 09:47

Date of Inspection: 02/15/2019

Serial Number: 80-005655
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28818080A1 Exp: 12/05/2020
0.000	0.050	0.080	0.200	0.078
0.000	0.051	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.078

Number of Simulators Used: 5

Remarks:

A F / M A: TOO MUCH MA.

952 TALKED TO Ms PLATT
FROM FDLE - I MAY HAVE HAD MA Residual.
She said to give another test - if
no again then send instrument out.
Dec 2/15/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 03-043
Signature and Printed Name

02/15/2019
Date