

Florida Department of Law Enforcement Alcohol Testing Program

5/3/19
JL

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 03:58

Date of Inspection: 05/02/2019

Serial Number: 80-005935
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.048	0.075 / 0.076	0.194	0.080
0.000	0.048	0.076 / 0.076	0.198	0.080
0.000	0.048	/ 0.076	0.198	0.080

Number of Simulators Used: 5

Remarks:

08: RFI Detect, D/S WALKED IN W/O TURNING OFF RADIO.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

DANIEL E LYONS

Signature and Printed Name

05/02/2019

Date

Graham, Thomas

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Friday, May 03, 2019 3:58 PM
To: Graham, Thomas
Subject: FW: Inotx inspections for May 2019
Attachments: 02 May 2019 Monthly Intox Inspec.pdf

80-005935
5/3/19
JD

From: Dan Lyons
Sent: Thursday, May 02, 2019 5:46 AM
To: 'ThomasGraham@fdle.state.fl.us' <ThomasGraham@fdle.state.fl.us>
Subject: FW: Inotx inspections for May 2019

From: Dan Lyons
Sent: Thursday, May 02, 2019 5:45 AM
To: 'McKnight, Jamie (JMcknight@sao9.org)' <JMcknight@sao9.org>; 'Atkinson, Karsona' <KarsonaAtkinson@flhsmv.gov>; 'SherryForcier@flhsmv.gov' <SherryForcier@flhsmv.gov>
Cc: George Maroney <gmar2@OSCEOLA.ORG>; John Haydel <jhay@OSCEOLA.ORG>
Subject: Inotx inspections for May 2019

ALCON,

Agency Inspection Report conducted 05/02/2019 on the following instruments;
80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-005935, 0.08g/210L Test, Received an RFI reading when an Deputy walked in without turning off his radio. He was reprimanded and educated. The instrument was re-tested. The re-test results are with in standards. The inspection was not Re-Started.

Instruments 80-000966 is still out at CMI for Annual Inspection and repairs.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Florida Department of Law Enforcement Alcohol Testing Program

4/8/19
D

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 17:14

Date of Inspection: 04/03/2019

Serial Number: 80-005935
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.047	0.075	0.192	0.081
0.000	0.048	0.076	0.194	0.080
0.000	0.048	0.076	0.195	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DANIEL E LYONS

Signature and Printed Name

04/03/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/19/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 23:56

Date of Inspection: 03/18/2019

Serial Number: 80-005935
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.077	0.198	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

DANIEL E LYONS

Signature and Printed Name

03/18/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/14/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 17:55

Date of Inspection: 02/07/2019

Serial Number: 80-005935
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.045	0.074 / 0.076	0.191	0.080
0.000	0.046	0.075 / 0.076	0.192	0.080
0.000	0.047	0.075 / 0.076	0.192	0.080


Number of Simulators Used: 5

Remarks:

08:CONTROL OUTSIDE TOL/SIM LID LOOSE/TIGHTEN/RE-TESTED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIEL E LYONS
Signature and Printed Name

02/07/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

1/25/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 22:19

Date of Inspection: 01/24/2019

Serial Number: 80-005935
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:646645 Exp: 12/05/2019
0.000	0.048	0.077	0.194	0.080
0.000	0.047	0.076	0.195	0.079
0.000	0.048	0.077	0.195	0.079

Number of Simulators Used: 5

Remarks:
A F / M A: .

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIEL E LYONS

Signature and Printed Name

01/24/2019
Date