

RECEIVED

AMENDED ON 09.16.2019

6027
JR

SEP 25 2019

Florida Department of Law Enforcement
Alcohol Testing Program

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 12:22

Date of Inspection: 09/13/2019

Serial Number: 30-006027
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#: 201802E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22517080A5 Exp: 12/05/2019 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.081 | 0.200 | 0.077 |
| 0.000 | 0.050 | 0.080 | 0.201 | 0.077 |
| 0.000 | 0.050 | 0.080 | 0.201 | 0.077 |

Number of Simulators Used: 4 of #6467

Remarks:

*** ALCOHOL FREE SIMULATOR NOT ATTACHED. RETESTED TO ENSURE ACCURACY AND RELIABILITY.**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

J #6467 JASON C KARLECKE
Signature and Printed Name

09/13/2019
Date