

Florida Department of Law Enforcement Alcohol Testing Program

80-006235
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTORIA PD
Time of Inspection: 00:36

Date of Inspection: 07/29/2019

Serial Number: 80-006235
Software: 8100.27

Check or Test		YES	NO
Date and/or Time Adjusted			No
Diagnostic Check (Pre-Inspection): OK		Yes	
Alcohol Free Subject Test: 0.000		Yes	
Mouth Alcohol Test: slope Not Met		Yes	
Interferent Detect Test: Interferent Detect		Yes	
Diagnostic Check (Post-Inspection): OK		Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201808E Exp: 11/15/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Sed Test (g/210L) Lot#: 109923I Exp: 03/01/2022
0.000	0.045	0.078	0.196	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.079	0.199	0.079

Number of Simulations Used: 5

Remarks:
AMBIENT FAIL ON ALCOHOL FREE TEST

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOHN A. CALHOUN
Signature and Printed Name

07/29/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-006235
13K

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESTACOLA PD

Date of Inspection: 23:34

Date of Inspection: 07/28/2019

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Batteries and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulations Used: 5

Remarks

00: Ambient Fail. Non-compliance: .

The above instrument complies () does not comply (X) with Chapter 12D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 12D-8, FAC.



JOHN A. CALHOUN
Signature and Printed Name

07/28/2019
Date