

# Florida Department of Law Enforcement Alcohol Testing Program

4/12/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO  
Time of Inspection: 23:27

Date of Inspection: 04/10/2019

Serial Number: 80-006222  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.046	0.078	0.184 / 0.197	0.082
0.000	0.047	0.078	0.191 / 0.198	0.082
0.000	0.047	0.079	0.195 / 0.198	0.082

Number of Simulators Used: 5

**Remarks:**

20: Control Outside Tolerance/RECONNECTED AND CHECKED TEMP..

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




PITCHER / BINDER

Signature and Printed Name

04/10/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

3/8/19  
JA

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO  
Time of Inspection: 20:08

Date of Inspection: 03/06/2019

Serial Number: 80-006222  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.049	0.080	0.199	0.083
0.000	0.049	0.080	0.199	0.083
0.000	0.050	0.081	0.200	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JA 2502 MIA WJ 1437 BINDER / PITCHER  
Signature and Printed Name

03/06/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

BK  
80-006222

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO  
Time of Inspection: 20:29

Date of Inspection: 02/13/2019

Serial Number: 80-006222  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.049	0.079	0.198	0.000 / 0.081
0.000	0.049	0.079	0.199	0.000 / 0.082
0.000	0.049	0.079	0.199	0.000 / 0.082

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance RECONNECTED THE HOSE.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MIDWEST 1437

BINDER / PITCHER

---

Signature and Printed Name

02/13/2019  
Date