

Florida Department of Law Enforcement Alcohol Testing Program

6/3/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTACULA PD
Time of Inspection: 22:08

Date of Inspection: 05/31/2019

Serial Number: 80-006285
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: flope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 2018113 Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/21/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 108923F Exp: 03/01/2022
0.000	0.046	0.077	0.194	0.081
0.000	0.047	0.077	0.196	0.081
0.000	0.047	0.077	0.196	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOHN A. CALHOUN

Signature and Printed Name

05/31/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

5/2/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD
Time of Inspection: 22:42

Date of Inspection: 04/30/2019

Serial Number: 80-006235
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 2018111 Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2018	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.03 g/210L Dry Gas Std Test (g/210L) Lot#: 103923E Exp: 03/01/2022
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.199	0.081
0.000	0.048	0.079	0.199	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-3, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-3, FAC.



JOHN A. CALHOUN
Signature and Printed Name

04/30/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
FDLE

APR 02 2019

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Alcohol Testing Program

Agency: ASTATULA PD
Time of Inspection: 00:43

Date of Inspection: 03/30/2019

Serial Number: 80-006235
Software: 6100.27

4/8/19
JD

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802C Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1099238 Exp: 03/01/2022
0.000	0.048	0.078	0.192	0.081
0.000	0.048	0.080	0.192	0.081
0.000	0.048	0.079	0.192	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-3, FAC.



JOHN A CALHOUN
Signature and Printed Name

03/30/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/5/19
JEZ

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 19:47

Date of Inspection: 02/28/2019

Serial Number: 80-006235

Software: 6100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/05/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802C Exp: 02/22/2020	0.28 g/210L Dry Gas Std Test (g/210L) Lot#: 970670 Exp: 06/07/2021
0.000	0.044 / 0.046	0.078	0.197	0.080
0.000	0.045 / 0.047	0.078	0.198	0.080
0.000	0.045 / 0.047	0.079	0.198	0.080


Number of Simulators Used: 5

Remarks:

05 Control Outside Tolerance

The above instrument complies () does not comply () with Chapter 110-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 110-8, FAC.



 Signature and Printed Name: JOEN A. CALHOUN

02/28/2019
 Date