

0236  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 12:24

Date of Inspection: 08/27/2019

Serial Number: 80-006236  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) <i>SC</i> Lot#: <del>201811C</del> Exp: <del>11/13/2020</del>	0.20g/210L Test (g/210L) Lot#: 201811G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 572920 Exp: 09/19/2019
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079

Number of Simulators Used: 5

Remarks:  
NONE

*Corrected .08 solution numbers  
201808E  
Exp 8/22/2020*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GLENN CLINE

Signature and Printed Name

08/27/2019  
Date