

Florida Department of Law Enforcement Alcohol Testing Program

4/10/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 15:14

Date of Inspection: 04/04/2019

Serial Number: 80-006385
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31517080A1 Exp: 01/05/2020
0.000	0.076 / 0.048	0.077	0.195	0.081
0.000	0.077 / 0.047	0.077	0.195	0.081
0.000	0.078 / 0.047	0.077	0.195	0.081

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. 0.08 HOOKED UP BY MISTAKE.

AI JULIE DAIL #254236 PRESENT ENTIRE INSPECTION.

Julie Dail

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

ROGER G SKIPPER

04/04/2019
Date

Reviewed
4-8-19
MSI. J.K. [Signature]
4/8