

SP
60631

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 09:22

Date of Inspection: 09/16/2019
Jac

Serial Number: 80-006631
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| Diagnostic Check (Pre-Inspection): OK | | No |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:201908E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:840875 Exp: 09/19/2020 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.000 / 0.080 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.000 / 0.079 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.000 / 0.079 | 0.200 | 0.079 |

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. .08 NOT CONNECTED, RETRY OK

D.O.I. - corrected

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name GLENN CLINE

Date 09/16/2019

Florida Department of Law Enforcement Alcohol Testing Program

SP
6631

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 09:22

2010
Date of Inspection: 08/26/2010 *see*

Serial Number: 80-006631
Software: 8100 27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| Diagnostic Check (Pre-Inspection): OK | Yes | No |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 840875 Exp: 09/19/2020 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.079 | 0.199 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.078 |
| 0.000 | 0.050 | 0.079 | 0.199 | 0.079 |

Number of Simulators Used: 5

Remarks:

D.O.I. - corrected

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Glenn Cline

Signature and Printed Name GLENN CLINE

Date 08/26/2010