

# Florida Department of Law Enforcement Alcohol Testing Program

6/21/19  
JO

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 18:48

Date of Inspection: 04/29/2019

Serial Number: 80-006632  
Software: 8100.27


Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 02/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 972301 Exp: 06/15/2021
0.000	0.047	0.077	0.198	0.080
0.000	0.047	0.077	0.199	0.080
0.000	0.047	0.077	0.199	0.080

Number of Simulators Used: 5

Remarks: *NONE*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.  
I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name ALAIN HERNANDEZ

04/29/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

6/3/19  
JO

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 13:52

Date of Inspection: 05/29/2019

Serial Number: 80-006632  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 972301 Exp: 06/15/2021
0.000	0.044 / 0.045	0.075	0.196	0.081
0.000	0.045 / 0.046	0.075	0.196	0.081
0.000	0.045 / 0.046	0.075	0.196	0.081

Number of Simulators Used: 5

**Remarks:**

05: Control Outside Tolerance, ROOM TEMP TO LOW

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALAIN HERNANDEZ

Signature and Printed Name

05/29/2019  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

3/28/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 17:57

Date of Inspection: 03/28/2019

Serial Number: 80-006632  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201803B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 972301 Exp: 06/15/2021
0.000	0.046	0.075	0.198	0.081
0.000	0.047	0.076	0.198	0.081
0.000	0.047	0.077	0.198	0.081

Number of Simulators Used: 5

Remarks: NONE

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALAIN HERNANDEZ  
Signature and Printed Name

03/28/2019  
Date