

Florida Department of Law Enforcement Alcohol Testing Program

6/21/19
JA

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 18:28

Date of Inspection: 04/29/2019

Serial Number: 80-006637
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 915310 Exp: 03/07/2021
0.000	0.049	0.080	0.202	0.080
0.000	0.048	0.080	0.203	0.078
0.000	0.049	0.080	0.201	0.079

Number of Simulators Used: 5

Remarks: NONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name ALAIN HERNANDEZ

04/29/2019
 Date

Florida Department of Law Enforcement Alcohol Testing Program

6/3/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 13:36

Date of Inspection: 05/29/2019

Serial Number: 80-006637
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:915310 Exp: 03/07/2021
0.000	0.046	0.077	0.199	0.080
0.000	0.046	0.077	0.200	0.080
0.000	0.046	0.078	0.199	0.080

Number of Simulators Used: 5

Remarks: NONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALAIN HERNANDEZ

Signature and Printed Name

05/29/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2052

AGENCY INSPECTION REPORT - INTOXILYZER 8000

3/29/19
JA

Agency: FL HIGHWAY PATROL
Time of Inspection: 18:37

Date of Inspection: 03/28/2019

Serial Number: 80-006637
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 915310 Exp: 03/07/2021
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.080	0.202	0.081
0.000	0.050	0.080	0.203	0.081

Number of Simulators Used: 5

Remarks:
AF/MA: 1st Attempt, too much M/A, had to retry a second time for AF/MA which was successful. (A+)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name ALAIN HERNANDEZ

Date 03/28/2019

Florida Department of Law Enforcement Alcohol Testing Program

1 of 2

AGENCY INSPECTION REPORT - INTOXILYZER 8000

3/29/19 ja

Agency: FL HIGHWAY PATROL
Time of Inspection: 17:23

Date of Inspection: 03/28/2019

Serial Number: 80-006637
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: 5

Remarks:

SEE AI REPORT REMARKS Non-compliance: -
 DURING AFMA TEST, TOO MUCH MIA WAS USED, TEST WAS TERMINATED AND REATTEMPTED
 SEE SECOND TEST PAGE FOR THIS INSTRUMENT BEARING #80-006637 TEST WAS PERFORMED
 INSPECTION AT 18:37.
AH

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALAIN HERNANDEZ
Signature and Printed Name

03/28/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/11/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 19:05

Date of Inspection: 02/27/2019

Serial Number: 80-006637
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) <i>2018020</i> ^{AT} Lot#: 201808E Exp: 08/22/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201803G Exp: 02/23/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 915810 Exp: 03/07/2021
0.000	0.046	0.080	0.200	0.078
0.000	0.046	0.078	0.202	0.078
0.000	0.049	0.077	0.200	0.079

Number of Simulators Used: 6

Remarks: 0.05g/210L Test, Wrong Lot # inputted, was corrected on this form ^{AT} 2/28/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name ALAIN HERNANDEZ

02/27/2019
Date