

6761 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-006761

Time of Inspection: 11:15

Date of Inspection: 01/21/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201702A ^{OK} Exp: 02/23/2019 Lot # 201702B	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 867139 Exp: 11/15/2020
0.000	0.051	0.081	0.202	0.081
0.000	0.051	0.081	0.202	0.081
0.000	0.051	0.081	0.202	0.081

Number of Simulators Used: 5

Remarks:

DEP CLINE (MCSO) PRESENT

corrected lot # for 20 solution

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GLENN CLINE

Signature and Printed Name

01/21/2019

Date