

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD
Time of Inspection: 12:11

Date of Inspection: 11/13/2019

Serial Number: 80-006938
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A3 Exp: 03/05/2020
0.000	0.047	0.076	0.196	0.077
0.000	0.047	0.077	0.195	0.077
0.000	0.048	0.077	0.196	0.076

Number of Simulators Used: 5

Remarks:

J PITHIAN PRESENT AND OBSERVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

K. Keating

KAYLA C KEATING

Signature and Printed Name

11/13/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD
Time of Inspection: 13:46

Date of Inspection: 06/13/2019

Serial Number: 80-006938
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 06/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A3 Exp: 03/05/2020
0.000	0.047	0.071 / 0.077	0.195	0.078
0.000	0.047	0.072 / 0.077	0.196	0.078
0.000	0.047	0.072 / 0.078	0.196	0.078

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. BAD SEAL. A FRAZIER OBS.

Wrong exp. date entered for .08 Alcohol Reference Solution. Correct expiration date is 08/22/2020.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Kayla Keating _____ KAYLA C KEATING
Signature and Printed Name

06/13/2019
Date