

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ALACHUA COUNTY SO
Time of Inspection: 15:00

Date of Inspection: 02/20/2019

Serial Number: 80-000737
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.080	0.195	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.078

Standard Deviations	0.0003	0.0004	0.0008	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland

BRETT H KIRKLAND

Signature and Printed Name

02/20/2019
Date

*02/21/19
JB*

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
02/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:20
Control Test	0.047	12:21
Air Blank	0.000	12:21
Control Test	0.048	12:22
Air Blank	0.000	12:23
Control Test	0.050	12:23
Air Blank	0.000	12:24
Control Test Stats		
Average	0.0483	
Std Dev	0.0015	
Rel. Std Dev(%)	3.1614	

BK
Operator's Signature

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
02/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:25
Control Test	0.079	12:26
Air Blank	0.000	12:26
Control Test	0.079	12:27
Air Blank	0.000	12:27
Control Test	0.079	12:28
Air Blank	0.000	12:29
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

BK
Operator's Signature

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
02/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:30
Control Test	0.196	12:30
Air Blank	0.000	12:31
Control Test	0.196	12:32
Air Blank	0.000	12:32
Control Test	0.197	12:33
Air Blank	0.000	12:33
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2941	

BK
Operator's Signature

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
02/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.000	12:35
Air Blank	0.000	12:35
Control Test	0.080	12:36
Air Blank	0.000	12:36
Control Test	0.080	12:36
Air Blank	0.000	12:37
Control Test Stats		
Average	0.0533	
Std Dev	0.0462	
Rel. Std Dev(%)	86.6025	

DGS
Repeat
Forgot to plug in
hose.

BK
Operator's Signature

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
02/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:38
Control Test	0.080	12:38
Air Blank	0.000	12:39
Control Test	0.080	12:39
Air Blank	0.000	12:40
Control Test	0.080	12:40
Air Blank	0.000	12:40
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

BK
Operator's Signature

DGS
2/21/19
ja

80-000737
Stability Check
BK 2/20/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000737, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000737</u>	UNCERTAINTY* ±
Owning Agency:	<u>ALACHUA COUNTY SO</u>	0.050 g/ 210 L
Calibration Date:	<u>02/20/2019</u>	0.080 g/ 210 L
Calibration Time:	<u>15:00</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/20/2019 Date
 BRETT H KIRKLAND,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

2/21/19
JD