



INSTRUMENT PROCESSING SHEET

Agency NAS Jacksonville PD

S/N 80-000778

Florida Department of
Law Enforcement

Date In 05/03/2019 DI Completion Date 05/10/2019

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DP</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>JE</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>263</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.183</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804 11/14/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1018	201707D 07/25/2019	0.080	SD3962	201707E 07/25/2019	0.200	G2078	201707C 07/24/2019	0.080 DGS	N/A	AG831804 11/14/2020	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>JE</u> ☒ Lab Temp °C <u>21.3</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>SD1012</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1013</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	SD1018	201707D 07/25/2019																																																												
0.080	SD3962	201707E 07/25/2019																																																												
0.200	G2078	201707C 07/24/2019																																																												
0.080 DGS	N/A	AG831804 11/14/2020																																																												
Final Release Date FDLE MAY 13 2019 Alcohol Testing Program	Calibration Adjustment Performed By <u>JE</u> Barometric Pressure Gauge <u>1012</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>DR2035</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1022</td> <td>17410</td> <td>12/06/2019</td> </tr> <tr> <td>0.100</td> <td>SD3964</td> <td>18200</td> <td>07/03/2020</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>19040</td> <td>01/29/2021</td> </tr> <tr> <td>0.300</td> <td>SD1024</td> <td>18110</td> <td>04/02/2020</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>08/05/2019</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804</td> <td>11/14/2020</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	DR2035	N/A	N/A	0.040	SD1022	17410	12/06/2019	0.100	SD3964	18200	07/03/2020	0.200	SD1025	19040	01/29/2021	0.300	SD1024	18110	04/02/2020	0.080 DGS	N/A	17817080A2	08/05/2019	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1012	201707D	07/25/2019	0.080	DR1279	201707E	07/25/2019	0.200	SD1013	201707C	07/24/2019	0.080 DGS	N/A	AG831804	11/14/2020	Department Inspection Performed By <u>JE</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1015</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Extra 0.08 DGS</u> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>SP 5/10/19</u> Tech Review / Date <u>Brett Kirkland 5/13/19</u> Admin Review / Date	Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000	DR2035	N/A	N/A																																																											
0.040	SD1022	17410	12/06/2019																																																											
0.100	SD3964	18200	07/03/2020																																																											
0.200	SD1025	19040	01/29/2021																																																											
0.300	SD1024	18110	04/02/2020																																																											
0.080 DGS	N/A	17817080A2	08/05/2019																																																											
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050	SD1012	201707D	07/25/2019																																																											
0.080	DR1279	201707E	07/25/2019																																																											
0.200	SD1013	201707C	07/24/2019																																																											
0.080 DGS	N/A	AG831804	11/14/2020																																																											
Simulator	Serial Number																																																													
0.000	G2408																																																													
Interferent	G2882																																																													
0.050	SD1012																																																													
0.080	DR1279																																																													
0.200	SD1013																																																													
Notes/Suggested Service: <u>Please change the level 2 password to something unique.</u> <u>JE</u> _____ _____ _____ _____																																																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD
Time of Inspection: 09:58

Date of Inspection: 05/10/2019

Serial Number: 80-000778
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.052	0.079	0.200	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.078	0.200	0.079
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.078	0.201	0.079
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.079	0.200	0.079

Standard Deviations	0.0008	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

SP BK
5/13/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

THOMAS J GRAHAM

05/10/2019
Date

80-000778

5/8/19
JD

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
05/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:33
Control Test	0.052	07:34
Air Blank	0.000	07:34
Control Test	0.050	07:35
Air Blank	0.000	07:35
Control Test	0.051	07:36
Air Blank	0.000	07:37
Control Test Stats		
Average	0.0510	
Std Dev	0.0010	
Rel Std Dev(%)	1.9608	

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
05/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:38
Control Test	0.080	07:38
Air Blank	0.000	07:39
Control Test	0.080	07:40
Air Blank	0.000	07:40
Control Test	0.081	07:41
Air Blank	0.000	07:41
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
05/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:43
Control Test	0.202	07:43
Air Blank	0.000	07:44
Control Test	0.202	07:45
Air Blank	0.000	07:45
Control Test	0.203	07:46
Air Blank	0.000	07:46
Control Test Stats		
Average	0.2023	
Std Dev	0.0006	
Rel Std Dev(%)	0.2853	

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
05/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:49
Control Test	0.079	07:49
Air Blank	0.000	07:49
Control Test	0.076	07:50
Air Blank	0.000	07:50
Control Test	0.077	07:51
Air Blank	0.000	07:51
Control Test Stats		
Average	0.0773	
Std Dev	0.0015	
Rel Std Dev(%)	1.9753	

DGS

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
05/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:56
Control Test	0.079	07:56
Air Blank	0.000	07:57
Control Test	0.077	07:57
Air Blank	0.000	07:58
Control Test	0.078	07:58
Air Blank	0.000	07:59
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

DGS

#2

JD

Operator's Signature

SP RK 5/13/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000778, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000778</u>	UNCERTAINTY * $\pm$	
Owning Agency:	<u>NAS JACKSONVILLE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/10/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:58</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/10/2019

Date

THOMAS J GRAHAM,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

SR
BK
5/13/19

80-000778
Calibration
5/10/19
JD

NR5 JACKSONVILLE PD
Intoxilizer - Alcotest Analyzer
Model 8000
05/10/2019
SN 80-000778
07:01:46
Auto Calibration
Max Power Res Value = 22
Auto Range Res Value = 10

Sample % Abs (% Abs Ref)
Sample #1 = 1.5380 (-0.0150)
Sample #2 = 1.5330 (-0.0190)
Sample #3 = 1.5470 (-0.0250)
Sample #4 = 1.5300 (-0.0070)
Avg % Abs = 1.5367 (-0.0137)
STD DEV = 0.0091 (-0.0186)
REL STD DEV = 0.590 (135.974)

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12448, Sum Io = 12926
Sample % Abs (% Abs Ref)
Sample #1 = -0.1160 (-0.0600)
Sample #2 = 0.1250 (-0.2050)
Sample #3 = 0.1640 (-0.2220)
Sample #4 = 0.1170 (-0.1570)
Avg % Abs = 0.1353 (-0.1947)
STD DEV = 0.0251 (-0.0337)
REL STD DEV = 18.581 (17.317)

Sample % Abs (% Abs Ref)
Sample #1 = -0.0710 (-0.0390)
Sample #2 = 0.1330 (-0.1730)
Sample #3 = 0.1200 (-0.2090)
Sample #4 = 0.1120 (-0.1810)
Avg % Abs = 0.1217 (-0.1877)
STD DEV = 0.0106 (-0.0189)
REL STD DEV = 8.711 (10.073)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12459, Sum Io = 12944
Sample % Abs (% Abs Ref)
Sample #1 = 0.9120 (-0.0200)
Sample #2 = 0.8710 (-0.0300)
Sample #3 = 0.8720 (-0.0400)
Sample #4 = 0.9060 (-0.0670)
Avg % Abs = 0.8830 (-0.0367)
STD DEV = 0.0199 (-0.0321)
REL STD DEV = 2.256 (87.627)

Sample % Abs (% Abs Ref)
Sample #1 = 1.5380 (-0.0150)
Sample #2 = 1.5330 (-0.0190)
Sample #3 = 1.5470 (-0.0250)
Sample #4 = 1.5300 (-0.0070)
Avg % Abs = 1.5367 (-0.0137)
STD DEV = 0.0091 (-0.0186)
REL STD DEV = 0.590 (135.974)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12461, Sum Io = 12954
Sample % Abs (% Abs Ref)
Sample #1 = 2.1540 (-0.0100)
Sample #2 = 1.9860 (-0.1240)
Sample #3 = 2.0020 (-0.1000)
Sample #4 = 2.0590 (-0.0830)
Avg % Abs = 2.0157 (-0.1023)
STD DEV = 0.0384 (-0.0206)
REL STD DEV = 1.904 (20.130)

Sample % Abs (% Abs Ref)
Sample #1 = 3.7050 (0.0000)
Sample #2 = 3.5400 (0.1030)
Sample #3 = 3.6340 (0.0710)
Sample #4 = 3.6340 (0.0910)
Avg % Abs = 3.6027 (0.0883)
STD DEV = 0.0543 (0.0162)
REL STD DEV = 1.506 (18.301)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12453, Sum Io = 12946
Sample % Abs (% Abs Ref)
Sample #1 = 3.8970 (-0.0220)
Sample #2 = 3.8380 (0.0800)
Sample #3 = 3.8150 (-0.1270)
Sample #4 = 3.8790 (0.0790)
Avg % Abs = 3.8440 (0.0953)
STD DEV = 0.0324 (0.0274)
REL STD DEV = 0.843 (28.771)

Sample % Abs (% Abs Ref)
Sample #1 = 0.8990 (-0.0170)
Sample #2 = 6.8610 (0.0520)
Sample #3 = 6.8730 (0.0710)
Sample #4 = 6.9390 (0.0460)
Avg % Abs = 6.8910 (0.0563)
STD DEV = 0.0420 (0.0131)
REL STD DEV = 0.609 (23.168)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12444, Sum Io = 12944
Sample % Abs (% Abs Ref)
Sample #1 = 5.5910 (-0.0310)
Sample #2 = 5.5180 (0.0750)
Sample #3 = 5.5380 (0.1030)
Sample #4 = 5.5760 (0.1130)
Avg % Abs = 5.5440 (0.0970)
STD DEV = 0.0295 (0.0197)
REL STD DEV = 0.531 (20.307)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12444, Sum Io = 12944
Sample % Abs (% Abs Ref)
Sample #1 = 5.5910 (-0.0310)
Sample #2 = 5.5180 (0.0750)
Sample #3 = 5.5380 (0.1030)
Sample #4 = 5.5760 (0.1130)
Avg % Abs = 5.5440 (0.0970)
STD DEV = 0.0295 (0.0197)
REL STD DEV = 0.531 (20.307)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12444, Sum Io = 12944
Sample % Abs (% Abs Ref)
Sample #1 = 5.5910 (-0.0310)
Sample #2 = 5.5180 (0.0750)
Sample #3 = 5.5380 (0.1030)
Sample #4 = 5.5760 (0.1130)
Avg % Abs = 5.5440 (0.0970)
STD DEV = 0.0295 (0.0197)
REL STD DEV = 0.531 (20.307)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12444, Sum Io = 12944
Sample % Abs (% Abs Ref)
Sample #1 = 5.5910 (-0.0310)
Sample #2 = 5.5180 (0.0750)
Sample #3 = 5.5380 (0.1030)
Sample #4 = 5.5760 (0.1130)
Avg % Abs = 5.5440 (0.0970)
STD DEV = 0.0295 (0.0197)
REL STD DEV = 0.531 (20.307)

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0006
0.040 0.039 0.0007
0.100 0.100 0.0005
0.200 0.201 -0.0009
0.300 0.300 0.0003

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0003
0.100 0.100 0.0004
0.200 0.201 -0.0007
0.300 0.300 0.0003

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 2668.00
Sample #2 = 2671.00
Sample #3 = 2608.00
Sample #4 = 2758.00
Average Result = 2679.0000
STD DEV = 75.3193
REL STD DEV = 2.811
***** CHANNEL 2
Sample #1 = 3060.00
Sample #2 = 3050.00
Sample #3 = 3066.00
Sample #4 = 3133.00
Average Result = 3083.0000
STD DEV = 44.0341
REL STD DEV = 1.428

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1013
3 um H2O Adjust (mg/l*10,000) = 1130
9 um H2O Adjust (mg/l*10,000) = 726
**** AUTO CAL PASS

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 2668.00
Sample #2 = 2671.00
Sample #3 = 2608.00
Sample #4 = 2758.00
Average Result = 2679.0000
STD DEV = 75.3193
REL STD DEV = 2.811
***** CHANNEL 2
Sample #1 = 3060.00
Sample #2 = 3050.00
Sample #3 = 3066.00
Sample #4 = 3133.00
Average Result = 3083.0000
STD DEV = 44.0341
REL STD DEV = 1.428

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1013
3 um H2O Adjust (mg/l*10,000) = 1130
9 um H2O Adjust (mg/l*10,000) = 726
**** AUTO CAL PASS

SR BK 5/13/19

80-000778

Post Stability Checks

5/10/19
JD

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:39
Control Test	0.045	07:39
Air Blank	0.000	07:40
Control Test	0.049	07:40
Air Blank	0.000	07:41
Control Test	0.049	07:42
Air Blank	0.000	07:42
Control Test Stats		
Average	0.0477	
Std Dev	0.0023	
Rel Std Dev(%)	4.8449	

JD 0.05 #1

Possible cold tubing.
Repeat test.

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:43
Control Test	0.050	07:44
Air Blank	0.000	07:44
Control Test	0.050	07:45
Air Blank	0.000	07:45
Control Test	0.050	07:46
Air Blank	0.000	07:47
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD 0.05 #2

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:48
Control Test	0.082	07:48
Air Blank	0.000	07:49
Control Test	0.081	07:50
Air Blank	0.000	07:50
Control Test	0.081	07:51
Air Blank	0.000	07:51
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:52
Control Test	0.202	07:53
Air Blank	0.000	07:53
Control Test	0.200	07:54
Air Blank	0.000	07:55
Control Test	0.201	07:55
Air Blank	0.000	07:56
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

JD

Operator's Signature

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:57
Control Test	0.080	07:57
Air Blank	0.000	07:58
Control Test	0.079	07:58
Air Blank	0.000	07:58
Control Test	0.078	07:59
Air Blank	0.000	07:59
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

DGS

SP BK 5/13/19
JD

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD

Time of Inspection: 07:28

Date of Inspection: 05/08/2019

Serial Number: 80-000778

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

THOMAS J GRAHAM

THOMAS J GRAHAM

Signature and Printed Name

05/08/2019
Date

NAS JACKSONVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000778
05/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:00
Control Test	0.078	08:00
Air Blank	0.000	08:00
Control Test	0.079	08:01
Air Blank	0.000	08:01
Control Test	0.077	08:01
Air Blank	0.000	08:02
Control Test	0.079	08:02
Air Blank	0.000	08:03
Control Test	0.077	08:03
Air Blank	0.000	08:04
Control Test	0.077	08:04
Air Blank	0.000	08:04
Control Test	0.078	08:05
Air Blank	0.000	08:05
Control Test	0.079	08:06
Air Blank	0.000	08:06
Control Test	0.077	08:06
Air Blank	0.000	08:07
Control Test	0.078	08:07
Air Blank	0.000	08:08
Control Test	0.078	08:08
Air Blank	0.000	08:08
Control Test	0.079	08:09
Air Blank	0.000	08:09
Control Test	0.080	08:10
Air Blank	0.000	08:10
Control Test	0.078	08:10
Air Blank	0.000	08:11
Control Test	0.079	08:11
Air Blank	0.000	08:12
Control Test Stats		
Average	0.0782	
Std Dev	0.0009	
Rel Std Dev(%)	1.2035	


Operator's Signature

SR BK
5/13/19

0.08 g/210L
ARS
5/8/19
JD