



INSTRUMENT PROCESSING SHEET

Agency Sumter County


S/N 80-000816

Florida Department of
Law Enforcement

Date In 10/28/2019 DI Completion Date 10/31/19

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DP</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date FDLE NOV 04 2019 Alcohol Testing Program	Quality Checks Performed By <u>SP</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>215</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>1160</u> (.139 - .169) 36 mm <u>179</u> (.156 - .190) 53 mm <u>242</u> (.228 - .278) 103 mm <u>503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201905A</u> <u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201905B</u> <u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> <td><u>201904D</u> <u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG831804</u> <u>11-14-20</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201905A</u> <u>5-14-21</u>	0.080	<u>DR1279</u>	<u>201905B</u> <u>5-14-21</u>	0.200	<u>SD1013</u>	<u>201904D</u> <u>4-30-21</u>	0.080 DGS	N/A	<u>AG831804</u> <u>11-14-20</u>	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>SP</u> ☒ Lab Temp °C <u>21.3</u> External Digital Therm. ID#: <u>300603</u> ☒ 34°C ±.2 Serial #: <u>SD1012</u> ☒ 34°C ±.2 Serial #: <u>DR1279</u> ☒ 34°C ±.2 Serial #: <u>SD1013</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1012</u>	<u>201905A</u> <u>5-14-21</u>																																																											
0.080	<u>DR1279</u>	<u>201905B</u> <u>5-14-21</u>																																																											
0.200	<u>SD1013</u>	<u>201904D</u> <u>4-30-21</u>																																																											
0.080 DGS	N/A	<u>AG831804</u> <u>11-14-20</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1014</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2408</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2882</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1012</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>DP</u> 11/4/19 <u>Brett Kirkland</u> 11/4/19 Tech Review / Date Admin Review / Date	Simulator	Serial Number	0.000	<u>G2408</u>	Interferent	<u>G2882</u>	0.050	<u>SD1012</u>	0.080	<u>DR1279</u>	0.200	<u>SD1013</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	<u>G2408</u>																																																												
Interferent	<u>G2882</u>																																																												
0.050	<u>SD1012</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>SD1013</u>																																																												



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000816, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000816</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>SUMTER COUNTY SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>10/31/2019</u>	0.080 g/210 L	0.004
Calibration Time:	<u>14:27</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/31/2019

Date

Shayla Platt

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

QDM
TSK
11/4/19

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 14:27

Date of Inspection: 10/31/2019

Serial Number: 80-000816
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) <i>201905A SP</i> Lot#: 201905B- Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#: 201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#: 201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG831804 Exp: 11/14/2020
0.000	0.047	0.077	0.196	0.079
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.079

Standard Deviations	0.0003	0.0008	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SPM
1312
11/4/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

10/31/2019
Date

Stability Checks # 80-000816

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
10/31/2019
Software: 8100.27
SN 80-000816

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.048	10:41
Air Blank	0.000	10:41
Control Test	0.049	10:42
Air Blank	0.000	10:42
Control Test	0.047	10:43
Air Blank	0.000	10:44
Control Test Stats		
Average	0.0480	
Std Dev	0.0010	
Rel Std Dev(%)	2.0833	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
10/31/2019
Software: 8100.27
SN 80-000816

Test	g/210L	Time
Air Blank	0.000	10:29
Control Test	0.079	10:30
Air Blank	0.000	10:30
Control Test	0.078	10:31
Air Blank	0.000	10:31
Control Test	0.079	10:32
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
10/31/2019
Software: 8100.27
SN 80-000816

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.196	10:25
Air Blank	0.000	10:26
Control Test	0.198	10:26
Air Blank	0.000	10:27
Control Test	0.198	10:28
Air Blank	0.000	10:28
Control Test Stats		
Average	0.1973	
Std Dev	0.0012	
Rel Std Dev(%)	0.5852	

SP

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
10/31/2019
Software: 8100.27
SN 80-000816

Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.079	10:34
Air Blank	0.000	10:35
Control Test	0.078	10:35
Air Blank	0.000	10:36
Control Test	0.080	10:36
Air Blank	0.000	10:36
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2656	

SP

SP

Operator's Signature

OGM
BCK
11/4/19

Return Material Authorization

Ship to:

☐

CMI, Inc.

☒

Enforcement Electronics

Shipment to repair facility authorized by: Sgt. Gary Forinash on 5/21/2019

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000816

Bill To Address:

Sumter County SO

219 E. Anderson Ave.

Bushnell, FL 33513

Ship to Address:

FDLE Off-site Mail Facility

Alcohol Testing Program

813-B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Please evaluate flow sensor.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sgt. Gary Forinash

Phone #: 352-569-1713 Email: GaryForinash@sumtercountysheriff

ATP Contact Name: TJ Graham ATP Email: thomasgraham@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Sumter County SOS/N 80-000816Florida Department of
Law EnforcementDate In 05/21/2019 DI Completion Date 05/21/2019☐ Ship ☐ P/U ☐ H/D ☐ CMI ☒ EE

Intake		Quality Checks		Flow Calibration																																									
Performed By <u>DP</u>		Performed By <u>JD</u>		Performed By <u>JD</u>																																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>152</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.089</u> (.139 - .169) 36 mm <u>.113</u> (.156 - .190) 53 mm <u>.187</u> (.228 - .278) 103 mm <u>.453</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>SD1012</td><td>201707D 07/25/2019</td></tr><tr><td>0.080</td><td>DR1279</td><td>201707E 07/25/2019</td></tr><tr><td>0.200</td><td>SD1013</td><td>201707C 07/24/2019</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG831804 11/14/2020</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	SD1013	201707C 07/24/2019	0.080 DGS	N/A	AG831804 11/14/2020	Flow Column # <u>ATP-105</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>148</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.480</u> (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>JD</u> ☒ Lab Temp °C <u>21.0</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>SD1012</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1013</u>																										
Simulator	Serial #	Lot #/Exp																																											
0.050	SD1012	201707D 07/25/2019																																											
0.080	DR1279	201707E 07/25/2019																																											
0.200	SD1013	201707C 07/24/2019																																											
0.080 DGS	N/A	AG831804 11/14/2020																																											
Final Release Date																																													
Calibration Adjustment		Performed By _____		Department Inspection																																									
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28662</u> Gauge <u>1013</u> Instrument <u>1006</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u>																																									
<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>G2408</td></tr><tr><td>Interferent</td><td>G2882</td></tr><tr><td>0.050</td><td>SD1012</td></tr><tr><td>0.080</td><td>DR1279</td></tr><tr><td>0.200</td><td>SD1013</td></tr></tbody></table>		Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	G2408																																												
Interferent	G2882																																												
0.050	SD1012																																												
0.080	DR1279																																												
0.200	SD1013																																												
☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: <u>R value dropped to a post</u> <u>Department Inspection value of 136. I suggest</u> <u>sending this instrument to a repair facility to</u> <u>evaluate the flow sensor.</u> <u>JD</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																									
				8P 5/21/19 <u>Brett Kirkland</u> 5/21/19 Tech Review / Date Admin Review / Date																																									

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 13:58

Date of Inspection: 05/21/2019

Serial Number: 80-000816
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.050	0.079	0.196	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.079	0.197	0.080
0.000	0.049	0.080	0.196	0.079
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.080

Standard Deviations	0.0004	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP BK
5/21/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



THOMAS J GRAHAM
Signature and Printed Name

05/21/2019
Date

80-000816

5/18/19
JQ

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000816
05/21/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:10
Control Test	0.050	12:10
Air Blank	0.000	12:11
Control Test	0.049	12:12
Air Blank	0.000	12:12
Control Test	0.050	12:13
Air Blank	0.000	12:14
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

JQ

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000816
05/21/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.080	12:15
Air Blank	0.000	12:16
Control Test	0.079	12:17
Air Blank	0.000	12:17
Control Test	0.079	12:18
Air Blank	0.000	12:18
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

JQ

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000816
05/21/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.198	12:20
Air Blank	0.000	12:21
Control Test	0.197	12:21
Air Blank	0.000	12:22
Control Test	0.197	12:22
Air Blank	0.000	12:23
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

JQ

Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000816
05/21/2019
Software: 8100.27

Flow Calibration
5/21/19
JQ

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SORT(Diff) = 6.926
2: Rate (Liters/min) = 15
SORT(Diff) = 11.660
3: Rate (Liters/min) = 30
SORT(Diff) = 20.902
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 590
Rounded Intercept = -657792
Correlation = 0.99761

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000816
05/21/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.079	12:24
Air Blank	0.000	12:25
Control Test	0.080	12:25
Air Blank	0.000	12:26
Control Test	0.080	12:26
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

SP BX
5/21/19

JQ

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000816, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000816</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>SUMTER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/21/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:58</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

05/21/2019

Date


THOMAS J GRAHAM,
Department Inspector

Service • Integrity • Respect • Quality

SP BK 5/24/19