FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Citrus County S/N 80-000819

Date In 11/27/2019 DI Completion Date 12-04-2019 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By DP

☑ Annual
☐ Registration
☑ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☑ Power cord ☐ Printer Cable
☑ Static Bag ☐ 12V DC Cable

Notes:

Final Release Date

FDLE
DEC 09 2019
Alcohol Testing Program

Quality Checks Performed By MK

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 210
☑ Flow Verification (L/s)
Flow Column # ATP 102
32 mm 0.148 (.139 - .169)
36 mm 0.160 (.156 - .190)
53 mm 0.234 (.228 - .278)
103 mm 0.500 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 30793
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5088	201905A 05/14/2021
0.080	MP5089	201905B 05/14/2021
0.200	MP5090	201904D 04/30/2021
0.080 DGS	N/A	AG916501 06/14/2021

Flow Calibration Performed By

Flow Column #
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value
☐ Post Calibration Verification (L/s)
Flow Column #
32 mm (.139 - .169)
36 mm (.156 - .190)
53 mm (.228 - .278)
103 mm (.447 - .547)

Maintenance Performed By

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other

Temperature Checks Performed By MK

☑ Lab Temp °C 22.00 / 21.76
External Digital Therm. ID#: 300505 / 300505
☑ 34°C +-.2 Serial #: MP5088 / MP5088
☑ 34°C +-.2 Serial #: MP5089 / MP5089
☑ 34°C +-.2 Serial #: MP5090 / MP5090

Calibration Adjustment Performed By MK

Barometric Pressure Gauge 1019 ID # 28421

Simulator	Serial Number	Lot Number	Expiration
0.000	G8144	N/A	N/A
0.040	G2403	19080	03/04/2021
0.100	G2879	19160	07/09/2021
0.200	G3709	19046	01/29/2021
0.300	G8149	19010	01/03/2021
0.080 DGS	N/A	0351908044	04/05/2021

☑ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	MP5088	201905A	05/14/2021
0.080	MP5089	201905B	05/14/2021 MK
0.200	MP5090	201904D	~~06/30/2021~~
0.080 DGS	N/A	AG916501	06/14/2021

Department Inspection Performed By MK

Barometric Pressure ID# 26932 / 26932
Gauge 1014 / 1019 Instrument 1011 / 1020
Mouth Alcohol Solution Lot # 2019-B / 2019-B
Acetone Stock Solution Lot # 2019-A / 2019-A

Simulator	Serial Number
0.000	MP5086 / MP5086
Interferent	MP5087 / MP5087
0.050	MP5088 / MP5088
0.080	MP5089 / MP5089
0.200	MP5090 / MP5090

Attachments

☑ Form 41 ☑ Post-Stability Checks
☑ Stability Checks ☐ Flow Calibration
☑ Calibration Certificate ☑ Form 40
☑ Calibration Adjustment ☐ Other

Notes/Suggested Service: Post Cal 0.200 Exp: 04/30/2021

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☑ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
☑ **Conduct an Agency Inspection Before Evidentiary Use**

SP 12/~~4~~ 5/19 SP Brett Kirkland 12/6/19
Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO
Time of Inspection: 14:37 Date of Inspection: 12/05/2019
Serial Number: 80-000819
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.078

Standard Deviations	0.0004	0.0004	0.0000	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks: Post Calibration Adjustment DI

SP
BK
12/6/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael D Haughey MICHAEL D HAUGHEY
Signature and Printed Name

12/05/2019
Date

FDLE/ATP Form 41 –Revised August 2005

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO
Time of Inspection: 14:18   Date of Inspection: 12/04/2019   Serial Number: 80-000819
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.079	0.198	0.081
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.081
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.081
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080

Standard Deviations	0.0000	0.0000	0.0000	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

SP BK 12/6/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D HAUGHEY
Signature and Printed Name

12/04/2019
Date

80-000819 Stability Check 12/04/2019 MX

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:22
Control Test	0.049	11:23
Air Blank	0.000	11:23
Control Test	0.048	11:24
Air Blank	0.000	11:24
Control Test	0.048	11:25
Air Blank	0.000	11:25
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

MX
Operator's Signature

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:41
Control Test	0.077	11:41
Air Blank	0.000	11:42
Control Test	0.077	11:42
Air Blank	0.000	11:43
Control Test	0.077	11:44
Air Blank	0.000	11:44
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MX
Operator's Signature

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:57
Control Test	0.198	11:58
Air Blank	0.000	11:58
Control Test	0.197	11:59
Air Blank	0.000	11:59
Control Test	0.198	12:00
Air Blank	0.000	12:01
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

MX
Operator's Signature

DGS

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:49
Control Test	0.081	11:49
Air Blank	0.000	11:50
Control Test	0.081	11:50
Air Blank	0.000	11:51
Control Test	0.080	11:51
Air Blank	0.000	11:51
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

MX
Operator's Signature

SP
BK
12/6/19

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000819 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000819		UNCERTAINTY* ±
Owning Agency:	CITRUS COUNTY SO	0.050 g/ 210 L	0.004
Calibration Date:	12/05/2019	0.080 g/ 210 L	0.004
Calibration Time:	14:37	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

12/05/2019
Date

MICHAEL D HAUGHEY,
Department Inspector

SP
TSK
~~12/7~~ 12/6/19 RK

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/05/2019 08:33:30

Auto Calibration
Max Power Res Value = 36
Auto Range Res Value = 25

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12836, 9um Io = 13655

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.0830	(-0.0130)
Sample #2 =	0.1030	(0.0100)
Sample #3 =	0.0990	(0.0020)
Sample #4 =	0.1120	(0.0360)

Avg % Abs = 0.1047 (0.0160)
STD DEV = 0.0067 (0.0178)
REL STD DEV = 6.361 (111.102)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.0940	(-0.0070)
Sample #2 =	0.1070	(-0.0030)
Sample #3 =	0.1030	(-0.0110)
Sample #4 =	0.1010	(0.0080)

Avg % Abs = 0.1037 (-0.0020)
STD DEV = 0.0031 (0.0095)
REL STD DEV = 2.947 (476.970)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12827, 9um Io = 13650

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.8350	(0.0010)
Sample #2 =	0.8490	(0.0240)
Sample #3 =	0.8440	(0.0100)
Sample #4 =	0.8360	(0.0210)

Avg % Abs = 0.8430 (0.0183)
STD DEV = 0.0066 (0.0074)
REL STD DEV = 0.778 (40.206)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5380	(-0.0070)
Sample #2 =	1.5240	(0.0090)
Sample #3 =	1.5520	(-0.0070)
Sample #4 =	1.5450	(0.0000)

Avg % Abs = 1.5403 (0.0007)
STD DEV = 0.0146 (0.0080)
REL STD DEV = 0.946 (1203.121)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12819, 9um Io = 13646

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.9270	(-0.0160)
Sample #2 =	1.9240	(0.0180)
Sample #3 =	1.9540	(0.0010)
Sample #4 =	1.9180	(0.0080)

Avg % Abs = 1.9320 (0.0090)
STD DEV = 0.0193 (0.0085)
REL STD DEV = 0.998 (94.933)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.5900	(-0.0070)
Sample #2 =	3.5980	(0.0000)
Sample #3 =	3.6120	(-0.0140)
Sample #4 =	3.5750	(0.0000)

Avg % Abs = 3.5950 (-0.0047)
STD DEV = 0.0187 (0.0081)
REL STD DEV = 0.520 (173.205)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12815, 9um Io = 13641

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.6910	(-0.0160)
Sample #2 =	3.6730	(0.0020)
Sample #3 =	3.6780	(0.0410)
Sample #4 =	3.6610	(0.0450)

Avg % Abs = 3.6707 (0.0293)
STD DEV = 0.0087 (0.0238)
REL STD DEV = 0.238 (80.985)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	6.8760	(-0.0290)
Sample #2 =	6.8610	(0.0000)
Sample #3 =	6.8790	(0.0040)
Sample #4 =	6.8610	(0.0120)

Avg % Abs = 6.8670 (0.0053)
STD DEV = 0.0104 (0.0061)
REL STD DEV = 0.151 (114.564)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12809, 9um Io = 13642

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	5.3740	(-0.0190)
Sample #2 =	5.3550	(0.0050)
Sample #3 =	5.3710	(0.0310)
Sample #4 =	5.3660	(0.0150)

Avg % Abs = 5.3640 (0.0170)
STD DEV = 0.0082 (0.0131)
REL STD DEV = 0.153 (77.146)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	9.9830	(0.0020)
Sample #2 =	9.9560	(0.0390)
Sample #3 =	9.9570	(0.0480)
Sample #4 =	9.9620	(0.0430)

Avg % Abs = 9.9583 (0.0433)
STD DEV = 0.0032 (0.0045)
REL STD DEV = 0.032 (10.406)

80-000819
Cal Adj
12/05/2019
MK

***** AUTO CAL DATA *****

<<<<< CHANNEL 1 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.01 Rel Std Dev = 6.36
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.843
Std Dev = 0.01 Rel Std Dev = 0.78
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.932
Std Dev = 0.02 Rel Std Dev = 1.00
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.671
Std Dev = 0.01 Rel Std Dev = 0.24
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.364
Std Dev = 0.01 Rel Std Dev = 0.15
Zero Order Coef = -271.80
First Order Coef = 2555.71
Second Order Coef = 29.66
Standard Deviation = 10.870678

<<<<< CHANNEL 2 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.104
Std Dev = 0.00 Rel Std Dev = 2.95
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.540
Std Dev = 0.01 Rel Std Dev = 0.95
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.595
Std Dev = 0.02 Rel Std Dev = 0.52
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.867
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.958
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -142.17
First Order Coef = 1314.71
Second Order Coef = 13.47
Standard Deviation = 6.208895

Solution Stats Quadratic Fit Chan 1

Act g/210L	Fit g/210L	Residual g/210L
0.000	-0.000	0.0001
0.040	0.040	0.0000
0.100	0.100	-0.0003
0.200	0.200	0.0003
0.300	0.300	-0.0001

Solution Stats Quadratic Fit Chan 2

Act g/210L	Fit g/210L	Residual g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0001
0.200	0.200	0.0001
0.300	0.300	-0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1
Sample #1 = 3259.00
Sample #2 = 3291.00
Sample #3 = 3256.00
Sample #4 = 3196.00
Average Result = 3247.6667
STD DEV = 48.0451
REL STD DEV = 1.479

****

***** CHANNEL 2
Sample #1 = 3467.00
Sample #2 = 3444.00
Sample #3 = 3446.00
Sample #4 = 3446.00
Average Result = 3445.3333
STD DEV = 1.1547
REL STD DEV = 0.034

****

Dry Gas H20 Adjust Results **********
Barometric Pressure = 1020
3 um H20 Adjust (mg/l*10,000) = 562
9 um H20 Adjust (mg/l*10,000) = 364

**** AUTO CAL PASS

SP
BK 12/6/19

80-000819 Post Cal Stability Checks 12/05/2019 MK

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	0.050	09:26
Air Blank	0.000	09:27
Control Test	0.050	09:27
Air Blank	0.000	09:28
Control Test	0.049	09:29
Air Blank	0.000	09:29
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

MK
Operator's Signature

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:33
Control Test	0.079	09:34
Air Blank	0.000	09:34
Control Test	0.079	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

MK
Operator's Signature

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.201	09:39
Air Blank	0.000	09:40
Control Test	0.200	09:40
Air Blank	0.000	09:41
Control Test	0.201	09:42
Air Blank	0.000	09:42
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

MK
Operator's Signature

DGS

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
12/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.080	09:45
Air Blank	0.000	09:45
Control Test	0.079	09:46
Air Blank	0.000	09:46
Control Test	0.079	09:46
Air Blank	0.000	09:47
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

MK
Operator's Signature

SP
BK
12/6/19

Return Material Authorization

Ship to: ☐ CMI, Inc.

☑ Enforcement Electronics

Shipment to repair facility authorized by: Mary Bowermaster on 9/6/19

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: ____________

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000819

Bill To Address:	Ship to Address:
Citrus CountySO	FDLE-ATP Tallahassee

Reason for Return:

DSP Fail on initial startup and ambient fails

Please choose one of the following options:

☐ 1. I ____________________, authorize all repairs.

☐ 2. I ____________________, authorize repairs up to $________.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Mary Bowermaster

Phone #: 352-573-0240 Email: mary.bowermaster@corecivic.com

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Citrus County S/N 80-000819

Date In 7/30/2019 DI Completion Date ____ ❑Ship ❑P/U ❑H/D ❑CMI ❑EE

Intake Performed By DP

☑ Annual
❑ Registration
❑ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☑ Power cord ❑ Printer Cable
☑ Static Bag ❑ 12V DC Cable

Notes: ____

Final Release Date

Quality Checks Performed By SP

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value 2016
☒ Flow Verification (L/s)

Flow Column # ATP103

32 mm .156 (.139 - .169)
36 mm .175 (.156 - .190)
53 mm .242 (.228 - .278)
103 mm .496 (.447 - .547)

☒ Barometric Pressure Check
Gauge ID # 210932

❑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050		
0.080		
0.200		
0.080 DGS	N/A	

Flow Calibration Performed By ____

Flow Column # ____
❑ 5L/min – 17mm
❑ 15L/min – 53mm
❑ 30L/min – 103mm
❑ R-Value ____
❑ Post Calibration Verification (L/s)

Flow Column # ____

32 mm ____ (.139 - .169)
36 mm ____ (.156 - .190)
53 mm ____ (.228 - .278)
103 mm ____ (.447 - .547)

Maintenance Performed By ____

❑ Battery Replacement
❑ Dry Gas Regulator Replacement
❑ Breath Tube Replacement
❑ Other ____

Temperature Checks Performed By ____

❑ Lab Temp °C ____
External Digital Therm. ID#: ____
❑ 34°C +-.2 Serial #: ____
❑ 34°C +-.2 Serial #: ____
❑ 34°C +-.2 Serial #: ____

Calibration Adjustment Performed By ____

Barometric Pressure Gauge ____ ID # ____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

❑ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Notes/Suggested Service: ____

Department Inspection Performed By ____

Barometric Pressure ID# ____
Gauge ____ Instrument ____
Mouth Alcohol Solution Lot # ____
Acetone Stock Solution Lot # ____

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

❑ Form 41 ❑ Post-Stability Checks
❑ Stability Checks ❑ Flow Calibration
❑ Calibration Certificate ❑ Form 40
❑ Calibration Adjustment ❑ Other ____

❑ Instrument Complies with Chapter 11D-8, FAC
❑ Instrument Does Not Comply with Chapter 11D-8, FAC
❑ Return to/Place into Evidentiary Use
❑ Remain Out of Evidentiary Use
❑ Conduct an Agency Inspection Before Evidentiary Use

Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO
Time of Inspection: 09:55

Date of Inspection: 07/31/2019

Serial Number: 80-000819
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT

N/A COMPLIANCE NOT DETERMINED.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt SHAYLA D PLATT
Signature and Printed Name

07/31/2019
Date

FDLE/ATP Form 40 – March 2004

CITRUS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000819
07/31/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	09:56
Air Blank	AMB*	09:57

*Ambient Fail

SP

Operator's Signature

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO
Time of Inspection: 09:50
Date of Inspection: 12/04/2019
Serial Number: 80-000819
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:
BY-PASSED AI.COMPLIANCE NOT DETERMINED.

SP
BK
12/6/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael D Haughey MICHAEL D HAUGHEY
Signature and Printed Name

12/04/2019
Date

FDLE/ATP Form 40 – March 2004