



# INSTRUMENT PROCESSING SHEET

Agency Mount Dora PDS/N 80-000825

Florida Department of Law Enforcement

Date In 1/15/2019DI Completion Date 1/15/2019 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>JA</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>JA</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>210</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	DR3856	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD1012	201707D 07/25/2019																																																											
0.080	DR1279	201707E 07/25/2019																																																											
0.200	DR3856	201707C 07/24/2019																																																											
0.080 DGS	N/A	AG805701 02/26/2020																																																											
<b>Final Release Date</b>  <div style="text-align: center;"> <b>FDLE</b>   <b>JAN 16 2019</b>   <b>Alcohol Testing Program</b> </div>	<b>Maintenance</b> Performed By <u>JA</u> <input checked="" type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>JA</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.6</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>DR3856</u>																																																												
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>JA</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1023</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	DR3856
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	G2408																																																												
Interferent	G2882																																																												
0.050	SD1012																																																												
0.080	DR1279																																																												
0.200	DR3856																																																												
Notes/Suggested Service: <u>Please change level 2 password to something unique.</u> _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: center;"> <u>8P 1/15/19</u> <u>Brett Kirkland 1/16/19</u>          Tech Review / Date                      Admin Review / Date       </div>																																																												

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MOUNT DORA P.D.  
Time of Inspection: 14:08

Date of Inspection: 01/15/2019

Serial Number: 80-000825  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.051	0.081	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.079
0.000	0.051	0.080	0.199	0.078
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.080	0.199	0.079

Standard Deviations	0.0004	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP BK  
1/16/19

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

THOMAS J GRAHAM

01/15/2019  
 Date



80-000825  
 11/15/19  
 JD

MOUNT DORA P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000825  
 01/15/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:36
Control Test	0.049	11:36
Air Blank	0.000	11:37
Control Test	0.049	11:38
Air Blank	0.000	11:38
Control Test	0.048	11:39
Air Blank	0.000	11:40
Control Test Stats		
Average	0.0487	
Std Dev	0.0005	
Rel Std Dev(%)	1.1863	

JD  
 Operator's Signature

MOUNT DORA P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000825  
 01/15/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:41
Control Test	0.078	11:42
Air Blank	0.000	11:43
Control Test	0.078	11:43
Air Blank	0.000	11:44
Control Test	0.079	11:44
Air Blank	0.000	11:45
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

JD  
 Operator's Signature

MOUNT DORA P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000825  
 01/15/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:46
Control Test	0.196	11:47
Air Blank	0.000	11:47
Control Test	0.196	11:48
Air Blank	0.000	11:49
Control Test	0.197	11:49
Air Blank	0.000	11:50
Control Test Stats		
Average	0.1970	
Std Dev	0.0010	
Rel Std Dev(%)	0.5076	

JD  
 Operator's Signature

MOUNT DORA P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000825  
 01/15/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:04
Control Test	0.079	12:04
Air Blank	0.000	12:04
Control Test	0.079	12:05
Air Blank	0.000	12:05
Control Test	0.079	12:06
Air Blank	0.000	12:06
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

JD 11/15/19  
 JD  
 Operator's Signature

MOUNT DORA P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000825  
 01/15/2019  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:31
Control Test	INT*	11:31
Air Blank	0.000	11:32
Control Test	INT*	11:33
Air Blank	0.000	11:33
Control Test	INT*	11:34
Air Blank	0.000	11:34
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Acetone  
 Simulator  
 Test

\*Interferent Detect

Operator's Signature  
 JD

SP  
 BX  
 11/16/19

INTOXILYZER 8000  
 Instrument Initialization  
 11:15 01/15/2019



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000825, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000825</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MOUNT DORA P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/15/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:08</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

SP  
Bk  
1/16/19

01/15/2019

Date

Thomas J Graham  
**THOMAS J GRAHAM,**  
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality