



# INSTRUMENT PROCESSING SHEET

Agency Martin County SOS/N 80-000831Florida Department of  
Law EnforcementDate In 02/15/2019DI Completion Date 2/20/19 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>JA</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>120</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.149</u> (.139 - .169) 36 mm <u>.183</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
<b>Final Release Date</b> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FEB 20 2019</div> <div style="text-align: center; font-weight: bold;">Alcohol Testing Program</div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805701</u> <u>2-26-20</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	201707D 07/25/2019	0.080	<u>DR1279</u>	201707E 07/25/2019	0.200	<u>DR3856</u>	201707C 07/24/2019	0.080 DGS	N/A	<u>AG805701</u> <u>2-26-20</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>JA</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.6</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR3856</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1012</u>	201707D 07/25/2019																																																											
0.080	<u>DR1279</u>	201707E 07/25/2019																																																											
0.200	<u>DR3856</u>	201707C 07/24/2019																																																											
0.080 DGS	N/A	<u>AG805701</u> <u>2-26-20</u>																																																											
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1017</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2408</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2882</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1012</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2408</u>	Interferent	<u>G2882</u>	0.050	<u>SD1012</u>	0.080	<u>DR1279</u>	0.200	<u>DR3856</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	<u>G2408</u>																																																												
Interferent	<u>G2882</u>																																																												
0.050	<u>SD1012</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>DR3856</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td> <input checked="" type="checkbox"/> Form 41  <input checked="" type="checkbox"/> Stability Checks  <input checked="" type="checkbox"/> Calibration Certificate  <input type="checkbox"/> Calibration Adjustment         </td> <td> <input type="checkbox"/> Post-Stability Checks  <input type="checkbox"/> Flow Calibration  <input type="checkbox"/> Form 40  <input type="checkbox"/> Other _____         </td> </tr> </table> <div style="border: 1px solid black; padding: 5px;"> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC  <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC  <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use  <input type="checkbox"/> Remain Out of Evidentiary Use  <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use         </div> <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> <u>Scott Kirkland 2/20/19</u>              Tech Review / Date           </div> <div style="text-align: center;"> <u>J. [Signature]</u>              Admin Review / Date           </div> </div>		<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																									
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																												

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO  
Time of Inspection: 10:58

Date of Inspection: 02/20/2019

Serial Number: 80-000831  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.078	0.196	0.077
0.000	0.049	0.079	0.197	0.077
0.000	0.048	0.079	0.197	0.077
0.000	0.049	0.079	0.196	0.077
0.000	0.048	0.079	0.197	0.077
0.000	0.049	0.079	0.196	0.077
0.000	0.049	0.079	0.197	0.077
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.197	0.077
0.000	0.049	0.079	0.197	0.077

Standard Deviations	0.0004	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

02/20/2019  
Date

*2/20/19 JB*      *JBK 2/20/19*

# STABILITY CHECKS #80-000831

MARTIN COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000831  
02/20/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:39
Control Test	0.049	08:40
Air Blank	0.000	08:41
Control Test	0.049	08:41
Air Blank	0.000	08:42
Control Test	0.049	08:42
Air Blank	0.000	08:43
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*SP*  
Operator's Signature

*SP*  
*5/10/19*

MARTIN COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000831  
02/20/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.077	08:52
Air Blank	0.000	08:52
Control Test	0.078	08:53
Air Blank	0.000	08:54
Control Test	0.078	08:54
Air Blank	0.000	08:55
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

*SP*  
Operator's Signature

MARTIN COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000831  
02/20/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.195	08:46
Air Blank	0.000	08:46
Control Test	0.195	08:47
Air Blank	0.000	08:48
Control Test	0.195	08:48
Air Blank	0.000	08:49
Control Test Stats		
Average	0.1950	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*SP*  
Operator's Signature

MARTIN COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000831  
02/20/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.078	08:56
Air Blank	0.000	08:57
Control Test	0.077	08:57
Air Blank	0.000	08:58
Control Test	0.078	08:58
Air Blank	0.000	08:59
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

*DGS*

*SP*  
Operator's Signature

*SP*  
*2/20/19*



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000831, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000831</u>	UNCERTAINTY * ±
Owning Agency:	<u>MARTIN COUNTY SO</u>	0.050 g/ 210 L
Calibration Date:	<u>02/20/2019</u>	0.080 g/ 210 L
Calibration Time:	<u>10:58</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/20/2019

Date

*Shayla D Platt*

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*2/20/19*

*BK 2/20/19*