



INSTRUMENT PROCESSING SHEET

Agency Lynn Haven PD

S/N 80-000840

Florida Department of Law Enforcement

Date In 02/19/2019

DI Completion Date 2/26/19

Ship P/U H/D CMI EE

Intake Performed By <u>JD</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____	Quality Checks Performed By <u>PSM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>175</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.476</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)															
Final Release Date FDLE FEB 26 2019 Alcohol Testing Program	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1021</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1275</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804 11/14/20</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1021	201707D 07/25/2019	0.080	DR1275	201707E 07/25/2019	0.200	SD1013	201707C 07/24/2019	0.080 DGS	N/A	AG831804 11/14/20	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>PSM</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.0</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1021</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1275</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1013</u>
Simulator	Serial #	Lot #/Exp															
0.050	SD1021	201707D 07/25/2019															
0.080	DR1275	201707E 07/25/2019															
0.200	SD1013	201707C 07/24/2019															
0.080 DGS	N/A	AG831804 11/14/20															

Calibration Adjustment Performed By <u>PSM</u> Barometric Pressure Gauge <u>1019</u> ID # <u>28421</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>62834</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1022</td> <td>17410</td> <td>12/6/19</td> </tr> <tr> <td>0.100</td> <td>SD3964</td> <td>18070</td> <td>2/26/20</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>19040</td> <td>1/29/21</td> </tr> <tr> <td>0.300</td> <td>SD1024</td> <td>12110</td> <td>4/2/20</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A</td> <td>8/5/19</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1021</td> <td>201707D</td> <td>7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1275</td> <td>201707E</td> <td>7/25/19</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C</td> <td>7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804</td> <td>11/14/20</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	62834	N/A	N/A	0.040	SD1022	17410	12/6/19	0.100	SD3964	18070	2/26/20	0.200	SD1025	19040	1/29/21	0.300	SD1024	12110	4/2/20	0.080 DGS	N/A	17817080A	8/5/19	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1021	201707D	7/25/19	0.080	DR1275	201707E	7/25/19	0.200	SD1013	201707C	7/24/19	0.080 DGS	N/A	AG831804	11/14/20	Department Inspection Performed By <u>PSM</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1020</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G11621</td> </tr> <tr> <td>Interferent</td> <td>DR3855</td> </tr> <tr> <td>0.050</td> <td>SD1021</td> </tr> <tr> <td>0.080</td> <td>DR1275</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G11621	Interferent	DR3855	0.050	SD1021	0.080	DR1275	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	62834	N/A	N/A																																																										
0.040	SD1022	17410	12/6/19																																																										
0.100	SD3964	18070	2/26/20																																																										
0.200	SD1025	19040	1/29/21																																																										
0.300	SD1024	12110	4/2/20																																																										
0.080 DGS	N/A	17817080A	8/5/19																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD1021	201707D	7/25/19																																																										
0.080	DR1275	201707E	7/25/19																																																										
0.200	SD1013	201707C	7/24/19																																																										
0.080 DGS	N/A	AG831804	11/14/20																																																										
Simulator	Serial Number																																																												
0.000	G11621																																																												
Interferent	DR3855																																																												
0.050	SD1021																																																												
0.080	DR1275																																																												
0.200	SD1013																																																												
Notes/Suggested Service: <u>Calibrated adjusted to bring values closer to nominal.</u> <u>DI completion date amended at time of technical review 2/26/19</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>JD 2/26/19</u> <u>Britt Kirkland 2/26/19</u> Tech Review / Date Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LYNN HAVEN PD
Time of Inspection: 11:11

Date of Inspection: 02/22/2019

Serial Number: 80-000840
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.048	0.080	0.200	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.049	0.081	0.202	0.079
0.000	0.049	0.081	0.202	0.079
0.000	0.049	0.081	0.202	0.080
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.081	0.203	0.079
0.000	0.050	0.081	0.203	0.079

Standard Deviations	0.0006	0.0004	0.0008	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

02/22/2019
Date

*2/22/19
JK*

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/19/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:23
Control Test	0.051	08:24
Air Blank	0.000	08:24
Control Test	0.050	08:25
Air Blank	0.000	08:25
Control Test	0.050	08:26
Air Blank	0.000	08:27
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

P Murphy

Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/19/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:28
Control Test	0.082	08:29
Air Blank	0.000	08:30
Control Test	0.083	08:30
Air Blank	0.000	08:31
Control Test	0.083	08:31
Air Blank	0.000	08:32
Control Test Stats		
Average	0.0827	
Std Dev	0.0006	
Rel Std Dev(%)	0.6984	

P Murphy

Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/19/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:33
Control Test	0.202	08:34
Air Blank	0.000	08:35
Control Test	0.202	08:35
Air Blank	0.000	08:36
Control Test	0.202	08:36
Air Blank	0.000	08:37
Control Test Stats		
Average	0.2020	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/19/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.077	08:39
Air Blank	0.000	08:39
Control Test	0.076	08:39
Air Blank	0.000	08:40
Control Test	0.077	08:40
Air Blank	0.000	08:41
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

DGS

P Murphy

Operator's Signature

2/26/19
TSK



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000840, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000840</u>	UNCERTAINTY* ±
Owning Agency:	<u>LYNN HAVEN PD</u>	0.050 g/ 210 L
Calibration Date:	<u>02/22/2019</u>	0.080 g/ 210 L
Calibration Time:	<u>11:11</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/22/2019

Date

Patrick J. Murphy
PATRICK J MURPHY,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

2/26/19
JBJ
BSK

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/22/2019 08:03:51

Auto Calibration
Max Power Res Value = 17
Auto Range Res Value = 13

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12824, 9um Io = 12982

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1020 (-0.0250)
Sample #2 = 0.0830 (0.0100)
Sample #3 = 0.1020 (0.0220)
Sample #4 = 0.1000 (0.0160)
Avg % Abs = 0.0950 (0.0160)
STD DEU = 0.0104 (0.0060)
REL STD DEU = 10.990 (37.500)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1480 (-0.0250)
Sample #2 = 0.1270 (0.0030)
Sample #3 = 0.1540 (0.0000)
Sample #4 = 0.1520 (-0.0060)
Avg % Abs = 0.1443 (-0.0010)
STD DEU = 0.0150 (0.0046)
REL STD DEU = 10.423 (458.258)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12819, 9um Io = 12980

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7980 (-0.0180)
Sample #2 = 0.7920 (0.0030)
Sample #3 = 0.7930 (0.0330)
Sample #4 = 0.7830 (0.0200)
Avg % Abs = 0.7893 (0.0187)
STD DEU = 0.0055 (0.0150)
REL STD DEU = 0.698 (80.595)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5470 (-0.0120)
Sample #2 = 1.5520 (-0.0040)
Sample #3 = 1.5760 (0.0080)
Sample #4 = 1.5590 (0.0200)
Avg % Abs = 1.5623 (0.0080)
STD DEU = 0.0123 (0.0120)
REL STD DEU = 0.790 (150.000)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12815, 9um Io = 12977

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8140 (-0.0120)
Sample #2 = 1.8370 (0.0200)
Sample #3 = 1.8350 (-0.0020)
Sample #4 = 1.8100 (0.0230)
Avg % Abs = 1.8273 (0.0137)
STD DEU = 0.0150 (0.0137)
REL STD DEU = 0.823 (99.881)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6010 (-0.0030)
Sample #2 = 3.6110 (0.0120)
Sample #3 = 3.6220 (-0.0040)
Sample #4 = 3.6270 (0.0040)
Avg % Abs = 3.6200 (0.0040)
STD DEU = 0.0082 (0.0080)
REL STD DEU = 0.226 (200.000)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12808, 9um Io = 12972

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5250 (-0.0260)
Sample #2 = 3.5180 (-0.0340)
Sample #3 = 3.5350 (0.0110)
Sample #4 = 3.5120 (-0.0050)
Avg % Abs = 3.5217 (-0.0093)
STD DEU = 0.0119 (0.0228)
REL STD DEU = 0.339 (244.402)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8550 (-0.0250)
Sample #2 = 6.9010 (-0.0360)
Sample #3 = 6.9070 (-0.0160)
Sample #4 = 6.8970 (-0.0110)
Avg % Abs = 6.9017 (-0.0210)
STD DEU = 0.0050 (0.0132)
REL STD DEU = 0.073 (62.994)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12805, 9um Io = 12973

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.0580 (-0.0140)
Sample #2 = 5.0730 (0.0000)
Sample #3 = 5.0750 (-0.0070)
Sample #4 = 5.0760 (0.0150)
Avg % Abs = 5.0747 (0.0027)
STD DEU = 0.0015 (0.0112)
REL STD DEU = 0.030 (421.493)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.8680 (-0.0070)
Sample #2 = 9.8970 (-0.0030)
Sample #3 = 9.9350 (-0.0130)
Sample #4 = 9.9260 (0.0120)
Avg % Abs = 9.9193 (-0.0013)
STD DEU = 0.0199 (0.0126)
REL STD DEU = 0.200 (943.729)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.095
Std Dev = 0.01 Rel Std Dev = 10.99
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.789
Std Dev = 0.01 Rel Std Dev = 0.70
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.827
Std Dev = 0.02 Rel Std Dev = 0.82
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.522
Std Dev = 0.01 Rel Std Dev = 0.34
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.075
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -218.49
First Order Coef = 2619.85
Second Order Coef = 46.13
Standard Deviation = 41.162186

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.144
Std Dev = 0.02 Rel Std Dev = 10.42
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.562
Std Dev = 0.01 Rel Std Dev = 0.79
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.620
Std Dev = 0.01 Rel Std Dev = 0.23
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.902
Std Dev = 0.01 Rel Std Dev = 0.07
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.919
Std Dev = 0.02 Rel Std Dev = 0.20
Zero Order Coef = -175.31
First Order Coef = 1301.12
Second Order Coef = 15.69
Standard Deviation = 20.136606

2/22/19
20
JDK

80-000840

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0006
0.040	0.039	0.0006
0.100	0.099	0.0008
0.200	0.201	-0.0012
0.300	0.300	0.0004

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0002
0.100	0.100	0.0005
0.200	0.201	-0.0006
0.300	0.300	0.0002

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1
Sample #1 = 3074.00
Sample #2 = 3029.00
Sample #3 = 2942.00
Sample #4 = 2918.00
Average Result = 2963.0000
STD DEU = 58.4038
REL STD DEU = 1.971

***** CHANNEL 2
Sample #1 = 3310.00
Sample #2 = 3273.00
Sample #3 = 3281.00
Sample #4 = 3262.00
Average Result = 3272.0000
STD DEU = 9.5394
REL STD DEU = 0.292

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
3 um H2O Adjust (mg/l*10,000) = 846
9 um H2O Adjust (mg/l*10,000) = 537
**** AUTO CAL PASS

2/26/19
JE
BK

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.048	08:52
Air Blank	0.000	08:52
Control Test	0.049	08:53
Air Blank	0.000	08:54
Control Test	0.048	08:54
Air Blank	0.000	08:55
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

P Murphy
Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.079	08:57
Air Blank	0.000	08:57
Control Test	0.080	08:58
Air Blank	0.000	08:58
Control Test	0.080	08:59
Air Blank	0.000	09:00
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

P Murphy
Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.197	09:02
Air Blank	0.000	09:03
Control Test	0.199	09:04
Air Blank	0.000	09:04
Control Test	0.200	09:05
Air Blank	0.000	09:05
Control Test Stats		
Average	0.1987	
Std Dev	0.0015	
Rel Std Dev(%)	0.7689	

P Murphy
Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
02/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:07
Control Test	0.080	09:07
Air Blank	0.000	09:08
Control Test	0.080	09:08
Air Blank	0.000	09:09
Control Test	0.080	09:09
Air Blank	0.000	09:09
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P Murphy
Operator's Signature

POST CALIBRATION ADJUST
STABILITIES

2/26/19
JK
BK