



INSTRUMENT PROCESSING SHEET

Agency Collier County Sheriff's Office

S/N 80-000942

Florida Department of Law Enforcement

Date In 01/03/2019 DI Completion Date 01/03/2019

Ship P/U H/D CMI EE

Intake Performed By <u>Deer</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>Deer</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>218</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.488</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date <p style="text-align: center;">FDLE</p> <p style="text-align: center;">JAN 14 2019</p> <p style="text-align: center;">Alcohol Testing Program</p>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Maintenance Performed By <u>Deer</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Form Load/Pass Changed</u>																								
Simulator	Serial #	Lot #/Exp																																									
0.050	SD3967	201707D 07/25/2019																																									
0.080	SD3968	201707E 07/25/2019																																									
0.200	SD3969	201707C 07/24/2019																																									
0.080 DGS	N/A	AG805701 02/26/2020																																									
Calibration Adjustment Performed By <u>Deer</u> Barometric Pressure Gauge <u>1019</u> ID # <u>28663</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>2235</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>2108</td> <td>17410</td> <td>12/06/2019</td> </tr> <tr> <td>0.100</td> <td>MP4863</td> <td>18070</td> <td>02/26/2020</td> </tr> <tr> <td>0.200</td> <td>MP4864</td> <td>17340</td> <td>10/09/2019</td> </tr> <tr> <td>0.300</td> <td>SD3963</td> <td>18110</td> <td>04/02/2020</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>34416080A2</td> <td>02/05/2019</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks		Simulator	Serial Number	Lot Number	Expiration	0.000	2235	N/A	N/A	0.040	2108	17410	12/06/2019	0.100	MP4863	18070	02/26/2020	0.200	MP4864	17340	10/09/2019	0.300	SD3963	18110	04/02/2020	0.080 DGS	N/A	34416080A2	02/05/2019	Department Inspection Performed By <u>Deer</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1018</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	2235	N/A	N/A																																								
0.040	2108	17410	12/06/2019																																								
0.100	MP4863	18070	02/26/2020																																								
0.200	MP4864	17340	10/09/2019																																								
0.300	SD3963	18110	04/02/2020																																								
0.080 DGS	N/A	34416080A2	02/05/2019																																								
Simulator	Serial Number																																										
0.000	SD3965																																										
Interferent	SD3966																																										
0.050	SD3967																																										
0.080	SD3968																																										
0.200	SD3969																																										
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701</td> <td>02/26/2020</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050	SD3967	201707D	07/25/2019	0.080	SD3968	201707E	07/25/2019	0.200	SD3969	201707C	07/24/2019	0.080 DGS	N/A	AG805701	02/26/2020	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																								
0.050	SD3967	201707D	07/25/2019																																								
0.080	SD3968	201707E	07/25/2019																																								
0.200	SD3969	201707C	07/24/2019																																								
0.080 DGS	N/A	AG805701	02/26/2020																																								
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/>		<p style="text-align: center; color: red; font-weight: bold; font-size: 2em;">APPROVED</p> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
_____ _____ _____ _____		<p style="text-align: center;"> <u>P. P. P. 1/14/19</u> <u>[Signature]</u> 1/14/19 Tech Review / Date Admin Review / Date </p>																																									

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 14:34

Date of Inspection: 01/03/2019

Serial Number: 80-000942
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.197	0.078
0.000	0.050	0.081	0.199	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.077
0.000	0.050	0.081	0.200	0.077
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.080	0.200	0.078

Standard Deviations	0.0003	0.0004	0.0009	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

Pgm

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E. Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

01/03/2019
Date

*1/14/19
JO*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000942	Collier County Sheriff's Office	01/03/2019	<i>Jerr</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.08c 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.08c 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.09c 0.194 to 0.206 <input checked="" type="checkbox"/>	Lot AG805701 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000942 01/03/2019 Software: 8100.27	COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000942 01/03/2019 Software: 8100.27	COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000942 01/03/2019 Software: 8100.27	COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000942 01/03/2019 Software: 8100.27																																																																																																																																																
<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	10:39	Control Test	0.050	10:40	Air Blank	0.000	10:41	Control Test	0.050	10:41	Air Blank	0.000	10:42	Control Test	0.050	10:42	Air Blank	0.000	10:43	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0820</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	10:51	Control Test	0.082	10:51	Air Blank	0.000	10:52	Control Test	0.082	10:53	Air Blank	0.000	10:53	Control Test	0.082	10:54	Air Blank	0.000	10:54	Control Test Stats			Average	0.0820		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.201</td><td>10:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr> <tr><td>Control Test</td><td>0.201</td><td>10:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr> <tr><td>Control Test</td><td>0.201</td><td>10:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2010</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	10:56	Control Test	0.201	10:57	Air Blank	0.000	10:57	Control Test	0.201	10:58	Air Blank	0.000	10:59	Control Test	0.201	10:59	Air Blank	0.000	11:00	Control Test Stats			Average	0.2010		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	11:02	Control Test	0.078	11:02	Air Blank	0.000	11:03	Control Test	0.078	11:03	Air Blank	0.000	11:03	Control Test	0.079	11:04	Air Blank	0.000	11:04	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:39																																																																																																																																																	
Control Test	0.050	10:40																																																																																																																																																	
Air Blank	0.000	10:41																																																																																																																																																	
Control Test	0.050	10:41																																																																																																																																																	
Air Blank	0.000	10:42																																																																																																																																																	
Control Test	0.050	10:42																																																																																																																																																	
Air Blank	0.000	10:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:51																																																																																																																																																	
Control Test	0.082	10:51																																																																																																																																																	
Air Blank	0.000	10:52																																																																																																																																																	
Control Test	0.082	10:53																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test	0.082	10:54																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0820																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test	0.201	10:57																																																																																																																																																	
Air Blank	0.000	10:57																																																																																																																																																	
Control Test	0.201	10:58																																																																																																																																																	
Air Blank	0.000	10:59																																																																																																																																																	
Control Test	0.201	10:59																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2010																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.078	11:02																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.078	11:03																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.079	11:04																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
<i>Full</i> Operator's Signature <i>1/4/19</i>	<i>Full</i> Operator's Signature	<i>Jerr</i> Operator's Signature	<i>Jerr</i> Operator's Signature																																																																																																																																																

4800



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000942, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000942</u>	UNCERTAINTY* ±	
Owning Agency:	<u>COLLIER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/03/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:34</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/03/2019

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

1/4/19

WSP

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5490 (-0.0070)
Sample #2 = 1.5160 (0.0050)
Sample #3 = 1.5160 (0.0120)
Sample #4 = 1.5560 (0.0000)
Avg % Abs = 1.5293 (0.0057)
STD DEV = 0.0231 (0.0060)
REL STD DEV = 1.510 (106.371)

COLLIER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2019
SN 80-000942
11:28:25

Auto Calibration
Max Power Res Value = 32
Auto Range Res Value = 16

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12640, Sum Io = 14260
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1380 (-0.0200)
Sample #2 = 0.1090 (-0.0050)
Sample #3 = 0.1360 (0.0030)
Sample #4 = 0.1100 (0.0160)
Avg % Abs = 0.1183 (0.0047)
STD DEV = 0.0153 (0.0106)
REL STD DEV = 12.936 (227.116)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7310 (-0.0180)
Sample #2 = 3.7120 (0.0110)
Sample #3 = 3.7230 (0.0070)
Sample #4 = 3.7360 (0.0190)
Avg % Abs = 3.7237 (0.0123)
STD DEV = 0.0120 (0.0061)
REL STD DEV = 0.323 (49.541)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12633, Sum Io = 14259
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.8350 (-0.0240)
Sample #2 = 3.8400 (-0.0090)
Sample #3 = 3.8140 (0.0170)
Sample #4 = 3.8600 (-0.0060)
Avg % Abs = 3.8380 (0.0007)
STD DEV = 0.0231 (0.0142)
REL STD DEV = 0.601 (2133.657)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12638, Sum Io = 14263
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8800 (-0.0250)
Sample #2 = 0.8660 (-0.0100)
Sample #3 = 0.8790 (0.0200)
Sample #4 = 0.8630 (0.0230)
Avg % Abs = 0.8693 (0.0110)
STD DEV = 0.0085 (0.0182)
REL STD DEV = 0.978 (165.894)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.118
Std Dev = 0.02 Rel Std Dev = 12.94
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.869
Std Dev = 0.01 Rel Std Dev = 0.98
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.028
Std Dev = 0.01 Rel Std Dev = 0.30
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.838
Std Dev = 0.02 Rel Std Dev = 0.60
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.585
Std Dev = 0.01 Rel Std Dev = 0.26
Zero Order Coef = -269.02
First Order Coef = 2428.73
Second Order Coef = 31.70
Standard Deviation = 24.654371
<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.078
Std Dev = 0.01 Rel Std Dev = 11.58
Sol Val = 0.1915 mg/l or 0.040 g/210L
% Abs = 1.529
Std Dev = 0.02 Rel Std Dev = 1.51
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.724
Std Dev = 0.01 Rel Std Dev = 0.32
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.095
Std Dev = 0.01 Rel Std Dev = 0.13
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.306
Std Dev = 0.01 Rel Std Dev = 0.09
Zero Order Coef = -85.55
First Order Coef = 1260.71
Second Order Coef = 12.99
Standard Deviation = 22.678667

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12637, Sum Io = 14258
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.6180 (-0.0030)
Sample #2 = 5.5960 (0.0220)
Sample #3 = 5.5690 (0.0570)
Sample #4 = 5.5910 (0.0660)
Avg % Abs = 5.5853 (0.0483)
STD DEV = 0.0144 (0.0232)
REL STD DEV = 0.257 (48.093)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.3610 (-0.0230)
Sample #2 = 10.3160 (0.0500)
Sample #3 = 10.2980 (0.0610)
Sample #4 = 10.3050 (0.0700)
Avg % Abs = 10.3063 (0.0603)
STD DEV = 0.0091 (0.0100)
REL STD DEV = 0.088 (16.602)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12635, Sum Io = 14260
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0090 (-0.0150)
Sample #2 = 2.0220 (0.0120)
Sample #3 = 2.0270 (-0.0010)
Sample #4 = 2.0340 (0.0200)
Avg % Abs = 2.0277 (0.0103)
STD DEV = 0.0060 (0.0106)
REL STD DEV = 0.297 (102.568)

Optical Calibration	
SN: 80-000942	
Agency: Collier County SO	
Date: 01/03/2019	
Quadratic Fit: +/- 0.002g/210L	
By: <i>Jell</i>	

1/4/19
Jell

Comp

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12640, Sum Io = 14260
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3110.00
Sample #2 = 3235.00
Sample #3 = 3105.00
Sample #4 = 3204.00
Average Result = 3181.3333
STD DEV = 67.8994
REL STD DEV = 2.134

***** CHANNEL 2 *****
Sample #1 = 3523.00
Sample #2 = 3534.00
Sample #3 = 3523.00
Sample #4 = 3532.00
Average Result = 3529.6667
STD DEV = 5.8595
REL STD DEV = 0.166

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1017
3 um H2O Adjust (mg/l*10,000) = 628
9 um H2O Adjust (mg/l*10,000) = 280
***** AUTO CAL PASS *****

Lab 80

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-000942	Collier County Sheriff's Office	01/03/2019	Deek

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.08c 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.08c 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.09c 0.194 to 0.206 <input checked="" type="checkbox"/>	Lot AG805701 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000942 01/03/2019 Software: 8100.27	COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000942 01/03/2019 Software: 8100.27	COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000942 01/03/2019 Software: 8100.27	COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000942 01/03/2019 Software: 8100.27																																																																																																																																																
<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	12:37	Control Test	0.049	12:38	Air Blank	0.000	12:38	Control Test	0.050	12:39	Air Blank	0.000	12:39	Control Test	0.050	12:40	Air Blank	0.000	12:41	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.082</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	12:42	Control Test	0.081	12:43	Air Blank	0.000	12:44	Control Test	0.081	12:44	Air Blank	0.000	12:45	Control Test	0.082	12:45	Air Blank	0.000	12:46	Control Test Stats			Average	0.0813		Std Dev	0.0005		Rel Std Dev(%)	0.7099		<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7689</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	12:47	Control Test	0.197	12:48	Air Blank	0.000	12:48	Control Test	0.199	12:49	Air Blank	0.000	12:49	Control Test	0.200	12:50	Air Blank	0.000	12:51	Control Test Stats			Average	0.1987		Std Dev	0.0015		Rel Std Dev(%)	0.7689		<table border="0"> <tr><td>Test</td><td>g/210L</td><td>Time</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	12:52	Control Test	0.078	12:52	Air Blank	0.000	12:53	Control Test	0.079	12:53	Air Blank	0.000	12:54	Control Test	0.079	12:54	Air Blank	0.000	12:54	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test	0.049	12:38																																																																																																																																																	
Air Blank	0.000	12:38																																																																																																																																																	
Control Test	0.050	12:39																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test	0.050	12:40																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test	0.081	12:43																																																																																																																																																	
Air Blank	0.000	12:44																																																																																																																																																	
Control Test	0.081	12:44																																																																																																																																																	
Air Blank	0.000	12:45																																																																																																																																																	
Control Test	0.082	12:45																																																																																																																																																	
Air Blank	0.000	12:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0813																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel Std Dev(%)	0.7099																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:47																																																																																																																																																	
Control Test	0.197	12:48																																																																																																																																																	
Air Blank	0.000	12:48																																																																																																																																																	
Control Test	0.199	12:49																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.200	12:50																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7689																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test	0.078	12:52																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.079	12:53																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.079	12:54																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Deek Operator's Signature	Deek Operator's Signature	Deek Operator's Signature	Deek Operator's Signature																																																																																																																																																

2/14/19