



INSTRUMENT PROCESSING SHEET

Agency Osceola County SO corrections SK S/N 80-000968

Florida Department of Law Enforcement Date In 01/11/19 DI Completion Date 1/14/19 Ship P/U H/D CMI EE

Intake Performed By SQC Quality Checks Performed By JB Flow Calibration Performed By Maintenance Performed By Temperature Checks Performed By

Calibration Adjustment Performed By Department Inspection Performed By

Notes/Suggested Service: Change level 2 password to something unique. JB

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 11:26

Date of Inspection: 01/14/2019

Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.078	0.195	0.078
0.000	0.048	0.079	0.196	0.079
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.079	0.195	0.078
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.079	0.196	0.078
0.000	0.048	0.079	0.196	0.079
0.000	0.048	0.079	0.196	0.078
0.000	0.048	0.079	0.196	0.078
0.000	0.048	0.078	0.196	0.078

Standard Deviations	0.0000	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

Graham
BK
1/14/19
JK

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

THOMAS J GRAHAM

01/14/2019
 Date

80-000968

1/14/19 JD

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer SN 80-000968
 Model 8000
 01/14/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:35
Control Test	0.049	08:36
Air Blank	0.000	08:36
Control Test	0.049	08:37
Air Blank	0.000	08:38
Control Test	0.049	08:38
Air Blank	0.000	08:39
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD

Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer SN 80-000968
 Model 8000
 01/14/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.079	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:43
Control Test	0.078	08:44
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

JD

Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer SN 80-000968
 Model 8000
 01/14/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.196	08:31
Air Blank	0.000	08:32
Control Test	0.196	08:32
Air Blank	0.000	08:33
Control Test	0.196	08:34
Air Blank	0.000	08:34
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD

Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer SN 80-000968
 Model 8000
 01/14/2019
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Control Test	0.079	08:47
Air Blank	0.000	08:48
Control Test	0.079	08:48
Air Blank	0.000	08:49
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

JD

Operator's Signature

JPAM
 1/14/19
 157



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000968, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000968</u>	UNCERTAINTY* ±	
Owning Agency:	<u>OSCEOLA CORRECTIONS</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/14/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:26</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).


TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/14/2019 Date

THOMAS J GRAHAM,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Handwritten notes:
PBM
73K
1/14/19
2K