



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop E

S/N 80-001121

Florida Department of Law Enforcement

Date In 03/01/2019 DI Completion Date 03/04/2019

Ship P/U H/D CMI EE

Intake, Quality Checks, Flow Calibration, Maintenance sections with checkboxes and data entry fields.

Final Release Date FDLE MAR 11 2019 Alcohol Testing Program

Table with columns: Simulator, Serial #, Lot #/Exp. Rows include SD3967, SD3968, SD3969, and AG805702.

Calibration Adjustment section with Barometric Pressure Gauge 1015 ID # 28199 and Post Calibration Adjustment Stability Checks table.

Department Inspection section with Barometric Pressure ID# 68639 and a table of Simulators and Serial Numbers.

Attachments section with checkboxes for Form 41, Stability Checks, Calibration Certificate, Calibration Adjustment, Post-Stability Checks, Flow Calibration, Form 40, and Other.

Notes/Suggested Service: E-mailed Calibration adjustment to bring values closer to Nominal

Instrument Compliance section with checkboxes for Chapter 11D-8, FAC, Return to/Place into Evidentiary Use, and Conduct an Agency Inspection Before Evidentiary Use. Includes Tech Review and Admin Review dates.

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 12:59

Date of Inspection: 03/04/2019

Serial Number: 80-001121
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.080	0.196	0.078
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.078
0.000	0.048	0.080	0.197	0.079
0.000	0.048	0.080	0.197	0.078
0.000	0.048	0.080	0.197	0.079
0.000	0.048	0.080	0.198	0.079
0.000	0.048	0.080	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.080	0.197	0.080

Standard Deviations	0.0004	0.0003	0.0006	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

QDM

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera DAVID E REYES-RIVERA
Signature and Printed Name

03/04/2019
Date

3/11/19
JO

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001121	Florida Highway Patrol Troop E	03/04/2019	<i>JLL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.09C	SN: SD3968 Temp: 34.10C	SN: SD3969 Temp: 34.12C	Lot AG805702
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

<p>FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:37</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:38</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:40</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:40</td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:36	Control Test	0.049	06:37	Air Blank	0.000	06:37	Control Test	0.049	06:38	Air Blank	0.000	06:38	Control Test	0.049	06:39	Air Blank	0.000	06:40	Control Test	0.049	06:40	Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:41</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:42</td></tr> <tr><td>Control Test</td><td>0.060</td><td>06:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:44</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:45</td></tr> <tr><td>Control Test</td><td>0.087</td><td>06:45</td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:41	Control Test	0.081	06:42	Air Blank	0.000	06:42	Control Test	0.060	06:43	Air Blank	0.000	06:44	Control Test	0.081	06:44	Air Blank	0.000	06:45	Control Test	0.087	06:45	Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157		<p>FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr> <tr><td>Control Test</td><td>0.198</td><td>06:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr> <tr><td>Control Test</td><td>0.198</td><td>06:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr> <tr><td>Control Test</td><td>0.198</td><td>06:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr> <tr><td>Control Test</td><td>0.1980</td><td>06:52</td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:48	Control Test	0.198	06:49	Air Blank	0.000	06:50	Control Test	0.198	06:50	Air Blank	0.000	06:51	Control Test	0.198	06:51	Air Blank	0.000	06:52	Control Test	0.1980	06:52	Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr> <tr><td>Control Test</td><td>0.075</td><td>06:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:56</td></tr> <tr><td>Control Test</td><td>0.075</td><td>06:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr> <tr><td>Control Test</td><td>0.076</td><td>06:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr> <tr><td>Control Test</td><td>0.076</td><td>06:57</td></tr> <tr><td>Average</td><td>0.0760</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:55	Control Test	0.075	06:55	Air Blank	0.000	06:56	Control Test	0.075	06:56	Air Blank	0.000	06:57	Control Test	0.076	06:57	Air Blank	0.000	06:57	Control Test	0.076	06:57	Average	0.0760		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:36																																																																																																																																																	
Control Test	0.049	06:37																																																																																																																																																	
Air Blank	0.000	06:37																																																																																																																																																	
Control Test	0.049	06:38																																																																																																																																																	
Air Blank	0.000	06:38																																																																																																																																																	
Control Test	0.049	06:39																																																																																																																																																	
Air Blank	0.000	06:40																																																																																																																																																	
Control Test	0.049	06:40																																																																																																																																																	
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:41																																																																																																																																																	
Control Test	0.081	06:42																																																																																																																																																	
Air Blank	0.000	06:42																																																																																																																																																	
Control Test	0.060	06:43																																																																																																																																																	
Air Blank	0.000	06:44																																																																																																																																																	
Control Test	0.081	06:44																																																																																																																																																	
Air Blank	0.000	06:45																																																																																																																																																	
Control Test	0.087	06:45																																																																																																																																																	
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test	0.198	06:49																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test	0.198	06:50																																																																																																																																																	
Air Blank	0.000	06:51																																																																																																																																																	
Control Test	0.198	06:51																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test	0.1980	06:52																																																																																																																																																	
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test	0.075	06:55																																																																																																																																																	
Air Blank	0.000	06:56																																																																																																																																																	
Control Test	0.075	06:56																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.076	06:57																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.076	06:57																																																																																																																																																	
Average	0.0760																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature	<i>JLL</i> Operator's Signature																																																																																																																																																

3/11/19
JLL

WORO



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001121, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001121</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FHP TROOP E MIAMI</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/04/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:59</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/04/2019

Date

David Reyes-Rivera

DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

3/11/19
DA

UBPO

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5750 (-0.0190)
 Sample #2 = 1.5520 (-0.0030)
 Sample #3 = 1.5760 (-0.0220)
 Sample #4 = 1.5990 (-0.0080)
 Avg % Abs = 1.5757 (-0.0110)
 STD DEV = 0.0235 (0.0098)
 REL STD DEV = 1.492 (89.535)

Auto Calibration
 Max Power Res Value = 32
 Auto Range Res Value = 24

Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12798, Sum Io = 14085
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.9410 (-0.0040)
 Sample #2 = 1.9380 (0.0150)
 Sample #3 = 1.9720 (0.0040)
 Sample #4 = 1.9430 (0.0180)
 Avg % Abs = 1.9510 (0.0123)
 STD DEV = 0.0184 (0.0074)
 REL STD DEV = 0.941 (59.766)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.7160 (-0.0210)
 Sample #2 = 3.6840 (0.0070)
 Sample #3 = 3.7120 (-0.0140)
 Sample #4 = 3.7200 (-0.0110)
 Avg % Abs = 3.7053 (-0.0060)
 STD DEV = 0.0189 (0.0114)
 REL STD DEV = 0.510 (189.297)

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12795, Sum Io = 14085
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6920 (0.0040)
 Sample #2 = 3.6930 (0.0190)
 Sample #3 = 3.7180 (-0.0060)
 Sample #4 = 3.7730 (-0.0220)
 Avg % Abs = 3.7280 (-0.0030)
 STD DEV = 0.0409 (0.0207)
 REL STD DEV = 1.098 (688.799)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12803, Sum Io = 14090
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8580 (-0.0100)
 Sample #2 = 0.8310 (0.0220)
 Sample #3 = 0.8680 (0.0060)
 Sample #4 = 0.8610 (0.0130)
 Avg % Abs = 0.8533 (0.0137)
 STD DEV = 0.0197 (0.0180)
 REL STD DEV = 2.303 (58.689)

<<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.102
 Std Dev = 0.01 Rel Std Dev = 13.83
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.853
 Std Dev = 0.02 Rel Std Dev = 2.30
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.951
 Std Dev = 0.02 Rel Std Dev = 0.34
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.728
 Std Dev = 0.04 Rel Std Dev = 1.10
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.463
 Std Dev = 0.01 Rel Std Dev = 0.13
 Zero Order Coef = -266.58
 First Order Coef = 2534.11
 Second Order Coef = 23.88
 Standard Deviation = 8.972286

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.101
 Std Dev = 0.02 Rel Std Dev = 14.81
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.576
 Std Dev = 0.02 Rel Std Dev = 1.49
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.705
 Std Dev = 0.02 Rel Std Dev = 0.51
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 7.088
 Std Dev = 0.03 Rel Std Dev = 0.40
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.310
 Std Dev = 0.01 Rel Std Dev = 0.08
 Zero Order Coef = -134.79
 First Order Coef = 1279.73
 Second Order Coef = 11.56
 Standard Deviation = 5.77553

<<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.000
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.000
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000

<<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.000
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.102
 Std Dev = 0.01 Rel Std Dev = 13.83
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.853
 Std Dev = 0.02 Rel Std Dev = 2.30
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.951
 Std Dev = 0.02 Rel Std Dev = 0.34
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.728
 Std Dev = 0.04 Rel Std Dev = 1.10
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.463
 Std Dev = 0.01 Rel Std Dev = 0.13
 Zero Order Coef = -266.58
 First Order Coef = 2534.11
 Second Order Coef = 23.88
 Standard Deviation = 8.972286

***** CHANNEL 2 *****
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.101
 Std Dev = 0.02 Rel Std Dev = 14.81
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.576
 Std Dev = 0.02 Rel Std Dev = 1.49
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.705
 Std Dev = 0.02 Rel Std Dev = 0.51
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 7.088
 Std Dev = 0.03 Rel Std Dev = 0.40
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.310
 Std Dev = 0.01 Rel Std Dev = 0.08
 Zero Order Coef = -134.79
 First Order Coef = 1279.73
 Second Order Coef = 11.56
 Standard Deviation = 5.77553

***** CHANNEL 1 *****
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.000
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000

***** CHANNEL 2 *****
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.000
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000

***** CHANNEL 1 *****
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.000
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000
 Sol Val = 0.040 mg/l or 0.000 g/210L
 % Abs = 0.040
 Std Dev = 0.000

Optical Calibration	
SN:	80-001121
Agency:	FHP Troop E
Date:	03/04/2019
Quadratic Fit:	+/-0.002g/210L
By:	<i>[Signature]</i>

3/11/19
[Signature]

WMPA

W600

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-001121	Florida Highway Patrol Troop E	03/04/2019	<i>JEK</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.09C 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.10C 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.12C 0.194 to 0.206 <input checked="" type="checkbox"/>	Lot AG805702 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>FHP TROOP E MIAMI Intoxilyzer - Alconol Analyzer Model: 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:48</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:48	Control Test	0.049	09:48	Air Blank	0.000	09:49	Control Test	0.049	09:50	Air Blank	0.000	09:50	Control Test	0.049	09:51	Air Blank	0.000	09:51	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>FHP TROOP E MIAMI Intoxilyzer - Alconol Analyzer Model: 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:54</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:53	Control Test	0.081	09:53	Air Blank	0.000	09:54	Control Test	0.080	09:55	Air Blank	0.000	09:55	Control Test	0.081	09:56	Air Blank	0.000	09:56	Control Test Stats			Average	0.0807		Std Dev	0.0005		Rel. Std Dev(%)	0.7157		<p>FHP TROOP E MIAMI Intoxilyzer - Alconol Analyzer Model: 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:57	Control Test	0.198	09:58	Air Blank	0.000	09:59	Control Test	0.199	09:59	Air Blank	0.000	10:00	Control Test	0.198	10:00	Air Blank	0.000	10:01	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel. Std Dev(%)	0.2911		<p>FHP TROOP E MIAMI Intoxilyzer - Alconol Analyzer Model: 8000 03/04/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:06	Control Test	0.078	10:06	Air Blank	0.000	10:07	Control Test	0.078	10:07	Air Blank	0.000	10:08	Control Test	0.078	10:08	Air Blank	0.000	10:08	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:48																																																																																																																																																	
Control Test	0.049	09:48																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.049	09:50																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test	0.049	09:51																																																																																																																																																	
Air Blank	0.000	09:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:53																																																																																																																																																	
Control Test	0.081	09:53																																																																																																																																																	
Air Blank	0.000	09:54																																																																																																																																																	
Control Test	0.080	09:55																																																																																																																																																	
Air Blank	0.000	09:55																																																																																																																																																	
Control Test	0.081	09:56																																																																																																																																																	
Air Blank	0.000	09:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel. Std Dev(%)	0.7157																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:57																																																																																																																																																	
Control Test	0.198	09:58																																																																																																																																																	
Air Blank	0.000	09:59																																																																																																																																																	
Control Test	0.199	09:59																																																																																																																																																	
Air Blank	0.000	10:00																																																																																																																																																	
Control Test	0.198	10:00																																																																																																																																																	
Air Blank	0.000	10:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test	0.078	10:06																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test	0.078	10:07																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test	0.078	10:08																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
<p><i>JEK</i> Operator's Signature</p> <p>3/14/19 <i>JEK</i></p>	<p><i>JEK</i> Operator's Signature</p>	<p><i>JEK</i> Operator's Signature</p>	<p><i>JEK</i> Operator's Signature</p>																																																																																																																																																