Return Material Authorization

Ship to: ☐ CMI, Inc.

☑ Enforcement Electronics

Shipment to repair facility authorized by: _____ on _____

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001208

Bill To Address:	Ship to Address:
Lee County SO	FDLE ATP Tallahassee

Reason for Return:

Returning instrument due to multiple interferent detect exception messages.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☐ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: _____

Phone #: _____ Email: _____

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Lee County S/N 80-001208

Date In 11/1/2019 DI Completion Date 11/13/19 ☐Ship ☐P/U ☐H/D ☐CMI ☒EE

Intake Performed By DP

☑ Annual
☐ Registration
☑ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☐ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

Quality Checks Performed By SP

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value 202
☒ Flow Verification (L/s)
Flow Column # ATP103
32 mm .148 (.139 - .169)
36 mm .160 (.156 - .190)
53 mm .234 (.228 - .278)
103 mm .503 (.447 - .547)
☒ Barometric Pressure Check
Gauge ID # 30793
☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201905A 5-14-21
0.080	DR1279	201905B 5-14-21
0.200	SD1013	201904D 4-30-21
0.080 DGS	N/A	AG831804 11-14-20

Flow Calibration Performed By _____

Flow Column # _____
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value _____
☐ Post Calibration Verification (L/s)
Flow Column # _____
32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance Performed By _____

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other _____

Temperature Checks Performed By SP

☒ Lab Temp °C 21.4
External Digital Therm. ID#: 300502
☒ 34°C +-.2 Serial #: _____
☒ 34°C +-.2 Serial #: _____
☒ 34°C +-.2 Serial #: _____

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure ID# 30793
Gauge 1022 Instrument 1021
Mouth Alcohol Solution Lot # 2018-B
Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	G2408
Interferent	G2882
0.050	SD1012
0.080	DR1279
0.200	SD1013

Attachments

☒ Form 41
☒ Stability Checks
☐ Calibration Certificate
☐ Calibration Adjustment
☐ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other _____

Notes/Suggested Service: _____

☐ **Instrument Complies with Chapter 11D-8, FAC**
☒ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☐ **Return to/Place into Evidentiary Use**
☒ **Remain Out of Evidentiary Use**
☐ **Conduct an Agency Inspection Before Evidentiary Use**

Tech Review / Date _____ Admin Review / Date _____

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection:11:27 Date of Inspection:11/13/2019
Serial Number: 80-001208
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047 / 0.050			
0.000	INT / 0.049			
0.000	/ 0.051			
0.000	/ INT			
0.000				
0.000				
0.000				
0.000				
0.000				
0.000				
Standard Deviations				

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:
Int Det: Purge Fail. 05: Interferent Detect, Interferent Detect. Non-compliance: .

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

11/13/2019

Date

FDLE/ATP Form 41 -- Revised August 2005

Reprinted From Database

Stability Checks # 80-001208

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001208
11/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.047	08:21
Air Blank	0.000	08:22
Control Test	0.048	08:22
Air Blank	0.000	08:23
Control Test	0.050	08:24
Air Blank	0.000	08:24
Control Test Stats		
Average	0.0483	
Std Dev	0.0015	
Rel Std Dev(%)	3.1604	

SP
Operator's Signature

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001208
11/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.078	08:26
Air Blank	0.000	08:27
Control Test	0.078	08:27
Air Blank	0.000	08:28
Control Test	0.077	08:29
Air Blank	0.000	08:29
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP
Operator's Signature

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001208
11/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:55
Control Test	0.198	07:55
Air Blank	0.000	07:56
Control Test	0.197	07:57
Air Blank	0.000	07:57
Control Test	0.198	07:58
Air Blank	0.000	07:58
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

SP
Operator's Signature

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001208
11/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:51
Control Test	0.080	07:51
Air Blank	0.000	07:52
Control Test	0.080	07:52
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:54
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP
Operator's Signature