



INSTRUMENT PROCESSING SHEET

1 of 2

Florida Department of
Law Enforcement

Agency

Palm Beach County SO

S/N

80-001234

Date In

4-30-19

DI Completion Date

☐ Ship☐ P/U☐ H/D☒ CMI☐ EE

Intake

Performed By SP

☒ Annual☐ Registration☐ Return from CMI / EE

Visual Inspection:

☒ Case☒ Handle☒ Keyboard☒ Dry Gas Shelf☒ Feet☒ Breath Tube☒ Ports☒ Screws Tight

Other Equipment/ Accessories:

☐ Power cord☐ Printer Cable☐ Static Bag☐ 12V DC Cable

Notes:

Quality Checks

Performed By SP

☒ Breath Tube Screen☒ Replace External O-Rings☒ Instrument Set Up Verified☒ R-Value 212☒ Flow Verification (L/s)

Flow Column # ATP101

32 mm .156 (.139 - .169)

36 mm .183 (.156 - .190)

53 mm .242 (.228 - .278)

103 mm .484 (.447 - .547)

☒ Barometric Pressure Check

Gauge ID # 081039

☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD3967	2017070 7-25-19
0.080	SD3968	201707E 7-25-19
0.200	SD3969	201707C 7-24-19
0.080 DGS	N/A	AG831804 11-14-20

Flow Calibration Performed By

Flow Column #

☐ 5L/min - 17mm☐ 15L/min - 53mm☐ 30L/min - 103mm☐ R-Value☐ Post Calibration Verification (L/s)

Flow Column #

32 mm (.139 - .169)

36 mm (.156 - .190)

53 mm (.228 - .278)

103 mm (.447 - .547)

Maintenance

Performed By

☐ Battery Replacement☐ Dry Gas Regulator Replacement☐ Breath Tube Replacement☐ Other

Temperature Checks Performed By

☐ Lab Temp °C

External Digital Therm. ID#:

☐ 34°C +/- .2 Serial #:☐ 34°C +/- .2 Serial #:☐ 34°C +/- .2 Serial #:

Final Release Date

Calibration Adjustment

Performed By SP

Barometric Pressure Gauge

ID # 88199

Simulator	Serial Number	Lot Number	Expiration
0.000	G12100	N/A	N/A
0.040	MP5098	17410	12-6-19
0.100	MP5099	18200	7-3-20
0.200	MP5100	19040	1-29-21
0.300	MP5101	18110	4-2-20
0.080 DGS	N/A	1781708DA2	8-5-19

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	SD3967	2017070	7-25-19
0.080	SD3968	201707E	7-25-19
0.200	SD3969	201707C	7-24-19
0.080 DGS	N/A	AG831804	11-14-20

Department Inspection

Performed By

Barometric Pressure ID#

Gauge Instrument

Mouth Alcohol Solution Lot #

Acetone Stock Solution Lot #

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

☐ Form 41☒ Stability Checks☐ Calibration Certificate☐ Calibration Adjustment☐ Post-Stability Checks☐ Flow Calibration☐ Form 40☐ Other

Notes/Suggested Service:

Unable to bring values within the acceptable range. Sending to CMI for repair.

SP

7/16/19 - Agency decided to deregister instrument rather than repair. Deregistered as of 6/15/19. Last BT conducted on this instrument was 7/15/16. SP

☐ Instrument Complies with Chapter 11D-8, FAC☒ Instrument Does Not Comply with Chapter 11D-8, FAC☐ Return to/Place into Evidentiary Use☒ Remain Out of Evidentiary Use☐ Conduct an Agency Inspection Before Evidentiary Use

Tech Review / Date

Admin Review / Date



INSTRUMENT PROCESSING SHEET

Florida Department of
Law Enforcement

Agency

Palm Beach County SO

S/N

80-001234

Date In

DI Completion Date

☐ Ship☐ P/U☐ H/D☒ CMI☐ EE

Intake Performed By _____ ☐ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☐ Case ☐ Handle ☐ Keyboard ☐ Dry Gas Shelf ☐ Feet ☐ Breath Tube ☐ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date _____ _____ _____	Quality Checks Performed By _____ ☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td></tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By _____ ☐ Lab Temp °C _____ External Digital Therm. ID#: _____ ☐ 34°C +/-2 Serial #: _____ ☐ 34°C +/-2 Serial #: _____ ☐ 34°C +/-2 Serial #: _____																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge 1013 ID # 28427 <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>SD1016</td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td>SD1022</td><td>17410</td><td>12-6-19</td></tr> <tr><td>0.100</td><td>SD3964</td><td>18200</td><td>7-3-20</td></tr> <tr><td>0.200</td><td>SD1025</td><td>19040</td><td>1-29-21</td></tr> <tr><td>0.300</td><td>SD1024</td><td>18110</td><td>4-2-20</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>17817080A2</td><td>8-5-19</td></tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td>SD1012</td><td>201707D</td><td>7-25-19</td></tr> <tr><td>0.080</td><td>DR1279</td><td>201707E</td><td>7-25-19</td></tr> <tr><td>0.200</td><td>SD1013</td><td>201707C</td><td>7-24-19</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>AG831804</td><td>11-14-20</td></tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000	SD1016	N/A	N/A	0.040	SD1022	17410	12-6-19	0.100	SD3964	18200	7-3-20	0.200	SD1025	19040	1-29-21	0.300	SD1024	18110	4-2-20	0.080 DGS	N/A	17817080A2	8-5-19	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1012	201707D	7-25-19	0.080	DR1279	201707E	7-25-19	0.200	SD1013	201707C	7-24-19	0.080 DGS	N/A	AG831804	11-14-20	Department Inspection Performed By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td></tr> <tr><td>Interferent</td><td></td></tr> <tr><td>0.050</td><td></td></tr> <tr><td>0.080</td><td></td></tr> <tr><td>0.200</td><td></td></tr> </tbody> </table> Attachments ☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	SD1016	N/A	N/A																																																										
0.040	SD1022	17410	12-6-19																																																										
0.100	SD3964	18200	7-3-20																																																										
0.200	SD1025	19040	1-29-21																																																										
0.300	SD1024	18110	4-2-20																																																										
0.080 DGS	N/A	17817080A2	8-5-19																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD1012	201707D	7-25-19																																																										
0.080	DR1279	201707E	7-25-19																																																										
0.200	SD1013	201707C	7-24-19																																																										
0.080 DGS	N/A	AG831804	11-14-20																																																										
Simulator	Serial Number																																																												
0.000																																																													
Interferent																																																													
0.050																																																													
0.080																																																													
0.200																																																													

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001234	Palm Beach County SO	4-30-19	SP

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
PALM BEACH CO SO Intoxilizer - Alcohol Analyzer Model 8000 04/30/2019 Software: 8100.27 SN 80-001234 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr><tr><td>Control Test</td><td>0.033</td><td>09:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:43</td></tr><tr><td>Control Test</td><td>0.033</td><td>09:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:44</td></tr><tr><td>Control Test</td><td>0.030</td><td>09:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:45</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0320</td><td></td></tr><tr><td>Std Dev</td><td>0.0017</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>5.4127</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:41	Control Test	0.033	09:42	Air Blank	0.000	09:43	Control Test	0.033	09:43	Air Blank	0.000	09:44	Control Test	0.030	09:45	Air Blank	0.000	09:45	Control Test Stats			Average	0.0320		Std Dev	0.0017		Rel Std Dev(%)	5.4127		PALM BEACH CO SO Intoxilizer - Alcohol Analyzer Model 8000 04/30/2019 Software: 8100.27 SN 80-001234 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:36</td></tr><tr><td>Control Test</td><td>0.083</td><td>09:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:37</td></tr><tr><td>Control Test</td><td>0.082</td><td>09:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:39</td></tr><tr><td>Control Test</td><td>0.082</td><td>09:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:40</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0823</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7012</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:36	Control Test	0.083	09:37	Air Blank	0.000	09:37	Control Test	0.082	09:38	Air Blank	0.000	09:39	Control Test	0.082	09:39	Air Blank	0.000	09:40	Control Test Stats			Average	0.0823		Std Dev	0.0006		Rel Std Dev(%)	0.7012		PALM BEACH CO SO Intoxilizer - Alcohol Analyzer Model 8000 04/30/2019 Software: 8100.27 SN 80-001234 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:31</td></tr><tr><td>Control Test</td><td>0.187</td><td>09:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:32</td></tr><tr><td>Control Test</td><td>0.188</td><td>09:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:33</td></tr><tr><td>Control Test</td><td>0.188</td><td>09:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:34</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1877</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.3076</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:31	Control Test	0.187	09:31	Air Blank	0.000	09:32	Control Test	0.188	09:33	Air Blank	0.000	09:33	Control Test	0.188	09:34	Air Blank	0.000	09:34	Control Test Stats			Average	0.1877		Std Dev	0.0006		Rel Std Dev(%)	0.3076		PALM BEACH CO SO Intoxilizer - Alcohol Analyzer Model 8000 04/30/2019 Software: 8100.27 SN 80-001234 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:48</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table> DGS	Test	g/210L	Time	Air Blank	0.000	09:48	Control Test	0.078	09:48	Air Blank	0.000	09:49	Control Test	0.078	09:49	Air Blank	0.000	09:50	Control Test	0.079	09:50	Air Blank	0.000	09:50	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:41																																																																																																																																																	
Control Test	0.033	09:42																																																																																																																																																	
Air Blank	0.000	09:43																																																																																																																																																	
Control Test	0.033	09:43																																																																																																																																																	
Air Blank	0.000	09:44																																																																																																																																																	
Control Test	0.030	09:45																																																																																																																																																	
Air Blank	0.000	09:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0320																																																																																																																																																		
Std Dev	0.0017																																																																																																																																																		
Rel Std Dev(%)	5.4127																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:36																																																																																																																																																	
Control Test	0.083	09:37																																																																																																																																																	
Air Blank	0.000	09:37																																																																																																																																																	
Control Test	0.082	09:38																																																																																																																																																	
Air Blank	0.000	09:39																																																																																																																																																	
Control Test	0.082	09:39																																																																																																																																																	
Air Blank	0.000	09:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0823																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7012																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:31																																																																																																																																																	
Control Test	0.187	09:31																																																																																																																																																	
Air Blank	0.000	09:32																																																																																																																																																	
Control Test	0.188	09:33																																																																																																																																																	
Air Blank	0.000	09:33																																																																																																																																																	
Control Test	0.188	09:34																																																																																																																																																	
Air Blank	0.000	09:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1877																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.3076																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:48																																																																																																																																																	
Control Test	0.078	09:48																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.078	09:49																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test	0.079	09:50																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
SP Operator's Signature	SP Operator's Signature	SP Operator's Signature	SP Operator's Signature																																																																																																																																																

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
05/02/2019 08:24:05

Auto Calibration
Max Power Res Value = 40
Auto Range Res Value = 23

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12744, Sum Io = 12785

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.0200 (-0.0050)
Sample #2 = 0.0350 (-0.0200)
Sample #3 = 0.0300 (0.0010)
Sample #4 = 0.0280 (0.0000)
Avg % Abs = 0.0310 (-0.0063)
STD DEV = 0.0036 (0.0118)
REL STD DEV = 11.631 (187.046)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.0870 (-0.0200)
Sample #2 = 0.0610 (-0.0280)
Sample #3 = 0.0640 (-0.0180)
Sample #4 = 0.0720 (-0.0180)
Avg % Abs = 0.0657 (-0.0213)
STD DEV = 0.0057 (0.0058)
REL STD DEV = 8.659 (27.063)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12742, Sum Io = 12787

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.7820 (-0.0060)
Sample #2 = 0.7630 (-0.0020)
Sample #3 = 0.7450 (0.0160)
Sample #4 = 0.7820 (0.0100)
Avg % Abs = 0.7633 (0.0080)
STD DEV = 0.0185 (0.0092)
REL STD DEV = 2.424 (114.564)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5750 (0.0010)
Sample #2 = 1.5840 (0.0110)
Sample #3 = 1.5550 (0.0160)
Sample #4 = 1.5560 (0.0180)
Avg % Abs = 1.5650 (0.0150)
STD DEV = 0.0165 (0.0036)
REL STD DEV = 1.052 (24.037)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12737, Sum Io = 12783

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.7850 (-0.0140)
Sample #2 = 1.8050 (-0.0030)
Sample #3 = 1.7750 (0.0030)
Sample #4 = 1.7820 (0.0150)
Avg % Abs = 1.7687 (0.0050)
STD DEV = 0.0142 (0.0092)
REL STD DEV = 0.795 (183.303)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.6390 (-0.0160)
Sample #2 = 3.6220 (0.0120)
Sample #3 = 3.6080 (0.0210)
Sample #4 = 3.5910 (0.0300)
Avg % Abs = 3.6070 (0.0210)
STD DEV = 0.0155 (0.0090)
REL STD DEV = 0.430 (42.857)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12735, Sum Io = 12782

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.5190 (-0.0310)
Sample #2 = 3.5050 (-0.0190)
Sample #3 = 3.4920 (-0.0140)
Sample #4 = 3.4660 (0.0120)
Avg % Abs = 3.4877 (-0.0070)
STD DEV = 0.0199 (0.0166)
REL STD DEV = 0.569 (237.762)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.9580 (-0.0170)
Sample #2 = 6.9020 (0.0130)
Sample #3 = 6.8990 (0.0210)
Sample #4 = 6.9140 (0.0380)
Avg % Abs = 6.9050 (0.0240)
STD DEV = 0.0079 (0.0128)
REL STD DEV = 0.115 (53.196)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12733, Sum Io = 12781

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 4.9410 (-0.0230)
Sample #2 = 4.9710 (-0.0320)
Sample #3 = 4.9370 (-0.0290)
Sample #4 = 4.9480 (-0.0040)
Avg % Abs = 4.9520 (-0.0217)
STD DEV = 0.0173 (0.0154)
REL STD DEV = 0.350 (70.953)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.6900 (0.0190)
Sample #2 = 9.7500 (-0.0020)
Sample #3 = 9.6890 (-0.0030)
Sample #4 = 9.7410 (-0.0080)
Avg % Abs = 9.7267 (-0.0043)
STD DEV = 0.0329 (0.0032)
REL STD DEV = 0.339 (74.182)

Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.031
Std Dev = 0.00 Rel Std Dev = 11.63
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.763
Std Dev = 0.02 Rel Std Dev = 2.42
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.789
Std Dev = 0.01 Rel Std Dev = 0.80
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.488
Std Dev = 0.02 Rel Std Dev = 0.57
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 4.952
Std Dev = 0.02 Rel Std Dev = 0.35
Zero Order Coef = -49.42
First Order Coef = 2500.34
Second Order Coef = 78.15
Standard Deviation = 70.07663

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.066
Std Dev = 0.01 Rel Std Dev = 8.66
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.565
Std Dev = 0.02 Rel Std Dev = 1.05
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.607
Std Dev = 0.02 Rel Std Dev = 0.43
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.905
Std Dev = 0.01 Rel Std Dev = 0.11
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.727
Std Dev = 0.03 Rel Std Dev = 0.34
Zero Order Coef = -63.38
First Order Coef = 1224.20
Second Order Coef = 25.48
Standard Deviation = 59.127895

Channel 1
Sample #1 = 3168.00
Sample #2 = 3186.00
Sample #3 = 3195.00
Sample #4 = 3192.00
Average Result = 3191.0000
STD DEV = 4.5826
REL STD DEV = 0.144
Channel 2
Sample #1 = 3275.00
Sample #2 = 3290.00
Sample #3 = 3273.00
Sample #4 = 3246.00
Average Result = 3269.6667
STD DEV = 22.1886
REL STD DEV = 0.679

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1015
3 um H2O Adjust (mg/l*10,000) = 618
9 um H2O Adjust (mg/l*10,000) = 540
**** AUTO CAL PASS

CAL ADJUSTMENT #1
#80-001234 SP

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0006
0.040 0.040 0.0000
0.100 0.098 0.0019
0.200 0.202 -0.0021
0.300 0.299 0.0008

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0004
0.040 0.040 -0.0002
0.100 0.098 0.0016
0.200 0.202 -0.0017
0.300 0.299 0.0007

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

Channel 1

Sample #1 = 3168.00
Sample #2 = 3186.00
Sample #3 = 3195.00
Sample #4 = 3192.00
Average Result = 3191.0000
STD DEV = 4.5826
REL STD DEV = 0.144

Channel 2

Sample #1 = 3275.00
Sample #2 = 3290.00
Sample #3 = 3273.00
Sample #4 = 3246.00
Average Result = 3269.6667
STD DEV = 22.1886
REL STD DEV = 0.679

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities - Post Cal Adjust	80-001234	Palm Beach County SO	5-8-19	SP

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.05 PALM BEACH CO SO Intoxilyzer - Alconol Analyzer Model 8000 05/02/2019 Software: 8100.27 SN 80-001234 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:27</td></tr><tr><td>Control Test</td><td>0.037</td><td>09:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:28</td></tr><tr><td>Control Test</td><td>0.038</td><td>09:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:29</td></tr><tr><td>Control Test</td><td>0.038</td><td>09:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:30</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0377</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.5328</td><td></td></tr></tbody></table> Operator's Signature SP	Test	g/210L	Time	Air Blank	0.000	09:27	Control Test	0.037	09:27	Air Blank	0.000	09:28	Control Test	0.038	09:29	Air Blank	0.000	09:29	Control Test	0.038	09:30	Air Blank	0.000	09:30	Control Test Stats			Average	0.0377		Std Dev	0.0006		Rel Std Dev(%)	1.5328		0.08 PALM BEACH CO SO Intoxilyzer - Alconol Analyzer Model 8000 05/02/2019 Software: 8100.27 SN 80-001234 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:32</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:33</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:34</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:35</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table> Operator's Signature SP	Test	g/210L	Time	Air Blank	0.000	09:32	Control Test	0.050	09:33	Air Blank	0.000	09:33	Control Test	0.049	09:34	Air Blank	0.000	09:34	Control Test	0.049	09:35	Air Blank	0.000	09:35	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		PALM BEACH CO SO Intoxilyzer - Alconol Analyzer Model 8000 05/02/2019 Software: 8100.27 SN 80-001234 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:40</td></tr><tr><td>Control Test</td><td>0.198</td><td>09:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr><tr><td>Control Test</td><td>0.198</td><td>09:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:42</td></tr><tr><td>Control Test</td><td>0.197</td><td>09:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:44</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1977</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr></tbody></table> Operator's Signature SP	Test	g/210L	Time	Air Blank	0.000	09:40	Control Test	0.198	09:41	Air Blank	0.000	09:41	Control Test	0.198	09:42	Air Blank	0.000	09:42	Control Test	0.197	09:43	Air Blank	0.000	09:44	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		PALM BEACH CO SO Intoxilyzer - Alconol Analyzer Model 8000 05/02/2019 Software: 8100.27 SN 80-001234 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:46</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:47</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:48</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:48</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></tbody></table> DGS Operator's Signature SP	Test	g/210L	Time	Air Blank	0.000	09:46	Control Test	0.078	09:46	Air Blank	0.000	09:47	Control Test	0.079	09:47	Air Blank	0.000	09:48	Control Test	0.078	09:48	Air Blank	0.000	09:48	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:27																																																																																																																																																	
Control Test	0.037	09:27																																																																																																																																																	
Air Blank	0.000	09:28																																																																																																																																																	
Control Test	0.038	09:29																																																																																																																																																	
Air Blank	0.000	09:29																																																																																																																																																	
Control Test	0.038	09:30																																																																																																																																																	
Air Blank	0.000	09:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0377																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.5328																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:32																																																																																																																																																	
Control Test	0.050	09:33																																																																																																																																																	
Air Blank	0.000	09:33																																																																																																																																																	
Control Test	0.049	09:34																																																																																																																																																	
Air Blank	0.000	09:34																																																																																																																																																	
Control Test	0.049	09:35																																																																																																																																																	
Air Blank	0.000	09:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:40																																																																																																																																																	
Control Test	0.198	09:41																																																																																																																																																	
Air Blank	0.000	09:41																																																																																																																																																	
Control Test	0.198	09:42																																																																																																																																																	
Air Blank	0.000	09:42																																																																																																																																																	
Control Test	0.197	09:43																																																																																																																																																	
Air Blank	0.000	09:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:46																																																																																																																																																	
Control Test	0.078	09:46																																																																																																																																																	
Air Blank	0.000	09:47																																																																																																																																																	
Control Test	0.079	09:47																																																																																																																																																	
Air Blank	0.000	09:48																																																																																																																																																	
Control Test	0.078	09:48																																																																																																																																																	
Air Blank	0.000	09:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

PALM BEACH CO SO

Intoxilyzer - Alcohol Analyzer

SN 80-001234

10:15:12

Auto Calibration

Max Power Res Value = 40

Auto Range Res Value = 22

Sol Value = 0.000 g/210L ***

Fit value = 0.000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12643, Sum Io = 12724

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0350 (-0.0350)

Sample #2 = 0.0260 (-0.0090)

Sample #3 = 0.0480 (0.0070)

Sample #4 = 0.0570 (0.0060)

Aug % Abs = 0.0437 (0.0013)

STD DEV = 0.0159 (0.0090)

REL STD DEV = 36.522 (672.216)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1020 (-0.0270)

Sample #2 = 0.0810 (-0.0110)

Sample #3 = 0.0820 (-0.0100)

Sample #4 = 0.0930 (-0.0100)

Aug % Abs = 0.0853 (-0.0103)

STD DEV = 0.0067 (0.0066)

REL STD DEV = 7.803 (5.587)

Sol Value = 0.040 g/210L ***

Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12635, Sum Io = 12724

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.4570 (-0.0220)

Sample #2 = 0.4680 (0.0060)

Sample #3 = 0.4780 (-0.0040)

Sample #4 = 0.4590 (0.0170)

Aug % Abs = 0.4683 (0.0063)

STD DEV = 0.0095 (0.0105)

REL STD DEV = 2.029 (165.852)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.9110 (-0.0160)

Sample #2 = 0.9290 (0.0060)

Sample #3 = 0.9230 (0.0060)

Sample #4 = 0.9220 (0.0070)

Aug % Abs = 0.9247 (0.0053)

STD DEV = 0.0038 (0.0006)

REL STD DEV = 0.409 (9.116)

Sol Value = 0.100 g/210L ***

Fit value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12629, Sum Io = 12722

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.1830 (-0.0200)

Sample #2 = 1.1790 (0.0000)

Sample #3 = 1.1700 (0.0270)

Sample #4 = 1.1710 (0.0320)

Aug % Abs = 1.1733 (0.0197)

STD DEV = 0.0049 (0.0172)

REL STD DEV = 0.420 (87.531)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 2.3320 (0.0020)

Sample #2 = 2.3480 (0.0070)

Sample #3 = 2.3510 (0.0070)

Sample #4 = 2.3450 (0.0280)

Aug % Abs = 2.3480 (0.0140)

STD DEV = 0.0030 (0.0121)

REL STD DEV = 0.128 (86.603)

Sol Value = 0.200 g/210L ***

Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12623, Sum Io = 12716

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 2.6110 (-0.0110)

Sample #2 = 2.6120 (0.0120)

Sample #3 = 2.6120 (0.0200)

Sample #4 = 2.6060 (0.0080)

Aug % Abs = 2.6100 (0.0133)

STD DEV = 0.0035 (0.0061)

REL STD DEV = 0.133 (45.826)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 5.1800 (-0.0280)

Sample #2 = 5.1760 (-0.0080)

Sample #3 = 5.1720 (-0.0010)

Sample #4 = 5.1500 (0.0000)

Aug % Abs = 5.1660 (-0.0030)

STD DEV = 0.0140 (0.0044)

REL STD DEV = 0.271 (145.297)

Sol Value = 0.300 g/210L ***

Fit value = 1.4286 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12616, Sum Io = 12716

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.5760 (-0.0080)

Sample #2 = 3.5920 (-0.0100)

Sample #3 = 3.6170 (-0.0110)

Sample #4 = 3.6180 (0.0040)

Aug % Abs = 3.6090 (-0.0057)

STD DEV = 0.0147 (0.0084)

REL STD DEV = 0.408 (147.997)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 7.0670 (0.0100)

Sample #2 = 7.0760 (0.0170)

Sample #3 = 7.1070 (0.0140)

Sample #4 = 7.1310 (0.0090)

Aug % Abs = 7.1047 (0.0133)

STD DEV = 0.0276 (0.0040)

REL STD DEV = 0.388 (30.311)

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.044

Std Dev = 0.02 Rel Std Dev = 36.52

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.468

Std Dev = 0.01 Rel Std Dev = 2.03

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.173

Std Dev = 0.00 Rel Std Dev = 0.42

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 2.610

Std Dev = 0.00 Rel Std Dev = 0.13

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 3.609

Std Dev = 0.01 Rel Std Dev = 0.41

Zero Order Coef = 80.96

First Order Coef = 3575.79

Second Order Coef = 84.30

Standard Deviation = 341.299133

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.085

Std Dev = 0.01 Rel Std Dev = 7.80

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.925

Std Dev = 0.00 Rel Std Dev = 0.41

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 2.348

Std Dev = 0.00 Rel Std Dev = 0.13

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 5.166

Std Dev = 0.01 Rel Std Dev = 0.27

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 7.105

Std Dev = 0.03 Rel Std Dev = 0.39

Zero Order Coef = 91.18

First Order Coef = 1774.26

Second Order Coef = 27.42

Standard Deviation = 340.784515

CAL ADJUSTMENT #2

#80-001234 SP

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.005	-0.0050
0.040	0.037	0.0027
0.100	0.092	0.0078
0.200	0.210	-0.0097
0.300	0.296	0.0042

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.005	-0.0051
0.040	0.037	0.0031
0.100	0.093	0.0074
0.200	0.210	-0.0098
0.300	0.296	0.0043

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 4642.00

Sample #2 = 4598.00

Sample #3 = 4722.00

Sample #4 = 4678.00

Average Result = 4666.0000

STD DEV = 62.8649

REL STD DEV = 1.347

***** CHANNEL 2

Sample #1 = 4838.00

Sample #2 = 4842.00

Sample #3 = 4885.00

Sample #4 = 4872.00

Average Result = 4866.3335

STD DEV = 22.0530

REL STD DEV = 0.453

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1013

3 um H2O Adjust (mg/l*10,000) = -856

9 um H2O Adjust (mg/l*10,000) = -1056

**** AUTO CAL PASS

Post Cal-Adjust

Stability Checks #80-001234

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001234
06/03/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:00
Control Test	0.056	11:01
Air Blank	0.000	11:02
Control Test	0.058	11:03
Air Blank	0.000	11:03
Control Test	0.059	11:03
Air Blank	0.000	11:04
Control Test Stats		
Average	0.0577	
Std Dev	0.0015	
Rel Std Dev(%)	2.6489	

SP
Operator's Signature

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001234
06/03/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:09
Control Test	0.205	11:10
Air Blank	0.000	11:11
Control Test	0.210	11:11
Air Blank	0.000	11:12
Control Test	0.211	11:13
Air Blank	0.000	11:13
Control Test Stats		
Average	0.2087	
Std Dev	0.0032	
Rel Std Dev(%)	1.5405	

SP
Operator's Signature

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001234
06/03/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:18
Control Test	0.079	11:19
Air Blank	0.000	11:19
Control Test	0.079	11:19
Air Blank	0.000	11:20
Control Test	0.077	11:20
Air Blank	0.000	11:21
Control Test Stats		
Average	0.0783	
Std Dev	0.0012	
Rel Std Dev(%)	1.4741	

DGS

SP
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 10:14

Date of Inspection: 06/03/2019

Serial Number: 80-001234
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A COMPLIANCE
NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

06/03/2019
Date