



INSTRUMENT PROCESSING SHEET

Agency Indian Harbour Beach PD

S/N 80-001263

Florida Department of Law Enforcement

Date In 10/10/2019 DI Completion Date 10/17/19

Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>219</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.480</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date <p style="text-align: center;">FDLE</p> <p style="text-align: center;">OCT 22 2019</p> <p style="text-align: center;">Alcohol Testing Program</p>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201905A 5-14-21</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201905B 5-14-21</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201904D 4-30-21</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG916501 6-14-21</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201905A 5-14-21	0.080	DR1279	201905B 5-14-21	0.200	SD1013	201904D 4-30-21	0.080 DGS	N/A	AG916501 6-14-21	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	SD1012	201905A 5-14-21																																																													
0.080	DR1279	201905B 5-14-21																																																													
0.200	SD1013	201904D 4-30-21																																																													
0.080 DGS	N/A	AG916501 6-14-21																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1012</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	G2408																																																														
Interferent	G2882																																																														
0.050	SD1012																																																														
0.080	DR1279																																																														
0.200	SD1013																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____		Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																													
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
		<u>GDPM 10/17/19</u> <u>Brett Kirkland 10/18/19</u> Tech Review / Date Admin Review / Date																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001263, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001263</u>	UNCERTAINTY* ±
Owning Agency:	<u>INDIAN HARBOUR PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/17/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:16</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/17/2019

Date

Shayla Platt

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

PDAM
YBR 10/18/19

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN HARBOUR PD
Time of Inspection: 12:16

Date of Inspection: 10/17/2019

Serial Number: 80-001263
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.076	0.199	0.080
0.000	0.050	0.076	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.077	0.200	0.081
0.000	0.049	0.076	0.199	0.080
0.000	0.049	0.078	0.200	0.081
0.000	0.049	0.077	0.200	0.080
0.000	0.049	0.077	0.200	0.081
0.000	0.050	0.077	0.200	0.081
0.000	0.049	0.079	0.200	0.081

Standard Deviations	0.0005	0.0009	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

PJM
+BK
10/18/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt
SHAYLA D PLATT
Signature and Printed Name

10/17/2019
Date

INDIAN HARBOUR PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001263
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.049	09:41
Air Blank	0.000	09:41
Control Test	0.049	09:42
Air Blank	0.000	09:43
Control Test	0.049	09:43
Air Blank	0.000	09:44
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

INDIAN HARBOUR PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001263
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:28
Control Test	0.077	09:29
Air Blank	0.000	09:30
Control Test	0.078	09:30
Air Blank	0.000	09:31
Control Test	0.079	09:32
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

SP
Operator's Signature

INDIAN HARBOUR PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001263
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.200	09:33
Air Blank	0.000	09:34
Control Test	0.200	09:35
Air Blank	0.000	09:35
Control Test	0.200	09:36
Air Blank	0.000	09:36
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

INDIAN HARBOUR PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001263
10/17/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.080	09:37
Air Blank	0.000	09:38
Control Test	0.080	09:38
Air Blank	0.000	09:39
Control Test	0.080	09:39
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP
Operator's Signature

SP
TSK
10/18/19