



# INSTRUMENT PROCESSING SHEET

Agency Pinellas County SOS/N 80-001274Florida Department of  
Law EnforcementDate In 05/16/2019DI Completion Date 05/17/2019 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>KP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		<b>Quality Checks</b> Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>190</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
<b>Final Release Date</b> <p style="text-align: center;"><b>FDLE</b></p> <p style="text-align: center;">MAY 20 2019</p> <p style="text-align: center;">Alcohol Testing Program</p>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG831804 11/14/2020</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	SD1013	201707C 07/24/2019	0.080 DGS	N/A	AG831804 11/14/2020	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																		
0.050	SD1012	201707D 07/25/2019																		
0.080	DR1279	201707E 07/25/2019																		
0.200	SD1013	201707C 07/24/2019																		
0.080 DGS	N/A	AG831804 11/14/2020																		
		<b>Temperature Checks</b> Performed By <u>JD</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.0</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD1013</u>																		

<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____				<b>Department Inspection</b> Performed By <u>JD</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1016</u> Instrument <u>1007</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u>																																											
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	G2408																																														
Interferent	G2882																																														
0.050	SD1012																																														
0.080	DR1279																																														
0.200	SD1013																																														
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																											
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A																										
Simulator	Serial Number	Lot Number	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														

Notes/Suggested Service: <u>Agency Inspector Deane reported a 0.20 g/210L analysis resulted in a "control outside tolerance". He sent the instrument to me for an evaluation.</u> <u>JD</u>		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	
_____ _____ _____		<u>SP 5/20/19</u> <u>Butt Kirkland 5/20/19</u> Tech Review / Date Admin Review / Date	

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO  
Time of Inspection: 11:05

Date of Inspection: 05/17/2019

Serial Number: 80-001274  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.048	0.079	0.196	0.080
0.000	0.049	0.079	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.048	0.079	0.196	0.080
0.000	0.049	0.079	0.196	0.080
0.000	0.048	0.079	0.196	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080

Standard Deviations	0.0005	0.0005	0.0004	0.0000
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

SP  
BK  
5/17/19

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
\_\_\_\_\_  
Signature and Printed Name

THOMAS J GRAHAM

05/17/2019  
Date

80-001274

5/17/19  
SP

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001274  
05/17/2019  
Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:09
Control Test	0.048	08:09
Air Blank	0.000	08:10
Control Test	0.048	08:11
Air Blank	0.000	08:11
Control Test	0.049	08:12
Air Blank	0.000	08:12
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

SP

Operator's Signature

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001274  
05/17/2019  
Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:04
Control Test	0.078	08:05
Air Blank	0.000	08:05
Control Test	0.078	08:05
Air Blank	0.000	08:06
Control Test	0.078	08:07
Air Blank	0.000	08:08
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001274  
05/17/2019  
Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:18
Control Test	0.197	08:19
Air Blank	0.000	08:19
Control Test	0.196	08:20
Air Blank	0.000	08:21
Control Test	0.196	08:21
Air Blank	0.000	08:22
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

SP

Operator's Signature

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001274  
05/17/2019  
Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:15
Control Test	0.080	08:15
Air Blank	0.000	08:16
Control Test	0.081	08:16
Air Blank	0.000	08:17
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

SP  
BK  
5/17/19

SP

Operator's Signature



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001274, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001274</u>	UNCERTAINTY* ±
Owning Agency:	<u>PINELLAS COUNTY SO</u>	0.050 g/ 210 L      0.004
Calibration Date:	<u>05/17/2019</u>	0.080 g/ 210 L      0.004
Calibration Time:	<u>11:05</u>	0.200 g/ 210 L      0.007
		0.080 g/ 210 L Dry Gas Control      0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/17/2019

Date

  
**THOMAS J GRAHAM,**  
Department Inspector

SP  
13M  
5/17/19

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO

S/N 80-001274

Florida Department of Law Enforcement

Date In 02/08/2019 DI Completion Date 02/08/2019

Ship P/U H/D CMI EE

<b>Intake</b> Performed By <u>JP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>JP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	DR3856	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	SD1012	201707D 07/25/2019															
0.080	DR1279	201707E 07/25/2019															
0.200	DR3856	201707C 07/24/2019															
0.080 DGS	N/A	AG805701 02/26/2020															

**Final Release Date**

**FDLE**

FEB 11 2019

Alcohol Testing Program

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By JP

Lab Temp °C 21.3  
 External Digital Therm. ID#: 300503  
 34°C +/- .2 Serial #: SD1012  
 34°C +/- .2 Serial #: DR1279  
 34°C +/- .2 Serial #: DR3856

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By JP

Barometric Pressure ID# 26932  
 Gauge 1022 Instrument 1014  
 Mouth Alcohol Solution Lot # 2018-B  
 Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G2408
Interferent	G2882
0.050	SD1012
0.080	DR1279
0.200	DR3856

**Attachments**

Form 41       Post-Stability Checks  
 Stability Checks     Flow Calibration  
 Calibration Certificate     Form 40  
 Calibration Adjustment     Other \_\_\_\_\_

Notes/Suggested Service: Cleared bench area, changed dry gas standard (same lot #), and repeated 0.08 DGS test.

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

JP 2/8/19      Brett Kirkland 2/11/19  
 Tech Review / Date      Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO  
Time of Inspection: 15:34

Date of Inspection: 02/08/2019

Serial Number: 80-001274  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.078	0.196	0.079 / 0.079
0.000	0.048	0.078	0.197	0.079 / 0.079
0.000	0.048	0.078	0.196	0.078 / 0.079
0.000	0.048	0.079	0.196	0.079 / 0.079
0.000	0.048	0.078	0.196	0.079 / 0.079
0.000	0.048	0.078	0.196	0.079 / 0.079
0.000	0.048	0.078	0.196	0.079 / 0.078
0.000	0.048	0.078	0.196	0.075 / 0.079
0.000	0.048	0.078	0.195	INT / 0.079
0.000	0.049	0.078	0.196	/ 0.079
Standard Deviations	0.0003	0.0003	0.0004	/ 0.0003


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

**Remarks:**

08: Interferent Detect CHANGED TANK AND REPEATED TEST.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



THOMAS J GRAHAM  
Signature and Printed Name

02/08/2019  
Date

SP  
BK  
2/11/19

80-001274

2/8/19  
JSD

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
02/08/2019  
Software: 8100.27  
SN 80-001274

Test	g/210L	Time
Air Blank	0.000	12:43
Control Test	0.195	12:43
Air Blank	0.000	12:44
Control Test	0.196	12:44
Air Blank	0.000	12:45
Control Test	0.196	12:46
Air Blank	0.000	12:46
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
02/08/2019  
Software: 8100.27  
SN 80-001274

Test	g/210L	Time
Air Blank	0.000	13:02
Control Test	0.078	13:03
Air Blank	0.000	13:03
Control Test	0.079	13:04
Air Blank	0.000	13:04
Control Test	0.078	13:05
Air Blank	0.000	13:06
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
02/08/2019  
Software: 8100.27  
SN 80-001274

Test	g/210L	Time
Air Blank	0.000	12:57
Control Test	0.048	12:58
Air Blank	0.000	12:58
Control Test	0.048	12:59
Air Blank	0.000	12:59
Control Test	0.048	13:00
Air Blank	0.000	13:01
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
02/08/2019  
Software: 8100.27  
SN 80-001274

Test	g/210L	Time
Air Blank	0.000	12:48
Control Test	0.079	12:48
Air Blank	0.000	12:49
Control Test	0.079	12:49
Air Blank	0.000	12:50
Control Test	0.079	12:50
Air Blank	0.000	12:50
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

JSD  
Operator's Signature

JSD  
Operator's Signature

JSD  
Operator's Signature

JSD  
Operator's Signature

SP  
BK  
2/11/19



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001274, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001274</u>	Serial Number:	<u>80-001274</u>
Owning Agency:	<u>PINELLAS COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/08/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:34</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

UNCERTAINTY\* ±

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/08/2019

Date

*Th. J. Graham*

THOMAS J GRAHAM,  
Department Inspector

*SP BK 2/11/19*

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality