



INSTRUMENT PROCESSING SHEET

Agency Baker CountyS/N 80-001286Florida Department of
Law EnforcementDate In 9/4/2019DI Completion Date 9/10/19☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>[Signature]</u>	Quality Checks	Performed By <u>SP</u>	Flow Calibration	Performed By <u>SP</u>																																																										
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>175</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>148</u> (.139 - .169) 36 mm <u>164</u> (.156 - .190) 53 mm <u>238</u> (.228 - .278) 103 mm <u>500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE SEP 09 2019 Alcohol Testing Program		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>SD1012</td><td>201905A 5-14-21</td></tr><tr><td>0.080</td><td>DR1279</td><td>201905B 5-14-21</td></tr><tr><td>0.200</td><td>SD1013</td><td>201905D 4-30-21</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG916501 8-5-14-21</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201905A 5-14-21	0.080	DR1279	201905B 5-14-21	0.200	SD1013	201905D 4-30-21	0.080 DGS	N/A	AG916501 8-5-14-21	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>SP</u> ☒ Lab Temp °C <u>21.4</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/- .2 Serial #: <u>SD1012</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1013</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	SD1012	201905A 5-14-21																																																													
0.080	DR1279	201905B 5-14-21																																																													
0.200	SD1013	201905D 4-30-21																																																													
0.080 DGS	N/A	AG916501 8-5-14-21																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1011</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>G2408</td></tr><tr><td>Interferent</td><td>G2882</td></tr><tr><td>0.050</td><td>SD1012</td></tr><tr><td>0.080</td><td>DR1279</td></tr><tr><td>0.200</td><td>SD1013</td></tr></tbody></table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date <u>9/9/19</u> Admin Review / Date <u>9/9/19</u>		Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	G2408																																																														
Interferent	G2882																																																														
0.050	SD1012																																																														
0.080	DR1279																																																														
0.200	SD1013																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____																																																															

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
09/06/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:19
Control Test	0.047	09:20
Air Blank	0.000	09:20
Control Test	0.047	09:21
Air Blank	0.000	09:21
Control Test	0.047	09:22
Air Blank	0.000	09:23
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
09/06/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.077	10:05
Air Blank	0.000	10:06
Control Test	0.078	10:07
Air Blank	0.000	10:07
Control Test	0.078	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	



Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
09/06/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.196	09:06
Air Blank	0.000	09:07
Control Test	0.195	09:07
Air Blank	0.000	09:08
Control Test	0.196	09:09
Air Blank	0.000	09:09
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	



Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
09/06/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.079	09:35
Air Blank	0.000	09:36
Control Test	0.079	09:36
Air Blank	0.000	09:37
Control Test	0.079	09:37
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DAS



Operator's Signature

PDm BK 9/9/19

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO

Time of Inspection: 13:45

Date of Inspection: 09/06/2019

Serial Number: 80-001286

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.078	0.200	0.079
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.080	0.203	0.079
0.000	0.051	0.079	0.203	0.080
0.000	0.051	0.080	0.203	0.081
0.000	0.050	0.079	0.202	0.081
0.000	0.050	0.079	0.203	0.081
0.000	0.050	0.079	0.203	0.081
0.000	0.050	0.079	0.202	0.081
0.000	0.050	0.079	0.203	0.081

Standard Deviations	0.0008	0.0005	0.0009	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

PK
9/9/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

09/06/2019
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001286, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001286</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/06/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:45</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/06/2019 Shayla Platt
Date SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

rgm
RSK 9/9/19



INSTRUMENT PROCESSING SHEET

Agency Baker County SOS/N 80-001286Florida Department of
Law EnforcementDate In 11/5/18DI Completion Date 1/7/19☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>SQC</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____		Quality Checks Performed By <u>JE</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>180</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.257</u> (.228 - .278) 103 mm <u>.535</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE JAN 07 2019 Alcohol Testing Program		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805701</u> <u>2/26/20</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>DR3856</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG805701</u> <u>2/26/20</u>	Maintenance Performed By <u>JE</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced Internal Printer Paper</u> Temperature Checks Performed By <u>JE</u> ☒ Lab Temp °C <u>21.7</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>SD1012</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>DR3856</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	<u>SD1012</u>	<u>201707D</u> <u>7/25/19</u>																																																													
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>																																																													
0.200	<u>DR3856</u>	<u>201707C</u> <u>7/24/19</u>																																																													
0.080 DGS	N/A	<u>AG805701</u> <u>2/26/20</u>																																																													
Calibration Adjustment Performed By <u>JE</u> Barometric Pressure Gauge <u>1022</u> ID # <u>30793</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2834</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1022</u></td> <td><u>17410</u></td> <td><u>12/6/19</u></td> </tr> <tr> <td>0.100</td> <td><u>SD3964</u></td> <td><u>18070</u></td> <td><u>2/26/20</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>17340</u></td> <td><u>10/9/19</u></td> </tr> <tr> <td>0.300</td> <td><u>SD1024</u></td> <td><u>18110</u></td> <td><u>4/2/20</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817080A2</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805701</u></td> <td><u>2/26/20</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G2834</u>	N/A	N/A	0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>	0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>	0.200	<u>SD1025</u>	<u>17340</u>	<u>10/9/19</u>	0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>	0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1012</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>DR3856</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A	<u>AG805701</u>	<u>2/26/20</u>	Department Inspection Performed By <u>JE</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1021</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2408</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2882</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1012</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>G2408</u>	Interferent	<u>G2882</u>	0.050	<u>SD1012</u>	0.080	<u>DR1279</u>	0.200	<u>DR3856</u>
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000	<u>G2834</u>	N/A	N/A																																																												
0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>																																																												
0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>																																																												
0.200	<u>SD1025</u>	<u>17340</u>	<u>10/9/19</u>																																																												
0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>																																																												
0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>																																																												
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050	<u>SD1012</u>	<u>201707D</u>	<u>7/25/19</u>																																																												
0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>																																																												
0.200	<u>DR3856</u>	<u>201707C</u>	<u>7/24/19</u>																																																												
0.080 DGS	N/A	<u>AG805701</u>	<u>2/26/20</u>																																																												
Simulator	Serial Number																																																														
0.000	<u>G2408</u>																																																														
Interferent	<u>G2882</u>																																																														
0.050	<u>SD1012</u>																																																														
0.080	<u>DR1279</u>																																																														
0.200	<u>DR3856</u>																																																														
Notes/Suggested Service: <u>Change level 2 password to something unique. JE</u> _____ _____ _____ _____		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>DDM 1/7/19</u> Tech Review / Date <u>Brett Kirkland 1/7/19</u> Admin Review / Date																																																											

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 12:56

Date of Inspection: 01/07/2019

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.051	0.082	0.202	0.076
0.000	0.051	0.082	0.202	0.077
0.000	0.052	0.082	0.203	0.076
0.000	0.052	0.081	0.202	0.077
0.000	0.051	0.081	0.202	0.078
0.000	0.052	0.081	0.201	0.078
0.000	0.051	0.081	0.201	0.079
0.000	0.052	0.081	0.202	0.078
0.000	0.052	0.081	0.201	0.079
0.000	0.051	0.082	0.201	0.079
Standard Deviations	0.0005	0.0005	0.0006	0.0011

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

Baker
BK
1/7/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TJ Graham

THOMAS J GRAHAM
Signature and Printed Name

01/07/2019
Date

80-001286

11/3/19

Stability Checks

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2019
Software: 8100.27
SN 80-001286

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2019
Software: 8100.27
SN 80-001286

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2019
Software: 8100.27
SN 80-001286

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/03/2019
Software: 8100.27
SN 80-001286

Test	g/210L	Time
Air Blank	0.000	08:39
Control Test	0.047	08:40
Air Blank	0.000	08:41
Control Test	0.046	08:41
Air Blank	0.000	08:42
Control Test	0.046	08:43
Air Blank	0.000	08:43
Control Test Stats		
Average	0.0463	
Std Dev	0.0006	
Rel Std Dev(%)	1.2461	

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.076	08:45
Air Blank	0.000	08:46
Control Test	0.075	08:46
Air Blank	0.000	08:47
Control Test	0.076	08:47
Air Blank	0.000	08:48
Control Test Stats		
Average	0.0757	
Std Dev	0.0006	
Rel Std Dev(%)	0.7630	

Test	g/210L	Time
Air Blank	0.000	08:49
Control Test	0.196	08:50
Air Blank	0.000	08:50
Control Test	0.195	08:51
Air Blank	0.000	08:52
Control Test	0.194	08:52
Air Blank	0.000	08:53
Control Test Stats		
Average	0.1950	
Std Dev	0.0010	
Rel Std Dev(%)	0.5128	

Test	g/210L	Time
Air Blank	0.000	08:54
Control Test	0.076	08:54
Air Blank	0.000	08:55
Control Test	0.076	08:55
Air Blank	0.000	08:56
Control Test	0.076	08:56
Air Blank	0.000	08:56
Control Test Stats		
Average	0.0760	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD

Operator's Signature

JD

Operator's Signature

JD

Operator's Signature

JD

Operator's Signature

OPM

13X17/19

DS



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001286, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001286</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/07/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:56</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/07/2019

Date


THOMAS J GRAHAM,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1



80-001286
Calibration
Adjustment

```
<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.6760 (-0.0020)
Sample #2 = 1.6460 (0.0320)
Sample #3 = 1.6340 (0.0340)
Sample #4 = 1.6550 (0.0340)
Avg % Abs = 1.6450 (0.0333)
STD DEV = 0.0105 (0.0012)
REL STD DEV = 0.640 (3.464)
```

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
SN 80-001286
01/07/2019 07:48:34

Auto Calibration
Max Power Res Value = 40
Auto Range Res Value = 29

```
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/L %22%
Samples Taken = 4, Discarded = 1
Sum Lo = 12849, Sum Hi = 13210

<<<< CHANNEL 1 >>>>

Sample   % Abs Rel
Sample H1 = 0.2460 [-0.0050]
Sample H2 = 0.2200 [0.0360]
Sample H3 = 0.2240 [0.0630]
Sample H4 = 0.2340 [0.0920]
Avg % Abs = 0.2260 [0.0637]
STD DEV = 0.0072 [0.0280]
REL STD DEV = 3.191 [43.988]
```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 0.3070      (-0.0050)
Sample #2 = 0.2920      (0.0300)
Sample #3 = 0.2720      (0.0420)
Sample #4 = 0.2700      (0.0570)
Avg % Abs = 0.2780      (0.0430)
STD DEV = 0.0122      (0.0135)
REL STD DEV = 4.376      (3.460)

```

```
Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12833, Sum Io = 13201
<<<< CHANNEL 1 >>>>
```

Sample	% Abs	(% Abs Ref)
Sample #1	0.9340	(0.0000)
Sample #2	0.9510	(0.0300)
Sample #3	0.9380	(0.0430)
Sample #4	0.9430	(0.0350)
Avg % Abs =	0.9440	(0.0360)
STD DEV =	0.0066	(0.0066)
REL STD DEV =	0.695	(0.215)

6/12/19 JST
Rm

```
<<<<< CHANNEL 2
Sample      % Abs
Sample #1 = 1.6760
Sample #2 = 1.6460
Sample #3 = 1.6340
Sample #4 = 1.6550
Ave % Abs = 1.6450
STD DEV = 0.0105
REL STD DEV = 0.640
```

```

Vol Value = 0.100 g/21.0L ***
Tit value = 0.4762 mg/1 %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12821, Sum Io = 13195
<<<< CHANNEL 1 >>>>>
Sample % Abs
Sample #1 = 1.9810 (% Abs Ref)
Sample #2 = 2.0040 (-0.0110)
Sample #3 = 2.0050 (0.0080)
Sample #4 = 1.9750 (0.0070)
Avg % Abs = 1.9960 (0.0217)
STD DEV = 0.0147 (0.0245)
REL STD DEV = 0.738 (113.273)

```

```

<<<< CHANNEL 2 >>>>
(% Abs Ref)
Sample      % Abs
Sample #1 = 3.6670
Sample #2 = 3.6770
Sample #3 = 3.6820
Sample #4 = 3.6460
Avg % Abs = 3.6650 (0.0277)
STD DEV = 0.0164 (0.0095)
REL STD DEV = 0.5100 (34.353)

```

```

col Value = 0.200 g/210L ***
tit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum to = 12814, Num to = 13191
<<<<< CHANNEL 1 >>>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.7080      (-0.0050)
Sample #2 = 3.6760      (0.0150)
Sample #3 = 3.6690      (0.0370)
Sample #4 = 3.6720      (0.0340)
Avg % Abs = 3.6723      (0.0287)
STD DEV = 0.0035      (0.0119)
REL STD DEV = 0.096      (41.618)

```

```

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
1601 Ual = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.226
1st Std Dev = 0.01 Rel Std Dev = 3.19
1601 Ual = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.944
1st Std Dev = 0.01 Rel Std Dev = 0.69
1601 Ual = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.996
1st Std Dev = 0.01 Rel Std Dev = 0.74
1601 Ual = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.672
1st Std Dev = 0.00 Rel Std Dev = 0.10
1601 Ual = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.328
1st Std Dev = 0.01 Rel Std Dev = 0.12
Zero Order Coef = -613.51
First Order Coef = 2669.71
Second Order Coef = 27.93
Standard Deviation = 21.541843

```

```

<<<<< CHANNEL 2 >>>>>
1001 Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.278
Std Dev = 0.01 Rel Std Dev = 4.38
1002 Val = 0.1915 mg/l or 0.040 g/210L
% Abs = 1.645
Std Dev = 0.01 Rel Std Dev = 0.64
1003 Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.669
Std Dev = 0.02 Rel Std Dev = 0.20
1004 Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.782
Std Dev = 0.01 Rel Std Dev = 0.14
1005 Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.629
Std Dev = 0.02 Rel Std Dev = 0.17
Zero Order Coef = -389.85
First Order Coef = 1369.25
Second Order Coef = 12.73
Standard Deviation = 31.372949

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      [% Abs Ref]
Sample #1 = 6.8620   (-0.0120)
Sample #2 = 6.7930   (0.0410)
Sample #3 = 6.7760   (0.0560)
Sample #4 = 6.7770   (0.0470)
Avg % Abs = 6.7820   (0.0480)
STD DEV = 0.0095   (0.0075)
REL STD DEV = 0.141   (15.729)
-----

```

```

!o! Value = 0.300 g/210L ***
tit value = 1.4266 mg/l 22%
Samples Taken = 4, Discarded = 1
sum lo = 12808, sum lo = 13187
<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 5.3730      (-0.0170)
Sample #2 = 5.3200      (0.0280)
Sample #3 = 5.3320      (0.0440)
Sample #4 = 5.3310      (0.0710)
Avg % Abs = 5.3277      (0.0477)
STD DEV = 0.0067      (0.0217)
REF STD DEV = 0.125      (45.594)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 9.9160    (-0.13000)
Sample #2 = 9.8440    (0.0560)
Sample #3 = 9.8110    (0.1020)
Sample #4 = 9.8320    (0.0930)
Avg. % Abs = 9.8293  (0.0377)
STD DEV = 0.0168    (0.0244)
REL STD DEV = 0.171 (29.138)

```

Solution Stats Quadratic Fit Chan 1			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	-0.000	0.0003	
0.040	0.040	-0.0002	
0.100	0.101	-0.0005	
0.200	0.199	0.0006	
0.300	0.300	-0.0002	

Solution Stats Quadratic Fit Chan 2			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	-0.000	0.0002	
0.040	0.040	0.0002	
1.100	0.101	-0.0009	
2.200	0.199	0.0009	
3.300	0.300	-0.0003	

```

***
g/dl Value = 0.080 g/210L ***
tit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2667.00
Sample #2 = 2706.00
Sample #3 = 2769.00
Sample #4 = 2748.00
Average Result = 2739.5667
STD DEV = 31.7228
REL STD DEV = 1.158

```

```
***** CHANNEL 2
Sample #1 = 3059.00
Sample #2 = 3101.00
Sample #3 = 3132.00
Sample #4 = 3131.00
Average Result = 3121.3333
STD DEV = 17.6163
EL STD DEV = 0.564
```

```
*****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1022
3 um H2O Adjust (mg/1*10,000) = 1070
9 um H2O Adjust (mg/1*10,000) = 688
**** AUTO CAL PASS
```

80-001284

Post Cal. Adj. Stabilities

1/7/19
JD

INTOXILYZER 8000
Instrument Initialization
08:31 01/07/2019

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:00
Control Test	0.081	09:01
Air Blank	0.000	09:02
Control Test	0.080	09:02
Air Blank	0.000	09:03
Control Test	0.081	09:03
Air Blank	0.000	09:04
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

JD
Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.202	09:10
Air Blank	0.000	09:10
Control Test	0.201	09:11
Air Blank	0.000	09:11
Control Test	0.200	09:12
Air Blank	0.000	09:13
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

JD
Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.079	09:05
Air Blank	0.000	09:06
Control Test	0.080	09:06
Air Blank	0.000	09:07
Control Test	0.079	09:07
Air Blank	0.000	09:08
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

JD
Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.052	08:56
Air Blank	0.000	08:57
Control Test	0.051	08:58
Air Blank	0.000	08:58
Control Test	0.051	08:59
Air Blank	0.000	08:59
Control Test Stats		
Average	0.0513	
Std Dev	0.0006	
Rel Std Dev(%)	1.1247	

JD
Operator's Signature

PK 1/7/19

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 08:06

Date of Inspection: 01/03/2019

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

egm
BK
1/7/19

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas J. Graham

Signature and Printed Name

THOMAS J GRAHAM

01/03/2019
Date