



INSTRUMENT PROCESSING SHEET

Agency Bradford County SOS/N 80-001290Florida Department of
Law EnforcementDate In 03/29/2019 DI Completion Date 04/08/2019☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>KP</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date FDLE APR 10 2019 Alcohol Testing Program	Quality Checks Performed By <u>JO</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>122</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.132</u> (.139 - .169) 36 mm <u>.152</u> (.156 - .190) 53 mm <u>.222</u> (.228 - .278) 103 mm <u>.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	SD1013	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Flow Calibration Performed By <u>JO</u> Flow Column # <u>ATP-102</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>122</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>JO</u> ☒ Lab Temp °C <u>21.2</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>SD1012</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1013</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	SD1012	201707D 07/25/2019																																																												
0.080	DR1279	201707E 07/25/2019																																																												
0.200	SD1013	201707C 07/24/2019																																																												
0.080 DGS	N/A	AG805701 02/26/2020																																																												
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>JO</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1014</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☒ Flow Calibration ☒ Form 40 ☐ Other _____		Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	G2408																																																													
Interferent	G2882																																																													
0.050	SD1012																																																													
0.080	DR1279																																																													
0.200	SD1013																																																													
Notes/Suggested Service: <u>I placed a service center sticker on the instrument. I did not open the instrument. JO</u> _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>SP 4/10/19</u> Tech Review / Date <u>Brett Kirkland 4/10/19</u> Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BRADFORD COUNTY SO

Time of Inspection: 11:30

Date of Inspection: 04/08/2019

Serial Number: 80-001290

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.050	0.081	0.198	0.078
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.081	0.199	0.078
0.000	0.051	0.081	0.199	0.079
0.000	0.050	0.081	0.199	0.079
0.000	0.051	0.081	0.199	0.078
0.000	0.051	0.081	0.199	0.079
0.000	0.051	0.081	0.199	0.078

Standard Deviations	0.0004	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP
BK
4/10/19

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



THOMAS J GRAHAM
Signature and Printed Name

04/08/2019
Date

80-001290

4/8/19
JD

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
04/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.051	09:33
Air Blank	0.000	09:33
Control Test	0.050	09:34
Air Blank	0.000	09:35
Control Test	0.050	09:35
Air Blank	0.000	09:36
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

JD

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
04/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.081	09:38
Air Blank	0.000	09:38
Control Test	0.080	09:39
Air Blank	0.000	09:40
Control Test	0.081	09:40
Air Blank	0.000	09:41
Control Test Stats		
Average	0.0907	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

JD

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
04/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:42
Control Test	0.197	09:43
Air Blank	0.000	09:43
Control Test	0.198	09:44
Air Blank	0.000	09:44
Control Test	0.199	09:45
Air Blank	0.000	09:46
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

JD

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
04/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	0.080	09:47
Air Blank	0.000	09:47
Control Test	0.080	09:48
Air Blank	0.000	09:48
Control Test	0.079	09:49
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

JD

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/08/2019
Software: 8100.27
Flow Rate Calibration*****
1: Rate (Liters/min) = 5
2: Rate (Liters/min) = 15
3: Rate (Liters/min) = 30
SORT(Diff) = 20.688
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 664
Rounded Intercept = -482897
Correlation = 0.99764

Flow Calibration
4/8/19
JD

SP BK
4/10/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001290, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001290</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BRADFORD COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/08/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:30</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

04/08/2019

Date

THOMAS J GRAHAM,
Department Inspector

Service • Integrity • Respect • Quality

Page 1 of 1

SP
BK
4/10/19

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BRADFORD COUNTY SO
Time of Inspection: 09:27

Date of Inspection: 04/08/2019

Serial Number: 80-001290
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

THOMAS J GRAHAM

04/08/2019
Date