



INSTRUMENT PROCESSING SHEET

Agency Madison CountyS/N 80-001307Florida Department of
Law EnforcementDate In 7/17/2019DI Completion Date 7/18/19 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>165</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.158</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date <div style="text-align: center;"> FDLE JUL 18 2019 Alcohol Testing Program </div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201707D 7/25/19</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201707E 7/25/19</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201707C 7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG916501 6/14/21</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1018	201707D 7/25/19	0.080	SD3962	201707E 7/25/19	0.200	G2078	201707C 7/24/19	0.080 DGS	N/A	AG916501 6/14/21	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.0</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD1018</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>G2078</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD1018	201707D 7/25/19																																																											
0.080	SD3962	201707E 7/25/19																																																											
0.200	G2078	201707C 7/24/19																																																											
0.080 DGS	N/A	AG916501 6/14/21																																																											
Calibration Adjustment Performed By <u>[Signature]</u> Barometric Pressure Gauge <u>1017</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G8144</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2403</td> <td>19080</td> <td>3/4/21</td> </tr> <tr> <td>0.100</td> <td>G2879</td> <td>18206</td> <td>7/3/20</td> </tr> <tr> <td>0.200</td> <td>SD1019</td> <td>19046</td> <td>1/29/21</td> </tr> <tr> <td>0.300</td> <td>G8149</td> <td>18110</td> <td>4/2/20</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>8/5/19</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1018</td> <td>201707D</td> <td>7/25/19</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> <td>201707E</td> <td>7/25/19</td> </tr> <tr> <td>0.200</td> <td>G2078</td> <td>201707C</td> <td>7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG916501</td> <td>6/14/21</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G8144	N/A	N/A	0.040	G2403	19080	3/4/21	0.100	G2879	18206	7/3/20	0.200	SD1019	19046	1/29/21	0.300	G8149	18110	4/2/20	0.080 DGS	N/A	17817080A2	8/5/19	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1018	201707D	7/25/19	0.080	SD3962	201707E	7/25/19	0.200	G2078	201707C	7/24/19	0.080 DGS	N/A	AG916501	6/14/21	Department Inspection Performed By <u>[Signature]</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1017</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> </tr> <tr> <td>Interferent</td> <td>G6621</td> </tr> <tr> <td>0.050</td> <td>SD1018</td> </tr> <tr> <td>0.080</td> <td>SD3962</td> </tr> <tr> <td>0.200</td> <td>G2078</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G4444	Interferent	G6621	0.050	SD1018	0.080	SD3962	0.200	G2078
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G8144	N/A	N/A																																																										
0.040	G2403	19080	3/4/21																																																										
0.100	G2879	18206	7/3/20																																																										
0.200	SD1019	19046	1/29/21																																																										
0.300	G8149	18110	4/2/20																																																										
0.080 DGS	N/A	17817080A2	8/5/19																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD1018	201707D	7/25/19																																																										
0.080	SD3962	201707E	7/25/19																																																										
0.200	G2078	201707C	7/24/19																																																										
0.080 DGS	N/A	AG916501	6/14/21																																																										
Simulator	Serial Number																																																												
0.000	G4444																																																												
Interferent	G6621																																																												
0.050	SD1018																																																												
0.080	SD3962																																																												
0.200	G2078																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>SP 7/18/19</u> <u>[Signature]</u> <u>7/18/19</u> Tech Review / Date Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MADISON COUNTY SO
Time of Inspection: 12:54

Date of Inspection: 07/18/2019

Serial Number: 80-001307
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.079	0.199	0.079

Standard Deviations	0.0004	0.0005	0.0008	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

SP BK
7/18/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

07/18/2019
Date

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:00
Control Test	0.049	08:01
Air Blank	0.000	08:01
Control Test	0.048	08:02
Air Blank	0.000	08:03
Control Test	0.049	08:03
Air Blank	0.000	08:04
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

P Murphy
Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:05
Control Test	0.080	08:06
Air Blank	0.000	08:06
Control Test	0.080	08:07
Air Blank	0.000	08:08
Control Test	0.079	08:08
Air Blank	0.000	08:09
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

P Murphy
Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:10
Control Test	0.194	08:11
Air Blank	0.000	08:12
Control Test	0.196	08:12
Air Blank	0.000	08:13
Control Test	0.197	08:13
Air Blank	0.000	08:14
Control Test Stats		
Average	0.1957	
Std Dev	0.0015	
Rel Std Dev(%)	0.7807	

P Murphy
Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:40
Control Test	0.077	07:40
Air Blank	0.000	07:41
Control Test	0.078	07:41
Air Blank	0.000	07:42
Control Test	0.077	07:42
Air Blank	0.000	07:43
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

DGS

P Murphy
Operator's Signature

SP
BK
7/18/19

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019 09:54:23

Auto Calibration
Max Power Res Value = 41
Auto Range Res Value = 34

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12790, Sum lo = 12693

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0840 (-0.0300)
Sample #2 = 0.1240 (-0.0150)
Sample #3 = 0.1140 (-0.0190)
Sample #4 = 0.1110 (-0.0130)
Avg % Abs = 0.1163 (-0.0157)
STD DEV = 0.0068 (0.0031)
REL STD DEV = 5.851 (19.500)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0930 (-0.0210)
Sample #2 = 0.1170 (-0.0070)
Sample #3 = 0.1020 (-0.0080)
Sample #4 = 0.1060 (-0.0100)
Avg % Abs = 0.1083 (-0.0083)
STD DEV = 0.0078 (0.0015)
REL STD DEV = 7.170 (18.330)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12791, Sum lo = 12692

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8650 (-0.0060)
Sample #2 = 0.8550 (0.0300)
Sample #3 = 0.8640 (0.0110)
Sample #4 = 0.8620 (0.0410)
Avg % Abs = 0.8603 (0.0273)
STD DEV = 0.0047 (0.0152)
REL STD DEV = 0.549 (55.525)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5270 (-0.0230)
Sample #2 = 1.5480 (-0.0040)
Sample #3 = 1.5460 (-0.0180)
Sample #4 = 1.5290 (0.0120)
Avg % Abs = 1.5410 (-0.0033)
STD DEV = 0.0104 (0.0150)
REL STD DEV = 0.678 (450.333)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12787, Sum lo = 12690

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9440 (-0.0110)
Sample #2 = 1.9550 (0.0230)
Sample #3 = 1.9770 (0.0010)
Sample #4 = 1.9560 (0.0060)
Avg % Abs = 1.9627 (0.0100)
STD DEV = 0.0124 (0.0115)
REL STD DEV = 0.633 (115.326)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6030 (-0.0100)
Sample #2 = 3.6540 (-0.0120)
Sample #3 = 3.6400 (-0.0080)
Sample #4 = 3.6290 (-0.0050)
Avg % Abs = 3.6410 (-0.0083)
STD DEV = 0.0125 (0.0035)
REL STD DEV = 0.344 (42.143)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12781, Sum lo = 12686

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6640 (-0.0260)
Sample #2 = 3.6730 (-0.0300)
Sample #3 = 3.6780 (0.0040)
Sample #4 = 3.7030 (-0.0180)
Avg % Abs = 3.6847 (-0.0147)
STD DEV = 0.0161 (0.0172)
REL STD DEV = 0.436 (117.566)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.7930 (-0.0250)
Sample #2 = 6.8500 (-0.0400)
Sample #3 = 6.8530 (-0.0330)
Sample #4 = 6.8580 (-0.0380)
Avg % Abs = 6.8537 (-0.0370)
STD DEV = 0.0040 (0.0036)
REL STD DEV = 0.059 (9.745)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12783, Sum lo = 12689

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.3470 (-0.0180)
Sample #2 = 5.3320 (0.0030)
Sample #3 = 5.3470 (0.0050)
Sample #4 = 5.3230 (0.0290)
Avg % Abs = 5.3340 (0.0123)
STD DEV = 0.0121 (0.0145)
REL STD DEV = 0.227 (117.311)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.8590 (-0.0160)
Sample #2 = 9.8730 (-0.0020)
Sample #3 = 9.8810 (0.0010)
Sample #4 = 9.8660 (0.0180)
Avg % Abs = 9.8733 (0.0057)
STD DEV = 0.0075 (0.0108)
REL STD DEV = 0.076 (190.338)

Auto Cal Data:
Channel 1 data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.116
Std Dev = 0.01 Rel Std Dev = 5.85
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.860
Std Dev = 0.00 Rel Std Dev = 0.55
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.963
Std Dev = 0.01 Rel Std Dev = 0.63
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.685
Std Dev = 0.02 Rel Std Dev = 0.44
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.334
Std Dev = 0.01 Rel Std Dev = 0.23
Zero Order Coef = -287.94
First Order Coef = 2495.13
Second Order Coef = 44.58
Standard Deviation = 13.387872

Channel 2 data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.108
Std Dev = 0.01 Rel Std Dev = 7.17
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.541
Std Dev = 0.01 Rel Std Dev = 0.68
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.641
Std Dev = 0.01 Rel Std Dev = 0.34
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.854
Std Dev = 0.00 Rel Std Dev = 0.06
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.873
Std Dev = 0.01 Rel Std Dev = 0.08
Zero Order Coef = -134.99
First Order Coef = 1286.00
Second Order Coef = 17.71
Standard Deviation = 14.684698

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0001
0.040 0.040 0.0003
0.100 0.100 -0.0004
0.200 0.200 0.0003
0.300 0.300 -0.0001

SP 13K 7/18/19

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0003
0.100	0.100	-0.0004
0.200	0.200	0.0003
0.300	0.300	-0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1
 Sample #1 = 2928.00
 Sample #2 = 2957.00
 Sample #3 = 2987.00
 Sample #4 = 2862.00
 Average Result = 2935.3333
 STD DEV = 65.2559
 REL STD DEV = 2.223

 ***** CHANNEL 2
 Sample #1 = 3315.00
 Sample #2 = 3314.00
 Sample #3 = 3296.00
 Sample #4 = 3276.00
 Average Result = 3295.3333
 STD DEV = 19.0088
 REL STD DEV = 0.577

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1017
 3 um H2O Adjust (mg/l*10,000) = 874
 9 um H2O Adjust (mg/l*10,000) = 514
 **** AUTO CAL PASS

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.050	10:39
Air Blank	0.000	10:40
Control Test	0.049	10:40
Air Blank	0.000	10:41
Control Test	0.049	10:42
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

P. Murphy
 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.080	10:47
Air Blank	0.000	10:47
Control Test	0.080	10:48
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:50
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:57
Control Test	0.079	10:58
Air Blank	0.000	10:58
Control Test	0.079	10:58
Air Blank	0.000	10:59
Control Test	0.079	10:59
Air Blank	0.000	11:00
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P. Murphy
 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:52
Control Test	0.198	10:53
Air Blank	0.000	10:53
Control Test	0.199	10:54
Air Blank	0.000	10:55
Control Test	0.198	10:55
Air Blank	0.000	10:56
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

P. Murphy
 Operator's Signature

80-001307

Post Cal Adjust
 Stabilities

SP BK
 7/18/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001307, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001307</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MADISON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/18/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:54</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/18/2019

Date

Patrick J. Murphy

PATRICK J MURPHY,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

SP
BK
7/18/19



INSTRUMENT PROCESSING SHEET

Agency Madison CountyS/N 80-001307

Florida Department of Law Enforcement

Date In 7/17/2019DI Completion Date 7/18/19 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>165</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.158</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date <div style="text-align: center;"> FDLE JUL 18 2019 Alcohol Testing Program </div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u> <u>6/14/21</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>SD3962</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>G2078</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG916501</u> <u>6/14/21</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.0</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD1018</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>G2078</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1018</u>	<u>201707D</u> <u>7/25/19</u>																																																											
0.080	<u>SD3962</u>	<u>201707E</u> <u>7/25/19</u>																																																											
0.200	<u>G2078</u>	<u>201707C</u> <u>7/24/19</u>																																																											
0.080 DGS	N/A	<u>AG916501</u> <u>6/14/21</u>																																																											
Calibration Adjustment Performed By <u>[Signature]</u> Barometric Pressure Gauge <u>1017</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G8144</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>G2403</u></td> <td><u>19080</u></td> <td><u>3/4/21</u></td> </tr> <tr> <td>0.100</td> <td><u>G2879</u></td> <td><u>18206</u></td> <td><u>7/3/20</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1019</u></td> <td><u>19046</u></td> <td><u>1/29/21</u></td> </tr> <tr> <td>0.300</td> <td><u>G8149</u></td> <td><u>18110</u></td> <td><u>4/2/20</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817080A2</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u></td> <td><u>6/14/21</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G8144</u>	N/A	N/A	0.040	<u>G2403</u>	<u>19080</u>	<u>3/4/21</u>	0.100	<u>G2879</u>	<u>18206</u>	<u>7/3/20</u>	0.200	<u>SD1019</u>	<u>19046</u>	<u>1/29/21</u>	0.300	<u>G8149</u>	<u>18110</u>	<u>4/2/20</u>	0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1018</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>SD3962</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>G2078</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A	<u>AG916501</u>	<u>6/14/21</u>	Department Inspection Performed By <u>[Signature]</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1017</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G4444</u></td> </tr> <tr> <td>Interferent</td> <td><u>G6621</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1018</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G4444</u>	Interferent	<u>G6621</u>	0.050	<u>SD1018</u>	0.080	<u>SD3962</u>	0.200	<u>G2078</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>G8144</u>	N/A	N/A																																																										
0.040	<u>G2403</u>	<u>19080</u>	<u>3/4/21</u>																																																										
0.100	<u>G2879</u>	<u>18206</u>	<u>7/3/20</u>																																																										
0.200	<u>SD1019</u>	<u>19046</u>	<u>1/29/21</u>																																																										
0.300	<u>G8149</u>	<u>18110</u>	<u>4/2/20</u>																																																										
0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>SD1018</u>	<u>201707D</u>	<u>7/25/19</u>																																																										
0.080	<u>SD3962</u>	<u>201707E</u>	<u>7/25/19</u>																																																										
0.200	<u>G2078</u>	<u>201707C</u>	<u>7/24/19</u>																																																										
0.080 DGS	N/A	<u>AG916501</u>	<u>6/14/21</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>G4444</u>																																																												
Interferent	<u>G6621</u>																																																												
0.050	<u>SD1018</u>																																																												
0.080	<u>SD3962</u>																																																												
0.200	<u>G2078</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>SP 7/18/19</u> <u>[Signature]</u> <u>7/18/19</u> Tech Review / Date Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MADISON COUNTY SO
Time of Inspection: 12:54

Date of Inspection: 07/18/2019

Serial Number: 80-001307
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.079	0.199	0.079

Standard Deviations	0.0004	0.0005	0.0008	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

SP BK
7/18/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy
PATRICK J MURPHY

Signature and Printed Name

07/18/2019
Date

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:00
Control Test	0.049	08:01
Air Blank	0.000	08:01
Control Test	0.048	08:02
Air Blank	0.000	08:03
Control Test	0.049	08:03
Air Blank	0.000	08:04
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

P. Murphy

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:05
Control Test	0.080	08:06
Air Blank	0.000	08:06
Control Test	0.080	08:07
Air Blank	0.000	08:08
Control Test	0.079	08:08
Air Blank	0.000	08:09
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

P. Murphy

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:10
Control Test	0.194	08:11
Air Blank	0.000	08:12
Control Test	0.196	08:12
Air Blank	0.000	08:13
Control Test	0.197	08:13
Air Blank	0.000	08:14
Control Test Stats		
Average	0.1957	
Std Dev	0.0015	
Rel Std Dev(%)	0.7807	

P. Murphy

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:40
Control Test	0.077	07:40
Air Blank	0.000	07:41
Control Test	0.078	07:41
Air Blank	0.000	07:42
Control Test	0.077	07:42
Air Blank	0.000	07:43
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

DGS

SP
BK
7/18/19

P. Murphy

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
07/18/2019 09:54:23

Auto Calibration
Max Power Res Value = 41
Auto Range Res Value = 34

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12790, Sum lo = 12693

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0840 (-0.0300)
Sample #2 = 0.1240 (-0.0150)
Sample #3 = 0.1140 (-0.0190)
Sample #4 = 0.1110 (-0.0130)
Avg % Abs = 0.1163 (-0.0157)
STD DEU = 0.0068 (0.0031)
REL STD DEU = 5.851 (19.500)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0930 (-0.0210)
Sample #2 = 0.1170 (-0.0070)
Sample #3 = 0.1020 (-0.0080)
Sample #4 = 0.1060 (-0.0100)
Avg % Abs = 0.1083 (-0.0083)
STD DEU = 0.0078 (0.0015)
REL STD DEU = 7.170 (18.330)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12791, Sum lo = 12692

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8650 (-0.0060)
Sample #2 = 0.8550 (0.0300)
Sample #3 = 0.8640 (0.0110)
Sample #4 = 0.8620 (0.0410)
Avg % Abs = 0.8603 (0.0273)
STD DEU = 0.0047 (0.0152)
REL STD DEU = 0.549 (55.525)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5270 (-0.0230)
Sample #2 = 1.5480 (-0.0040)
Sample #3 = 1.5460 (-0.0180)
Sample #4 = 1.5290 (0.0120)
Avg % Abs = 1.5410 (-0.0033)
STD DEU = 0.0104 (0.0150)
REL STD DEU = 0.678 (450.333)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12787, Sum lo = 12690

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9440 (-0.0110)
Sample #2 = 1.9550 (0.0230)
Sample #3 = 1.9770 (0.0010)
Sample #4 = 1.9560 (0.0060)
Avg % Abs = 1.9627 (0.0100)
STD DEU = 0.0124 (0.0115)
REL STD DEU = 0.633 (115.326)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6030 (-0.0100)
Sample #2 = 3.6540 (-0.0120)
Sample #3 = 3.6400 (-0.0080)
Sample #4 = 3.6290 (-0.0050)
Avg % Abs = 3.6410 (-0.0083)
STD DEU = 0.0125 (0.0035)
REL STD DEU = 0.344 (42.143)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12781, Sum lo = 12686

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6640 (-0.0260)
Sample #2 = 3.6730 (-0.0300)
Sample #3 = 3.6780 (0.0040)
Sample #4 = 3.7030 (-0.0180)
Avg % Abs = 3.6847 (-0.0147)
STD DEU = 0.0161 (0.0172)
REL STD DEU = 0.436 (117.566)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.7930 (-0.0250)
Sample #2 = 6.8500 (-0.0400)
Sample #3 = 6.8530 (-0.0330)
Sample #4 = 6.8580 (-0.0380)
Avg % Abs = 6.8537 (-0.0370)
STD DEU = 0.0040 (0.0036)
REL STD DEU = 0.059 (9.745)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12783, Sum lo = 12689

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.3470 (-0.0180)
Sample #2 = 5.3320 (0.0030)
Sample #3 = 5.3470 (0.0050)
Sample #4 = 5.3230 (0.0290)
Avg % Abs = 5.3340 (0.0123)
STD DEU = 0.0121 (0.0145)
REL STD DEU = 0.227 (117.311)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.8590 (-0.0160)
Sample #2 = 9.8730 (-0.0020)
Sample #3 = 9.8810 (0.0010)
Sample #4 = 9.8660 (0.0180)
Avg % Abs = 9.8733 (0.0057)
STD DEU = 0.0075 (0.0108)
REL STD DEU = 0.076 (190.338)

Auto Cal Data:
Channel 1 data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.116
Std Dev = 0.01 Rel Std Dev = 5.85
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.860
Std Dev = 0.00 Rel Std Dev = 0.55
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.963
Std Dev = 0.01 Rel Std Dev = 0.63
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.685
Std Dev = 0.02 Rel Std Dev = 0.44
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.334
Std Dev = 0.01 Rel Std Dev = 0.23
Zero Order Coef = -287.94
First Order Coef = 2495.13
Second Order Coef = 44.58
Standard Deviation = 13.387872

Channel 2 data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.108
Std Dev = 0.01 Rel Std Dev = 7.17
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.541
Std Dev = 0.01 Rel Std Dev = 0.68
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.641
Std Dev = 0.01 Rel Std Dev = 0.34
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.854
Std Dev = 0.00 Rel Std Dev = 0.06
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.873
Std Dev = 0.01 Rel Std Dev = 0.08
Zero Order Coef = -134.99
First Order Coef = 1286.00
Second Order Coef = 17.71
Standard Deviation = 14.684698

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0001
0.040 0.040 0.0003
0.100 0.100 -0.0004
0.200 0.200 0.0003
0.300 0.300 -0.0001

SP 13K 7/18/19

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0003
0.100	0.100	-0.0004
0.200	0.200	0.0003
0.300	0.300	-0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1
 Sample #1 = 2928.00
 Sample #2 = 2957.00
 Sample #3 = 2987.00
 Sample #4 = 2862.00
 Average Result = 2935.3333
 STD DEV = 65.2559
 REL STD DEV = 2.223

 ***** CHANNEL 2
 Sample #1 = 3315.00
 Sample #2 = 3314.00
 Sample #3 = 3296.00
 Sample #4 = 3276.00
 Average Result = 3295.3333
 STD DEV = 19.0088
 REL STD DEV = 0.577

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1017
 3 um H2O Adjust (mg/l*10,000) = 874
 9 um H2O Adjust (mg/l*10,000) = 514
 **** AUTO CAL PASS

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.050	10:39
Air Blank	0.000	10:40
Control Test	0.049	10:40
Air Blank	0.000	10:41
Control Test	0.049	10:42
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

P. Murphy
 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.080	10:47
Air Blank	0.000	10:47
Control Test	0.080	10:48
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:50
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:57
Control Test	0.079	10:58
Air Blank	0.000	10:58
Control Test	0.079	10:58
Air Blank	0.000	10:59
Control Test	0.079	10:59
Air Blank	0.000	11:00
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P. Murphy
 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 07/18/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:52
Control Test	0.198	10:53
Air Blank	0.000	10:53
Control Test	0.199	10:54
Air Blank	0.000	10:55
Control Test	0.198	10:55
Air Blank	0.000	10:56
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

P. Murphy
 Operator's Signature

80-001307

Post Cal Adjust
 Stabilities

SP BK
 7/18/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001307, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001307</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MADISON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/18/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:54</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/18/2019

Date

Patrick J. Murphy

PATRICK J MURPHY,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

SP
BK
7/18/19



INSTRUMENT PROCESSING SHEET

Agency Madison County

S/N 80-001307

Florida Department of Law Enforcement

Date In 5/22/2019

DI Completion Date 5/28/19

Ship P/U H/D CMI EE

Intake Performed By <u>DP</u>		Quality Checks Performed By <u>PSM</u>		Flow Calibration Performed By _____																
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>165</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
Final Release Date		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>2017070</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG831804</u> <u>11/14/20</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>2017070</u> <u>7/25/19</u>	0.080	<u>SD3962</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>G2078</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG831804</u> <u>11/14/20</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																		
0.050	<u>SD1018</u>	<u>2017070</u> <u>7/25/19</u>																		
0.080	<u>SD3962</u>	<u>201707E</u> <u>7/25/19</u>																		
0.200	<u>G2078</u>	<u>201707C</u> <u>7/24/19</u>																		
0.080 DGS	N/A	<u>AG831804</u> <u>11/14/20</u>																		
FDLE MAY 29 2019 Alcohol Testing Program		Temperature Checks Performed By <u>PSM</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.0</u> External Digital Therm. ID#: <u>300563</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1018</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2078</u>																		

Calibration Adjustment Performed By _____				Department Inspection Performed By <u>PSM</u>																																			
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28427</u>																																			
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1014</u> Instrument <u>1006</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u>							
Simulator	Serial Number	Lot Number	Expiration																																				
0.000		N/A	N/A																																				
0.040																																							
0.100																																							
0.200																																							
0.300																																							
0.080 DGS	N/A																																						
<input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G4444</u></td> </tr> <tr> <td>Interferent</td> <td><u>G6621</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1018</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	<u>G4444</u>	Interferent	<u>G6621</u>	0.050	<u>SD1018</u>	0.080	<u>SD3962</u>	0.200	<u>G2078</u>
Simulator	Serial Number	Lot Number	Expiration																																				
0.050																																							
0.080																																							
0.200																																							
0.080 DGS	N/A																																						
Simulator	Serial Number																																						
0.000	<u>G4444</u>																																						
Interferent	<u>G6621</u>																																						
0.050	<u>SD1018</u>																																						
0.080	<u>SD3962</u>																																						
0.200	<u>G2078</u>																																						
				Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																			

Notes/Suggested Service: _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	
_____		<u>SP 5/29/19</u> <u>J. Zahra</u> <u>5/29/19</u> Tech Review / Date Admin Review / Date	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MADISON COUNTY SO
Time of Inspection: 12:25

Date of Inspection: 05/28/2019

Serial Number: 80-001307
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.050	0.080	0.197	0.077
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.081	0.197	0.078
0.000	0.050	0.080	0.197	0.078
0.000	0.050	0.080	0.198	0.077
0.000	0.050	0.080	0.198	0.077
0.000	0.050	0.081	0.198	0.078

Standard Deviations	0.0003	0.0004	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

05/28/2019
Date

5/29/19
JP

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/28/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.049	09:58
Air Blank	0.000	09:59
Control Test	0.049	10:00
Air Blank	0.000	10:00
Control Test	0.050	10:01
Air Blank	0.000	10:01
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

P Murphy

 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/28/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:14
Control Test	0.078	10:14
Air Blank	0.000	10:14
Control Test	0.078	10:15
Air Blank	0.000	10:15
Control Test	0.079	10:16
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

P Murphy

 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/28/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.080	10:04
Air Blank	0.000	10:04
Control Test	0.080	10:05
Air Blank	0.000	10:05
Control Test	0.080	10:06
Air Blank	0.000	10:06
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

 Operator's Signature

MADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/28/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.194	10:09
Air Blank	0.000	10:10
Control Test	0.197	10:10
Air Blank	0.000	10:11
Control Test	0.197	10:11
Air Blank	0.000	10:12
Control Test Stats		
Average	0.1960	
Std Dev	0.0017	
Rel Std Dev(%)	0.8837	

P Murphy

 Operator's Signature

SP

5/28/19
SP



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001307, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001307</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MADISON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/28/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Patrick J Murphy
PATRICK J MURPHY

05/28/2019

Date

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

5/29/19
JD