FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Okeechobee PD S/N 80-001322

Date In 8/12/2019 DI Completion Date 8/15/19 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake — Performed By DP

- ☒ Annual
- ☐ Registration
- ☒ Return from CMI / EE

Visual Inspection:
- ☒ Case ☒ Handle
- ☒ Keyboard ☒ Dry Gas Shelf
- ☒ Feet ☒ Breath Tube
- ☒ Ports ☒ Screws Tight

Other Equipment/ Accessories:
- ☐ Power cord ☐ Printer Cable
- ☒ Static Bag ☐ 12V DC Cable

Notes: ______

Quality Checks — Performed By SP

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 179
- ☒ Flow Verification (L/s)

Flow Column # ATP102

32 mm	.152	(.139 - .169)
36 mm	.171	(.156 - .190)
53 mm	.242	(.228 - .278)
103 mm	.515	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 26932

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201905A 2019 5-14-21 SP
0.080	DR1279	201905B 5-14-21
0.200	SD1013	201904D 4-30-21
0.080 DGS	N/A	AG831804 11-14-20

Flow Calibration — Performed By ______

Flow Column # ______
- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value ______
- ☐ Post Calibration Verification (L/s)

Flow Column # ______

32 mm	______	(.139 - .169)
36 mm	______	(.156 - .190)
53 mm	______	(.228 - .278)
103 mm	______	(.447 - .547)

Maintenance — Performed By ______

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other ______

Temperature Checks — Performed By SP

- ☒ Lab Temp °C 21.2

External Digital Therm. ID#: 300503
- ☒ 34°C +-.2 Serial #: SD1018
- ☒ 34°C +-.2 Serial #: SD3962
- ☒ 34°C +-.2 Serial #: G2078

Final Release Date

FDLE

AUG 16 2019

Alcohol Testing Program

Calibration Adjustment — Performed By SP

Barometric Pressure Gauge 1011 ID # 28427

Simulator	Serial Number	Lot Number	Expiration
0.000	G8144	N/A	N/A
0.040	G2403	19080	3-4-21
0.100	G2879	18200	7-3-20
0.200	G3709	19040	1-29-21
0.300	G8149	18110	4-2-20
0.080 DGS	N/A SP	~~AG831804~~	~~11-14-20~~

☒ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	SD1012	201905A	5-14-21
0.080	DR1279	201905B	5-14-21
0.200	SD1013	201905D	4-30-21
0.080 DGS	N/A	AG831804	11-14-20

Notes/Suggested Service: ______

DGS Lot # 0881908~~0~~0AF Exp. 6-5-21

Department Inspection — Performed By SP

Barometric Pressure ID# 26932

Gauge 1010 Instrument 1011

Mouth Alcohol Solution Lot # 2018-B

Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	G4444
Interferent	G6621
0.050	SD1018
0.080	SD3962
0.200	G2078

Attachments

- ☒ Form 41 ☒ Post-Stability Checks
- ☒ Stability Checks ☐ Flow Calibration
- ☒ Calibration Certificate ☐ Form 40
- ☒ Calibration Adjustment ☐ Other ______

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

DPM 8/16/19 — Tech Review / Date

Brett Kirkland 8/16/19 — Admin Review / Date

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:23
Control Test	0.047	14:23
Air Blank	0.000	14:24
Control Test	0.047	14:25
Air Blank	0.000	14:25
Control Test	0.046	14:26
Air Blank	0.000	14:26
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	

SP

Operator's Signature

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:55
Control Test	0.075	13:55
Air Blank	0.000	13:56
Control Test	0.076	13:57
Air Blank	0.000	13:57
Control Test	0.076	13:58
Air Blank	0.000	13:58
Control Test Stats		
Average	0.0757	
Std Dev	0.0006	
Rel Std Dev(%)	0.7630	

SP

Operator's Signature

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:16
Control Test	0.196	14:17
Air Blank	0.000	14:18
Control Test	0.196	14:18
Air Blank	0.000	14:19
Control Test	0.197	14:20
Air Blank	0.000	14:20
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

SP

Operator's Signature

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.078	13:51
Air Blank	0.000	13:52
Control Test	0.078	13:52
Air Blank	0.000	13:53
Control Test	0.079	13:53
Air Blank	0.000	13:53
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

SP

Operator's Signature

BK
8/16/19

OKEECH[illegible]
Intoxil[illegible]
Model 8000 SN 80-001322
08/13/2019 14:27:03

Auto Calibration
Max Power Res Value = 98
Auto Range Res Value = 76

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12761, 9um Io = 13065
<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1010	(0.0080)
Sample #2 =	0.0880	(0.0230)
Sample #3 =	0.0980	(0.0660)
Sample #4 =	0.0950	(0.0950)

Avg % Abs = 0.0937 (0.0613)
STD DEV = 0.0051 (0.0362)
REL STD DEV = 5.479 (59.064)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.0670	(-0.0220)
Sample #2 =	0.0750	(-0.0070)
Sample #3 =	0.0700	(0.0140)
Sample #4 =	0.0730	(0.0110)

Avg % Abs = 0.0727 (0.0060)
STD DEV = 0.0025 (0.0114)
REL STD DEV = 3.463 (189.297)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12737, 9um Io = 13053
<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.8600	(-0.0130)
Sample #2 =	0.8230	(0.0400)
Sample #3 =	0.8390	(0.0240)
Sample #4 =	0.8410	(0.0350)

Avg % Abs = 0.8343 (0.0330)
STD DEV = 0.0099 (0.0082)
REL STD [illegible]

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5010	(-0.0150)
Sample #2 =	1.4840	(0.0110)
Sample #3 =	1.4810	(0.0050)
Sample #4 =	1.4800	(-0.0020)

Avg % Abs = 1.4817 (0.0047)
STD DEV = 0.0021 (0.0065)
REL STD DEV = 0.140 (139.423)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12733, 9um Io = 13053
<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	2.0110	(-0.0200)
Sample #2 =	1.9890	(0.0290)
Sample #3 =	1.9810	(0.0470)
Sample #4 =	1.9680	(0.0340)

Avg % Abs = 1.9793 (0.0367)
STD DEV = 0.0106 (0.0093)
REL STD DEV = 0.535 (25.341)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.5970	(0.0000)
Sample #2 =	3.5650	(0.0510)
Sample #3 =	3.5490	(0.0560)
Sample #4 =	3.5430	(0.0440)

Avg % Abs = 3.5523 (0.0503)
STD DEV = 0.0114 (0.0060)
REL STD DEV = 0.320 (11.976)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12724, 9um Io = 13048
<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.6580	(-0.0120)
Sample #2 =	3.7600	(0.0120)
Sample #3 =	3.7900	(0.0050)
Sample #4 =	3.7870	(0.0320)

Avg % Abs = 3.7790 (0.0163)
STD DEV = 0.0165 (0.0140)
REL STD [illegible]

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	6.6550	(-0.0090)
Sample #2 =	6.8230	(0.0200)
Sample #3 =	6.8760	(0.0150)
Sample #4 =	6.8860	(0.0190)

Avg % Abs = 6.8617 (0.0180)
STD DEV = 0.0339 (0.0026)
REL STD DEV = 0.493 (14.699)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12720, 9um Io = 13049
<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	5.5190	(-0.0240)
Sample #2 =	5.5210	(0.0210)
Sample #3 =	5.5470	(0.0260)
Sample #4 =	5.5160	(0.0500)

Avg % Abs = 5.5280 (0.0323)
STD DEV = 0.0166 (0.0155)
REL STD DEV = 0.301 (47.946)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	9.9050	(-0.0120)
Sample #2 =	9.8910	(0.0940)
Sample #3 =	9.9330	(0.0600)
Sample #4 =	9.8680	(0.0960)

Avg % Abs = 9.8973 (0.0833)
STD DEV = [illegible]
REL STD DEV [illegible]

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.094
Std Dev = 0.01 Rel Std Dev = 5.48
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.834
Std Dev = 0.01 Rel Std Dev = 1.18
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.979
Std Dev = 0.01 Rel Std Dev = 0.54
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.779
Std Dev = 0.02 Rel Std Dev = 0.44
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.528
Std Dev = 0.02 Rel Std Dev = 0.30
Zero Order Coef = -214.09
First Order Coef = 2472.47
Second Order Coef = 27.23
Standard Deviation = 24.076254

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.073
Std Dev = 0.00 Rel Std Dev = 3.46
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.482
Std Dev = 0.00 Rel Std Dev = 0.14
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.552
Std Dev = 0.01 Rel Std Dev = 0.32
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.862
Std Dev = 0.03 Rel Std Dev = 0.49
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.897
Std Dev = 0.03 Rel Std Dev = 0.33
Zero Order Coef = -74.74
First Order Coef = 1301.63
Second Order Coef = 14.94
Standard [illegible]

Solution Stats [illegible] Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.039	0.0008
0.100	0.101	-0.0005
0.200	0.200	0.0001
0.300	0.300	-0.0000

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.040	0.0004
0.100	0.099	0.0005
0.200	0.201	-0.0008
0.300	0.300	0.0003

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3109.00
Sample #2 = 3193.00
Sample #3 = 3236.00
Sample #4 = 3227.00
Average Result = 3218.6667
STD DEV = 22.6789
REL STD DEV = 0.705
****
***** CHANNEL 2
Sample #1 = 3417.00
Sample #2 = 3434.00
Sample #3 = 3423.00
Sample #4 = 3436.00
Average Result = 3431.0000
STD DEV = 7.0000
REL STD DEV = 0.204
****
Dry Gas H2O Adjust Results **********
Barometric Pressure = 1009
3 um H2O Adjust (mg/l*10,000) = 591
9 um H2O Adjust (mg/l*10,000) = 378
**** AUTO CAL PASS

CAL ADJUSTMENT
80-001322 SP

RPM
BK 8/16/19

Post Cal Adjust Stability Checks #80-001322

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:08
Control Test	0.048	16:09
Air Blank	0.000	16:10
Control Test	0.048	16:10
Air Blank	0.000	16:11
Control Test	0.048	16:12
Air Blank	0.000	16:12
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:04
Control Test	0.078	16:05
Air Blank	0.000	16:05
Control Test	0.078	16:06
Air Blank	0.000	16:07
Control Test	0.078	16:07
Air Blank	0.000	16:08
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:00
Control Test	0.195	16:00
Air Blank	0.000	16:01
Control Test	0.198	16:02
Air Blank	0.000	16:02
Control Test	0.197	16:03
Air Blank	0.000	16:03
Control Test Stats		
Average	0.1967	
Std Dev	0.0015	
Rel Std Dev(%)	0.7767	

SP
Operator's Signature

OKEECHOBEE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001322
08/13/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:13
Control Test	0.080	16:13
Air Blank	0.000	16:14
Control Test	0.080	16:14
Air Blank	0.000	16:15
Control Test	0.080	16:15
Air Blank	0.000	16:16
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP
Operator's Signature

PBM
13K
8/16/19

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE P.D.
Time of Inspection: 15:53 Date of Inspection: 08/15/2019
Serial Number: 80-001322
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.048	0.077	0.194	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080

Standard Deviations	0.0007	0.0009	0.0015	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

BK
8/16/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt SHAYLA D PLATT
Signature and Printed Name

08/15/2019
Date

FDLE/ATP Form 41 –Revised August 2005

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001322 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001322
Owning Agency:	OKEECHOBEE P.D.
Calibration Date:	08/15/2019
Calibration Time:	15:53

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

08/15/2019
Date

Shayla Platt
SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

ADAM
BK
8/16/19

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Raul Marrero on ~~6/6/19~~ 7/1/19 SP

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: ______

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001322

Bill To Address:	Ship to Address:
Okeechobee PD	FDLE-ATP Tallahassee

Reason for Return:

Instrument will not power on.

Please choose one of the following options:

☐ 1. I ______, authorize all repairs.

☐ 2. I ______, authorize repairs up to $______.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Raul Marrero

Phone #: 863-381-0988 Email: rmarrero@cityofokeechobee.com

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Okeechobee PD S/N 80-001322

Date In 5/30/2019 DI Completion Date ________ ❑Ship ❑P/U ❑H/D ❑CMI ❑EE

Intake Performed By [signature]

☑ Annual
❑ Registration
❑ Return from CMI / EE

Visual Inspection:

☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

❑ Power cord ❑ Printer Cable
❑ Static Bag ❑ 12V DC Cable

Notes: ________

Final Release Date

Quality Checks Performed By ________

❑ Breath Tube Screen
❑ Replace External O-Rings
❑ Instrument Set Up Verified
❑ R-Value ________
❑ Flow Verification (L/s)

Flow Column # ________

32 mm ________ (.139 - .169)
36 mm ________ (.156 - .190)
53 mm ________ (.228 - .278)
103 mm ________ (.447 - .547)

❑ Barometric Pressure Check

Gauge ID # ________

❑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050		
0.080		
0.200		
0.080 DGS	N/A	

Flow Calibration Performed By ________

Flow Column # ________

❑ 5L/min – 17mm
❑ 15L/min – 53mm
❑ 30L/min – 103mm

❑ R-Value ________
❑ Post Calibration Verification (L/s)

Flow Column # ________

32 mm ________ (.139 - .169)
36 mm ________ (.156 - .190)
53 mm ________ (.228 - .278)
103 mm ________ (.447 - .547)

Maintenance Performed By ________

❑ Battery Replacement
❑ Dry Gas Regulator Replacement
❑ Breath Tube Replacement
❑ Other ________

Temperature Checks Performed By ________

❑ Lab Temp °C ________

External Digital Therm. ID#: ________

❑ 34°C +-.2 Serial #: ________
❑ 34°C +-.2 Serial #: ________
❑ 34°C +-.2 Serial #: ________

Calibration Adjustment Performed By ________

Barometric Pressure Gauge ________ ID # ________

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

❑ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Notes/Suggested Service:

INSTRUMENT WILL NOT POWER ON. SP

Department Inspection Performed By ________

Barometric Pressure ID# ________

Gauge ________ Instrument ________

Mouth Alcohol Solution Lot # ________

Acetone Stock Solution Lot # ________

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

❑ Form 41 ❑ Post-Stability Checks
❑ Stability Checks ❑ Flow Calibration
❑ Calibration Certificate ❑ Form 40
❑ Calibration Adjustment ❑ Other ________

❑ **Instrument Complies with Chapter 11D-8, FAC**
☒ **Instrument Does Not Comply with Chapter 11D-8, FAC**
❑ **Return to/Place into Evidentiary Use**
☒ **Remain Out of Evidentiary Use**
❑ **Conduct an Agency Inspection Before Evidentiary Use**

________ ________
Tech Review / Date Admin Review / Date