



INSTRUMENT PROCESSING SHEET

Agency Sarasota County SOS/N 80-001348

Florida Department of Law Enforcement

Date In 03/01/2019 DI Completion Date 3/14/19 Ship P/U H/D CMI EE

Intake Performed By <u>JA</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>225</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.177</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.527</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE MAR 15 2019 Alcohol Testing Program	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201707D</u> <u>07/25/2019</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>07/25/2019</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3850</u></td> <td><u>201707C</u> <u>07/24/2019</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG831804</u> <u>11-14-20</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201707D</u> <u>07/25/2019</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>07/25/2019</u>	0.200	<u>DR3850</u>	<u>201707C</u> <u>07/24/2019</u>	0.080 DGS	<u>N/A</u>	<u>AG831804</u> <u>11-14-20</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>JA</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.3</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR3850</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1012</u>	<u>201707D</u> <u>07/25/2019</u>																																																											
0.080	<u>DR1279</u>	<u>201707E</u> <u>07/25/2019</u>																																																											
0.200	<u>DR3850</u>	<u>201707C</u> <u>07/24/2019</u>																																																											
0.080 DGS	<u>N/A</u>	<u>AG831804</u> <u>11-14-20</u>																																																											
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1019</u> ID # <u>30793</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G4444</u></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td><u>SD1022</u></td> <td><u>17410</u></td> <td><u>12-6-19</u></td> </tr> <tr> <td>0.100</td> <td><u>SD3964</u></td> <td><u>18200</u></td> <td><u>7-3-20</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>SD1024</u></td> <td><u>18110</u></td> <td><u>4-2-20</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>17817080A2</u></td> <td><u>8-5-19</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201707D</u></td> <td><u>7-25-19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u></td> <td><u>7-25-19</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3850</u></td> <td><u>201707C</u></td> <td><u>7-24-19</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG831804</u></td> <td><u>11-14-20</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G4444</u>	<u>N/A</u>	<u>N/A</u>	0.040	<u>SD1022</u>	<u>17410</u>	<u>12-6-19</u>	0.100	<u>SD3964</u>	<u>18200</u>	<u>7-3-20</u>	0.200	<u>SD1025</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>SD1024</u>	<u>18110</u>	<u>4-2-20</u>	0.080 DGS	<u>N/A</u>	<u>17817080A2</u>	<u>8-5-19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1012</u>	<u>201707D</u>	<u>7-25-19</u>	0.080	<u>DR1279</u>	<u>201707E</u>	<u>7-25-19</u>	0.200	<u>DR3850</u>	<u>201707C</u>	<u>7-24-19</u>	0.080 DGS	<u>N/A</u>	<u>AG831804</u>	<u>11-14-20</u>	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>307SP 28421 SP</u> Gauge <u>1017</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2408</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2882</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1012</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3850</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2408</u>	Interferent	<u>G2882</u>	0.050	<u>SD1012</u>	0.080	<u>DR1279</u>	0.200	<u>DR3850</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>G4444</u>	<u>N/A</u>	<u>N/A</u>																																																										
0.040	<u>SD1022</u>	<u>17410</u>	<u>12-6-19</u>																																																										
0.100	<u>SD3964</u>	<u>18200</u>	<u>7-3-20</u>																																																										
0.200	<u>SD1025</u>	<u>19040</u>	<u>1-29-21</u>																																																										
0.300	<u>SD1024</u>	<u>18110</u>	<u>4-2-20</u>																																																										
0.080 DGS	<u>N/A</u>	<u>17817080A2</u>	<u>8-5-19</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>SD1012</u>	<u>201707D</u>	<u>7-25-19</u>																																																										
0.080	<u>DR1279</u>	<u>201707E</u>	<u>7-25-19</u>																																																										
0.200	<u>DR3850</u>	<u>201707C</u>	<u>7-24-19</u>																																																										
0.080 DGS	<u>N/A</u>	<u>AG831804</u>	<u>11-14-20</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>G2408</u>																																																												
Interferent	<u>G2882</u>																																																												
0.050	<u>SD1012</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>DR3850</u>																																																												
Notes/Suggested Service: <u>instrument needs new keyboard. used spare to conduct inspection. Performed cal adjust to bring values closer to nominal. SP</u> _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use _____ <u>SPM 3/15/19</u> <u>J. Debra</u> <u>3/15/19</u> Tech Review / Date Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 14:24

Date of Inspection: 03/14/2019

Serial Number: 80-001348
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.048	0.078	0.195	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.048	0.080	0.198	0.078
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.079

Standard Deviations	0.0005	0.0006	0.0009	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

03/14/2019
Date

3/15/19
[Signature]

STABILITY CHECKS # 80-001348

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/06/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:13
Control Test	0.049	09:14
Air Blank	0.000	09:14
Control Test	0.049	09:15
Air Blank	0.000	09:16
Control Test	0.050	09:16
Air Blank	0.000	09:17
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

SP

Operator's Signature

3/15/19
 [Signature]

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/06/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:18
Control Test	0.074	09:18
Air Blank	0.000	09:19
Control Test	0.075	09:19
Air Blank	0.000	09:20
Control Test	0.075	09:21
Air Blank	0.000	09:21
Control Test Stats		
Average	0.0747	
Std Dev	0.0006	
Rel Std Dev(%)	0.7732	

SP

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/06/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:22
Control Test	0.198	09:23
Air Blank	0.000	09:23
Control Test	0.198	09:24
Air Blank	0.000	09:25
Control Test	0.198	09:25
Air Blank	0.000	09:26
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/06/2019
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:27
Control Test	0.081	09:28
Air Blank	0.000	09:28
Control Test	0.080	09:28
Air Blank	0.000	09:29
Control Test	0.080	09:29
Air Blank	0.000	09:30
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SP

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001348, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001348</u>	UNCERTAINTY* ±	
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/14/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:24</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

03/14/2019

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Room
3/15/19
[Signature]

SARASOTA COUNTY 50

Intoxilyzer - Alconol Analyzer

Model 8000 SN 80-001348
03/14/2019 09:55:32

Auto Calibration

Max Power Res Value = 43
Auto Range Res Value = 19

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1
Sum Io = 12541, Sum Io = 13432

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 0.1380 (-0.0110)
Sample #2 = 0.1420 (-0.0650)
Sample #3 = 0.1260 (-0.1260)
Sample #4 = 0.1180 (-0.1770)
Aug % Abs = 0.1287 (-0.1227)
STD DEV = 0.0122 (-0.0561)
REL STD DEV = 9.498 (45.713)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 0.1370 (-0.0030)
Sample #2 = 0.1230 (-0.0130)
Sample #3 = 0.1090 (-0.0320)
Sample #4 = 0.1370 (-0.0310)
Aug % Abs = 0.1230 (-0.0253)
STD DEV = 0.0140 (-0.0107)
REL STD DEV = 11.382 (42.208)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1
Sum Io = 12509, Sum Io = 13423

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 0.8550 (-0.0080)
Sample #2 = 0.8850 (-0.0230)
Sample #3 = 0.8650 (-0.0540)
Sample #4 = 0.8720 (-0.0920)
Aug % Abs = 0.8740 (-0.0563)
STD DEV = 0.0101 (-0.0346)
REL STD DEV = 1.161 (61.348)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 0.1370 (-0.0030)
Sample #2 = 0.1230 (-0.0130)
Sample #3 = 0.1090 (-0.0320)
Sample #4 = 0.1370 (-0.0310)
Aug % Abs = 0.1230 (-0.0253)
STD DEV = 0.0140 (-0.0107)
REL STD DEV = 11.382 (42.208)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 7.1170 (-0.0070)
Sample #2 = 7.0990 (-0.0260)
Sample #3 = 7.1390 (-0.0200)
Sample #4 = 7.1370 (-0.0440)
Aug % Abs = 7.1250 (-0.0300)
STD DEV = 0.0225 (-0.0125)
REL STD DEV = 0.316 (41.633)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%

Samples Taken = 4, Discarded = 1
Sum Io = 12478, Sum Io = 13407

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 5.4720 (-0.0230)
Sample #2 = 5.4830 (-0.0050)
Sample #3 = 5.4880 (-0.0220)
Sample #4 = 5.4950 (-0.0150)
Aug % Abs = 5.4887 (-0.0107)
STD DEV = 0.0060 (-0.0140)
REL STD DEV = 0.110 (131.362)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 10.2110 (-0.0130)
Sample #2 = 10.2770 (-0.0020)
Sample #3 = 10.2870 (-0.0060)
Sample #4 = 10.2820 (-0.0110)
Aug % Abs = 10.2820 (-0.0050)
STD DEV = 0.0050 (-0.0065)
REL STD DEV = 0.049 (131.149)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1
Sum Io = 12486, Sum Io = 13417

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 3.8070 (-0.0190)
Sample #2 = 3.7810 (-0.0090)
Sample #3 = 3.7730 (-0.0390)
Sample #4 = 3.7930 (-0.0430)
Aug % Abs = 3.7823 (-0.0303)
STD DEV = 0.0101 (-0.0186)
REL STD DEV = 0.266 (61.263)

***** AUTO CAL DATA *****

<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.129
Std Dev = 0.01 Rel Std Dev = 9.50
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.874
Std Dev = 0.01 Rel Std Dev = 1.16
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.984
Std Dev = 0.03 Rel Std Dev = 1.62

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.123
Std Dev = 0.01 Rel Std Dev = 11.38
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.586
Std Dev = 0.01 Rel Std Dev = 0.64
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.747
Std Dev = 0.01 Rel Std Dev = 0.36
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.125
Std Dev = 0.02 Rel Std Dev = 0.32
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.282
Std Dev = 0.01 Rel Std Dev = 0.05
Zero Order Coef = -142.38
First Order Coef = 1256.03
Second Order Coef = 14.28
Standard Deviation = 12.485214

<<<<< CHANNEL 2 >>>>>

Sol Val = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3147.00
Sample #2 = 3106.00
Sample #3 = 3075.00
Sample #4 = 3098.00
Average Result = 3093.0000
STD DEV = 16.0935
REL STD DEV = 0.520
***** CHANNEL 2 *****
Sample #1 = 3386.00
Sample #2 = 3416.00
Sample #3 = 3422.00
Sample #4 = 3427.00
Average Result = 3421.6667
STD DEV = 5.5076
REL STD DEV = 0.161

***** CHANNEL 2 *****

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
3 um H2O Adjust (mg/l*10,000) = 716
9 um H2O Adjust (mg/l*10,000) = 388
***** AUTO CAL PASS *****

CAL ADJUSTMENT
80-001348
SR

3/15/19
SR

POST CAL ADJUSTMENT
80-001348
STABILITY CHECKS

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001348
03/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:56
Control Test	0.048	10:56
Air Blank	0.000	10:57
Control Test	0.049	10:58
Air Blank	0.000	10:58
Control Test	0.049	10:59
Air Blank	0.000	10:59
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

SP

Operator's Signature

3/15/19
0/2

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001348
03/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:51
Control Test	0.079	10:51
Air Blank	0.000	10:52
Control Test	0.079	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001348
03/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.197	10:44
Air Blank	0.000	10:44
Control Test	0.198	10:45
Air Blank	0.000	10:46
Control Test	0.198	10:46
Air Blank	0.000	10:47
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

SP

Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001348
03/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:01
Control Test	0.079	11:02
Air Blank	0.000	11:02
Control Test	0.079	11:02
Air Blank	0.000	11:03
Control Test	0.079	11:03
Air Blank	0.000	11:04
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

SP

Operator's Signature