FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Orange County S/N 80-001418

Date In 11/18/2019 DI Completion Date 12/03/2019 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By DP

- ☑ Annual
- ☐ Registration
- ☑ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

DEC 09 2019

Alcohol Testing Program

Quality Checks Performed By MK

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 215
- ☑ Flow Verification (L/s)

Flow Column # ATP 102

32 mm 0.164 (.139 - .169)
36 mm 0.175 (.156 - .190)
53 mm 0.246 (.228 - .278)
103 mm 0.516 (.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 30793

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5088	201905A
		05/14/2021
0.080	MP5089	201905B
		05/14/2021
0.200	MP5090	201904D
		04/30/2021
0.080 DGS	N/A	AG916501
		06/14/2021

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By MK

- ☑ Lab Temp °C MK ~~23.83~~ 21.83
- External Digital Therm. ID#: 300505
- ☑ 34°C +-.2 Serial #: MP5088
- ☑ 34°C +-.2 Serial #: MP5089
- ☑ 34°C +-.2 Serial #: MP5090

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By MK

Barometric Pressure ID# 26932
Gauge 1016 Instrument 1015
Mouth Alcohol Solution Lot # 2019-B
Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	MP5086
Interferent	MP5087
0.050	MP5088
0.080	MP5089
0.200	MP5090

Attachments

- ☑ Form 41
- ☑ Stability Checks
- ☑ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: _____

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

PJM 12/5/19 Brett Kirkland 12/6/19

Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 12:19
Date of Inspection: 12/03/2019
Serial Number: 80-001418
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.050	0.080	0.202	0.080
0.000	0.049	0.081	0.202	0.079
0.000	0.050	0.080	0.203	0.079
0.000	0.051	0.080	0.203	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.081	0.203	0.079
0.000	0.051	0.080	0.203	0.079
0.000	0.051	0.081	0.203	0.080
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.203	0.079

Standard Deviations	0.0006	0.0005	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

PGM
BK
12/6/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael D Haughey

MICHAEL D HAUGHEY
Signature and Printed Name

12/03/2019
Date

FDLE/ATP Form 41 –Revised August 2005

80-001418 Stability Check 12/03/2019 MK

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
12/03/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.050	10:06
Air Blank	0.000	10:06
Control Test	0.049	10:07
Air Blank	0.000	10:07
Control Test	0.049	10:08
Air Blank	0.000	10:09
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

MK
Operator's Signature

BBM
RSK
12/6/19

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
12/03/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.080	10:11
Air Blank	0.000	10:12
Control Test	0.079	10:12
Air Blank	0.000	10:13
Control Test	0.079	10:13
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

MK
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
12/03/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.202	10:19
Air Blank	0.000	10:20
Control Test	0.199	10:20
Air Blank	0.000	10:21
Control Test	0.200	10:22
Air Blank	0.000	10:22
Control Test Stats		
Average	0.2003	
Std Dev	0.0015	
Rel Std Dev(%)	0.7625	

MK
Operator's Signature

DES

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
12/03/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.080	10:24
Air Blank	0.000	10:25
Control Test	0.080	10:25
Air Blank	0.000	10:26
Control Test	0.080	10:26
Air Blank	0.000	10:27
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MK
Operator's Signature

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001418 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001418
Owning Agency:	ORANGE COUNTY S.O.
Calibration Date:	12/03/2019
Calibration Time:	12:19

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/03/2019
Date

MICHAEL D HAUGHEY,
Department Inspector

BK 12/6/19

Return Material Authorization

Ship to: ☑ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Kelly Melville on 10/7/19

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001418

Bill To Address:	Ship to Address:
Orange County SO	FDLE ATP Tallahassee

Reason for Return:

Dry gas regulator not working

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Kelly Melville

Phone #: 407-836-0800 Email: Kelly.Melville@ocfl.net

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Orange County S/N 80-001418

Date In 9/19/2019 DI Completion Date 10/3/19 ☐Ship ☐P/U ☐H/D ☒CMI ☐EE

Intake Performed By DP

- ☑ Annual
- ☐ Registration
- ☑ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

Quality Checks Performed By SP

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 214
- ☐ Flow Verification (L/s)

Flow Column # _____

- 32 mm _____ (.139 - .169)
- 36 mm _____ (.156 - .190)
- 53 mm _____ (.228 - .278)
- 103 mm _____ (.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 26932

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201905A 5-14-21
0.080	DR1279	201905B 5-14-21
0.200	SD1013	201904D 4-30-21
0.080 DGS	N/A	AG916501 6-14-21

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm

- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

- 32 mm _____ (.139 - .169)
- 36 mm _____ (.156 - .190)
- 53 mm _____ (.228 - .278)
- 103 mm _____ (.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By _____

- ☐ Lab Temp °C _____

External Digital Therm. ID#: _____

- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By _____

Barometric Pressure ID# _____

Gauge _____ Instrument _____

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

- ☐ Form 41
- ☐ Stability Checks
- ☐ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: DGS NOT WORKING. SP

- ☐ **Instrument Complies with Chapter 11D-8, FAC**
- ☒ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☐ **Return to/Place into Evidentiary Use**
- ☒ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

Tech Review / Date _____ Admin Review / Date _____

Stability Checks #80-001418

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
09/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:13
Control Test	0.048	10:14
Air Blank	0.000	10:14
Control Test	0.048	10:15
Air Blank	0.000	10:16
Control Test	0.048	10:16
Air Blank	0.000	10:17
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
09/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.078	10:03
Air Blank	0.000	10:03
Control Test	0.079	10:04
Air Blank	0.000	10:04
Control Test	0.078	10:05
Air Blank	0.000	10:06
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

SP
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
09/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.200	09:58
Air Blank	0.000	09:59
Control Test	0.199	09:59
Air Blank	0.000	10:00
Control Test	0.198	10:01
Air Blank	0.000	10:01
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

SP
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
09/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.000	10:19
Air Blank	0.000	10:19
Control Test	0.000	10:19
Air Blank	0.000	10:20
Control Test	0.000	10:20
Air Blank	0.000	10:21
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP
Operator's Signature

Return Material Authorization

Ship to: ☑ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Kelly Melville on 7/2/19

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001418

Bill To Address:	Ship to Address:
Orange County SO	FDLE-ATP Tallahassee

Reason for Return:

Instrument not reading dry gas standard.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Kelly Melville

Phone #: 407-836-0804 Email: Kelly.Melville@ocfl.net

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us

INSTRUMENT PROCESSING SHEET

FDLE Florida Department of Law Enforcement

Agency Orange County S/N 80-001418

Date In 6/24/2019 DI Completion Date _____ ☐Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By DP

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

Quality Checks Performed By SP

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 214
- ☒ Flow Verification (L/s)

Flow Column # ATP103

32 mm	.144	(.139 - .169)
36 mm	.164	(.156 - .190)
53 mm	.226	(.228 - .278)
103 mm	.496	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 26932

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201707D 7-25-19
0.080	DR1279	201707E 7-25-19
0.200	SD1013	201707C 7-24-19
0.080 DGS	N/A	AG916501 6-14-21

Flow Calibration Performed By SP

Flow Column # ATP102

- ☒ 5L/min – 17mm
- ☒ 15L/min – 53mm
- ☒ 30L/min – 103mm
- ☒ R-Value 217
- ☒ Post Calibration Verification (L/s)

Flow Column # ATP103

32 mm	.148	(.139 - .169)
36 mm	.167	(.156 - .190)
53 mm	.234	(.228 - .278)
103 mm	.496	(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By _____

- ☐ Lab Temp °C _____

External Digital Therm. ID#: _____

- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By _____

Barometric Pressure ID# _____

Gauge _____ Instrument _____

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

- ☐ Form 41
- ☐ Stability Checks
- ☐ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: _____

- ☐ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☐ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

Tech Review / Date _____ Admin Review / Date _____

stability checks #80-001418

RANGE COUNTY S.O.
ntoxilyzer - Alcohol Analyzer
odel 8000 SN 80-001418
5/28/2019
oftware: 8100.27

est	g/210L	Time
ir Blank	0.000	10:16
ontrol Test	0.048	10:17
ir Blank	0.000	10:17
ontrol Test	0.049	10:18
ir Blank	0.000	10:18
ontrol Test	0.049	10:19
ir Blank	0.000	10:20

ontrol Test Stats

Average	0.0487
Std Dev	0.0006
Rel Std Dev(%)	1.1863

SP
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
06/28/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.079	10:12
Air Blank	0.000	10:13
Control Test	0.079	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:15
Air Blank	0.000	10:15

Control Test Stats

Average	0.0793
Std Dev	0.0006
Rel Std Dev(%)	0.7277

SP
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
06/28/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:06
Control Test	0.201	10:07
Air Blank	0.000	10:08
Control Test	0.199	10:08
Air Blank	0.000	10:09
Control Test	0.198	10:10
Air Blank	0.000	10:10

Control Test Stats

Average	0.1993
Std Dev	0.0015
Rel Std Dev(%)	0.7663

SP
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
06/28/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.000	10:23
Air Blank	0.000	10:23
Control Test	0.000	10:23
Air Blank	0.000	10:24
Control Test	0.000	10:24
Air Blank	0.000	10:25

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DGS

SP
Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
06/28/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:57
Control Test	0.000	10:58
Air Blank	0.000	10:58
Control Test	0.000	10:58
Air Blank	0.000	10:59
Control Test	0.080	10:59
Air Blank	0.000	11:00

Control Test Stats

Average	0.0267
Std Dev	0.0462
Rel Std Dev(%)	173.2051

DGS

SP
Operator's Signature

FLOW CAL ADJUSTMENT

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001418
06/28/2019
Software: 8100.27

Flow Rate Calibration********
1: Rate (Liters/min) = 5
SQRT(Diff)) = 6.926
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.355
3: Rate (Liters/min) = 30
SQRT(Diff)) = 21.094
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 675
Rounded Intercept = -601025
Correlation = 0.99519

SP