



INSTRUMENT PROCESSING SHEET

Agency Charlotte County SO

S/N 80-001739

Florida Department of Law Enforcement

Date In 02/04/2019 DI Completion Date 2/7/19

Ship  P/U  H/D  CMI  EE

**Intake** Performed By JD

Annual  
 Registration  
 Return from CMI / EE

Visual Inspection:  
 Case  Handle  
 Keyboard  Dry Gas Shelf  
 Feet  Breath Tube  
 Ports  Screws Tight

Other Equipment/ Accessories:  
 Power cord  Printer Cable  
 Static Bag  12V DC Cable

Notes: \_\_\_\_\_

**Quality Checks** Performed By SP

Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 192  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32 mm .152 (.139 - .169)  
 36 mm .171 (.156 - .190)  
 53 mm .253 (.228 - .278)  
 103 mm .527 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	<u>SD1012</u>	201707D 07/25/2019
0.080	<u>DR1279</u>	201707E 07/25/2019
0.200	<u>DR3850</u>	201707C 07/24/2019
0.080 DGS	N/A	<u>AG805701</u> <u>2-26-20</u>

**Flow Calibration** Performed By \_\_\_\_\_

Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)

Flow Column # \_\_\_\_\_  
 32 mm \_\_\_\_\_ (.139 - .169)  
 36 mm \_\_\_\_\_ (.156 - .190)  
 53 mm \_\_\_\_\_ (.228 - .278)  
 103 mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By SP

Lab Temp °C 23.2  
 External Digital Therm. ID#: 300502  
 34°C +/-2 Serial #: SD1012  
 34°C +/-2 Serial #: DR1279  
 34°C +/-2 Serial #: DR3850

**Final Release Date**  
**FDLE**  
**FEB 07 2019**  
**Alcohol Testing Program**

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By SP

Barometric Pressure ID# 28421  
 Gauge 1018 Instrument 1017  
 Mouth Alcohol Solution Lot # 2018-B  
 Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	<u>G2408</u>
Interferent	<u>G2882</u>
0.050	<u>SD1012</u>
0.080	<u>DR1279</u>
0.200	<u>DR3850</u>

**Attachments**

Form 41  Post-Stability Checks  
 Stability Checks  Flow Calibration  
 Calibration Certificate  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_

Notes/Suggested Service: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Room 2/7/19 J. Zah 2/7/19  
 Tech Review / Date Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO  
Time of Inspection: 10:48

Date of Inspection: 02/07/2019

Serial Number: 80-001739  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.078	0.193	0.080
0.000	0.047	0.079	0.197	0.079
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.080	0.196	0.080
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.079	0.196	0.080

Standard Deviations	0.0009	0.0007	0.0013	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

*gpm*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT  
Signature and Printed Name

02/07/2019  
Date

*2/7/19  
jo*

# STABILITY CHECKS # 80-001739

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001739  
02/07/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.079	08:52
Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:54
Control Test Stats	0.0797	08:54
Average	0.0006	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP  
Operator's Signature

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001739  
02/07/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:00
Control Test	0.196	09:01
Air Blank	0.000	09:01
Control Test	0.197	09:02
Air Blank	0.000	09:02
Control Test	0.197	09:03
Air Blank	0.000	09:04
Control Test Stats	0.1967	
Average	0.0006	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

SP  
Operator's Signature

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001739  
02/07/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:55
Control Test	0.080	08:56
Air Blank	0.000	08:56
Control Test	0.080	08:57
Air Blank	0.000	08:57
Control Test	0.079	08:58
Air Blank	0.000	08:59
Control Test Stats	0.0797	
Average	0.0006	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP  
Operator's Signature

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001739  
02/07/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:12
Control Test	0.048	09:13
Air Blank	0.000	09:13
Control Test	0.048	09:14
Air Blank	0.000	09:15
Control Test	0.049	09:15
Air Blank	0.000	09:16
Control Test Stats	0.0483	
Average	0.0006	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

SP  
Operator's Signature

2/7/19  
2/7/19  
DGM



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001739, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001739</u>	UNCERTAINTY* ±	
Owning Agency:	<u>CHARLOTTE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/07/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:48</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/07/2019

Date

*Shayla Platt*

SHAYLA D PLATT,  
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*Handwritten initials/signature*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO  
Time of Inspection: 08:47

Date of Inspection: 02/07/2019

Serial Number: 80-001739  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASSED AI TO OPERATE INSTRUMENT

*gsm*

*N/A COMPLIANCE  
NOT DETERMINED*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT  
Signature and Printed Name

02/07/2019  
Date

*2/7/19  
JP*