INSTRUMENT PROCESSING SHEET

FDLE
Florida Department of Law Enforcement

Agency FFWCC S/N 80-003195

Date In 03/18/2019 DI Completion Date 3/27/19 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By JR

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☑ Power cord ☐ Printer Cable
- ☐ Static Bag ☑ 12V DC Cable

Notes: 1.) Pelican Case
2.) Peg that holds keyboard is missing on left side of the instrument.

Final Release Date

FDLE

MAR 27 2019

Alcohol Testing Program

Quality Checks Performed By BK

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 111
- ☒ Flow Verification (L/s)

Flow Column # ATP102

32 mm	.167	(.139 - .169)
36 mm	.179	(.156 - .190)
53 mm	.257	(.228 - .278)
103 mm	.1539	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 30793

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1021	201707D
		07/25/2019
0.080	DR1275	201707E
		07/25/2019
0.200	SD1011	201707C
		07/24/2019
0.080 DGS	N/A	AG805701
		2/26/20

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By BK

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☒ Other Keyboard peg left side

Temperature Checks Performed By BK

- ☒ Lab Temp °C 20.7/22.7

External Digital Therm. ID#: 300503/300502

- ☒ 34°C +-.2 Serial #: SD1021/SD1021
- ☒ 34°C +-.2 Serial #: DR1275/DR1275
- ☒ 34°C +-.2 Serial #: SD1011/SD1011

Calibration Adjustment Performed By BK

Barometric Pressure Gauge 1021 ID # 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	DR2035	N/A	N/A
0.040	SD1022	17410	12/6/19
0.100	SD3964	18200	7/3/20
0.200	SD1025	19040	1/29/21
0.300	SD1024	18110	4/2/20
0.080 DGS	N/A	1781708OA2	8/5/19

☒ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	SD1021	201707D	7/25/19
0.080	DR1275	201707E	7/25/19
0.200	SD1011	201707C	7/24/19
0.080 DGS	N/A	AG805701	2/26/20

Department Inspection Performed By BK

Barometric Pressure ID# 28421/30793

Gauge 1022/1022 Instrument 1021/1022

Mouth Alcohol Solution Lot # 2018-B/2018-B

Acetone Stock Solution Lot # 2018-A/2018-A

Simulator	Serial Number
0.000	G11621/G11621
Interferent	DR3855/DR3855
0.050	SD1021/SD1021
0.080	DR1275/DR1275
0.200	SD1011/SD1011

Attachments

- ☒ Form 41 X 2
- ☒ Stability Checks
- ☒ Calibration Certificate
- ☒ Calibration Adjustment
- ☒ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☒ Other Extra DGS stabilities

Notes/Suggested Service:
Cal adjust to bring values closer to nominal. BK

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

SP 3/27/19 [signature] 3/27/19

Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWC
Serial Number: 80-003195
Time of Inspection: 10:58
Date of Inspection: 03/27/2019
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.080	0.199	0.079
0.000	0.048	0.081	0.200	0.080
0.000	0.048	0.080	0.199	0.079
0.000	0.048	0.081	0.199	0.079
0.000	0.049	0.081	0.200	0.080
0.000	0.048	0.080	0.200	0.080
0.000	0.048	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.081	0.200	0.079

Standard Deviations	0.0005	0.0005	0.0005	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland
BRETT H KIRKLAND
Signature and Printed Name

03/27/2019
Date

3/27/19
SP

FDLE/ATP Form 41 –Revised August 2005

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWC
Serial Number: 80-003195
Time of Inspection: 12:29
Date of Inspection: 03/20/2019
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.079	0.195	0.000 / 0.084
0.000	0.050	0.080	0.197	0.000 / 0.084
0.000	0.050	0.080	0.197	0.000 / 0.084
0.000	0.049	0.081	0.197	0.087 / 0.084
0.000	0.050	0.080	0.197	0.086 / 0.084
0.000	0.049	0.080	0.198	0.086 / 0.084
0.000	0.049	0.080	0.198	0.086 / 0.084
0.000	0.050	0.080	0.198	0.085 / 0.084
0.000	0.049	0.080	0.197	0.085 / 0.084
0.000	0.049	0.080	0.198	0.084 / 0.083

Standard Deviations	0.0005	0.0004	0.0009	0.0413 / 0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:
08: Control Outside FORGOT TO PLUG IN HOSE

SP

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland BRETT H KIRKLAND
Signature and Printed Name

03/20/2019
Date

3/27/19
SD

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:58
Control Test	0.049	08:59
Air Blank	0.000	08:59
Control Test	0.049	09:00
Air Blank	0.000	09:01
Control Test	0.050	09:01
Air Blank	0.000	09:02
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

BK
Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.080	09:09
Air Blank	0.000	09:09
Control Test	0.079	09:10
Air Blank	0.000	09:11
Control Test	0.080	09:11
Air Blank	0.000	09:12
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

BK
Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:13
Control Test	0.196	09:14
Air Blank	0.000	09:14
Control Test	0.197	09:15
Air Blank	0.000	09:16
Control Test	0.197	09:16
Air Blank	0.000	09:17
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

BK
Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.080	08:51
Air Blank	0.000	08:51
Control Test	0.080	08:52
Air Blank	0.000	08:52
Control Test	0.080	08:53
Air Blank	0.000	08:53
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

BK
Operator's Signature

SP

80-003195
stability Check
BK
3/27/19

3/27/19
JP

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003195 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

			UNCERTAINTY* ±
Serial Number:	80-003195		
Owning Agency:	FWC	0.050 g/ 210 L	0.004
Calibration Date:	03/27/2019	0.080 g/ 210 L	0.004
Calibration Time:	10:58	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/27/2019
Date

Brett Kirkland
BRETT H KIRKLAND,
Department Inspector

SP

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

3/27/19
SP

```
FWC
Intoxilyzer - Alcohol Analyzer
Model 8000                SN 80-003195
03/27/2019                  07:51:52

Auto Calibration
Max Power Res Value = 27
Auto Range Res Value = 15

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12720, 9um Io = 12771
     <<<<< CHANNEL 1 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  0.0820      (-0.0120)
Sample #2 =  0.0810      (0.0050)
Sample #3 =  0.0330      (0.0370)
Sample #4 =  0.0950      (0.0270)
Avg % Abs = 0.0697  (0.0230)
STD DEV =  0.0325  (0.0164)
REL STD DEV = 46.675  (71.177)
-------------------

     <<<<< CHANNEL 2 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  0.1410      (-0.0080)
Sample #2 =  0.1180      (-0.0150)
Sample #3 =  0.1260      (-0.0230)
Sample #4 =  0.1320      (-0.0070)
Avg % Abs = 0.1253  (-0.0150)
STD DEV =  0.0070  (0.0080)
REL STD DEV = 5.604  (53.333)
-------------------

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12712, 9um Io = 12771
     <<<<< CHANNEL 1 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  0.8260      (-0.0140)
Sample #2 =  0.7880      (0.0220)
Sample #3 =  0.8170      (0.0240)
Sample #4 =  0.7870      (0.0380)
Avg % Abs = 0.7973  (0.0280)
STD DEV =  0.0170  (0.0087)
REL STD DEV = 2.137  (31.135)
-------------------

     <<<<< CHANNEL 2 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  1.5780      (-0.0100)
Sample #2 =  1.5560      (0.0020)
Sample #3 =  1.5900      (0.0070)
Sample #4 =  1.5690      (0.0060)
Avg % Abs = 1.5717  (0.0050)
STD DEV =  0.0172  (0.0026)
REL STD DEV = 1.092  (52.915)
-------------------

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12706, 9um Io = 12767
     <<<<< CHANNEL 1 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  1.9190      (-0.0230)
Sample #2 =  1.9140      (-0.0020)
Sample #3 =  1.8550      (0.0330)
Sample #4 =  1.8520      (0.0360)
Avg % Abs = 1.8737  (0.0223)
STD DEV =  0.0350  (0.0211)
REL STD DEV = 1.866  (94.597)
-------------------

     <<<<< CHANNEL 2 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  3.6360      (-0.0260)
Sample #2 =  3.6600      (-0.0240)
Sample #3 =  3.6420      (0.0000)
Sample #4 =  3.6320      (0.0100)
Avg % Abs = 3.6447  (-0.0047)
STD DEV =  0.0142  (0.0175)
REL STD DEV = 0.389  (374.438)
-------------------

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12699, 9um Io = 12763
     <<<<< CHANNEL 1 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  3.6080      (-0.0020)
Sample #2 =  3.6210      (0.0090)
Sample #3 =  3.6800      (-0.0170)
Sample #4 =  3.6210      (0.0240)
Avg % Abs = 3.6407  (0.0053)
STD DEV =  0.0341  (0.0207)
REL STD DEV = 0.936  (388.959)
-------------------

     <<<<< CHANNEL 2 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  6.8940      (0.0020)
Sample #2 =  6.9150      (0.0050)
Sample #3 =  6.9640      (-0.0130)
Sample #4 =  6.9740      (-0.0040)
Avg % Abs = 6.9510  (-0.0040)
STD DEV =  0.0316  (0.0090)
REL STD DEV = 0.454  (225.000)
-------------------

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12695, 9um Io = 12762
     <<<<< CHANNEL 1 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  5.2630      (0.0000)
Sample #2 =  5.3270      (-0.0100)
Sample #3 =  5.3000      (0.0260)
Sample #4 =  5.2800      (0.0290)
Avg % Abs = 5.3023  (0.0150)
STD DEV =  0.0236  (0.0217)
REL STD DEV = 0.445  (144.684)
-------------------

     <<<<< CHANNEL 2 >>>>>
 Sample        % Abs      (% Abs Ref)
Sample #1 =  9.9830      (0.0070)
Sample #2 = 10.0680      (-0.0070)
Sample #3 = 10.0320      (0.0100)
Sample #4 = 10.0350      (0.0020)
Avg % Abs = 10.0450  (0.0017)
STD DEV =  0.0200  (0.0085)
REL STD DEV = 0.199  (510.294)
-------------------

     ***** AUTO CAL DATA *****
     <<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs =    0.070
 Std Dev =   0.03 Rel Std Dev =  46.67
Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs =    0.797
 Std Dev =   0.02 Rel Std Dev =   2.14
Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs =    1.874
 Std Dev =   0.03 Rel Std Dev =   1.87
Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs =    3.641
 Std Dev =   0.03 Rel Std Dev =   0.94
Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs =    5.302
 Std Dev =   0.02 Rel Std Dev =   0.44
Zero Order Coef = -162.32
First Order Coef = 2554.31
Second Order Coef =  31.72
Standard Deviation =  24.343534
-------------------
     <<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs =    0.125
 Std Dev =   0.01 Rel Std Dev =   5.60
Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs =    1.572
 Std Dev =   0.02 Rel Std Dev =   1.09
Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs =    3.645
 Std Dev =   0.01 Rel Std Dev =   0.39
Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs =    6.951
 Std Dev =   0.03 Rel Std Dev =   0.45
Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs =   10.045
 Std Dev =   0.02 Rel Std Dev =   0.20
Zero Order Coef = -163.25
First Order Coef = 1297.32
Second Order Coef =  14.02
Standard Deviation =  7.342533
-------------------

-----------------------------------
| Solution Stats Quadratic Fit Chan 1 |
|  Act      Fit      Residual       |
| g/210L   g/210L     g/210L        |
| 0.000    0.000     -0.0003        |
| 0.040    0.040      0.0002        |
| 0.100    0.099      0.0006        |
| 0.200    0.201     -0.0007        |
| 0.300    0.300      0.0003        |
-----------------------------------

-----------------------------------
| Solution Stats Quadratic Fit Chan 2 |
|  Act      Fit      Residual       |
| g/210L   g/210L     g/210L        |
| 0.000   -0.000      0.0000        |
| 0.040    0.040     -0.0001        |
| 0.100    0.100      0.0002        |
| 0.200    0.200     -0.0002        |
| 0.300    0.300      0.0001        |
-----------------------------------

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l  %%%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 =  3366.00
Sample #2 =  3210.00
Sample #3 =  3369.00
Sample #4 =  3288.00
Average Result = 3289.0000
STD DEV = 79.5047
REL STD DEV = 2.417
**********
***** CHANNEL 2
Sample #1 =  3356.00
Sample #2 =  3281.00
Sample #3 =  3336.00
Sample #4 =  3303.00
Average Result = 3306.6667
STD DEV = 27.6827
REL STD DEV = 0.837
**********
Dry Gas H20 Adjust Results **********
  Barometric Pressure = 1021
  3 um H20 Adjust (mg/l*10,000) = 520
  9 um H20 Adjust (mg/l*10,000) = 503
**** AUTO CAL PASS
```

3/27/19
SP

80-003195
Cal Adjust
BK
3/27/19

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:47
Control Test	0.048	08:48
Air Blank	0.000	08:48
Control Test	0.049	08:49
Air Blank	0.000	08:49
Control Test	0.049	08:50
Air Blank	0.000	08:51
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

BK
Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.080	08:53
Air Blank	0.000	08:53
Control Test	0.080	08:54
Air Blank	0.000	08:55
Control Test	0.081	08:55
Air Blank	0.000	08:56
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

BK
Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.198	08:58
Air Blank	0.000	08:58
Control Test	0.199	08:59
Air Blank	0.000	08:59
Control Test	0.199	09:00
Air Blank	0.000	09:00
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

BK
Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/27/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:43
Control Test	0.079	08:43
Air Blank	0.000	08:43
Control Test	0.079	08:44
Air Blank	0.000	08:44
Control Test	0.079	08:45
Air Blank	0.000	08:45
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

~~DGS~~ BK
Operator's Signature

CR

3/27/19
JD

80-003195
Post-stabilities
BK
3/27/19

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:03
Control Test	0.083	13:04
Air Blank	0.000	13:04
Control Test	0.083	13:04
Air Blank	0.000	13:05
Control Test	0.083	13:05
Air Blank	0.000	13:06
Control Test	0.083	13:06
Air Blank	0.000	13:07
Control Test	0.083	13:07
Air Blank	0.000	13:07
Control Test	0.083	13:08
Air Blank	0.000	13:08
Control Test	0.082	13:08
Air Blank	0.000	13:09
Control Test	0.082	13:09
Air Blank	0.000	13:10
Control Test	0.082	13:10
Air Blank	0.000	13:11
Control Test	0.081	13:11
Air Blank	0.000	13:11
Control Test	0.081	13:12
Air Blank	0.000	13:12
Control Test	0.082	13:13
Air Blank	0.000	13:13
Control Test	0.081	13:13
Air Blank	0.000	13:14
Control Test	0.081	13:14
Air Blank	0.000	13:15
Control Test	0.080	13:15
Air Blank	0.000	13:15
Control Test	0.081	13:16
Air Blank	0.000	13:16
Control Test	0.081	13:17
Air Blank	0.000	13:17
Control Test	0.080	13:17
Air Blank	0.000	13:18
Control Test	0.081	13:18
Air Blank	0.000	13:19
Control Test	0.080	13:19
Air Blank	0.000	13:19

Control Test Stats

Average	0.0816
Std Dev	0.0011
Rel Std Dev(%)	1.3343

DGS

BK
Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003195
03/20/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:01
Control Test	0.084	13:01
Air Blank	0.000	13:02
Control Test	0.083	13:02
Air Blank	0.000	13:02

Control Test Stats

Average	0.0835
Std Dev	0.0007
Rel Std Dev(%)	0.8468

DGS

BK
Operator's Signature

SP

80-003195
Additional DGS
stabilities
BK
3/27/19
3/27/19
SP