FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Miami Beach Police Department S/N 80-003224

Date In 03/07/2019 DI Completion Date 03/07/2019 ☐Ship ☐P/U ☑H/D ☐CMI ☐EE

Intake Performed By DERR

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☐ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

MAR 15 2019

Alcohol Testing Program

Quality Checks Performed By DERR

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 223
- ☑ Flow Verification (L/s)

Flow Column # ATP 101

32 mm	.152	(.139 - .169)
36 mm	.171	(.156 - .190)
53 mm	.238	(.228 - .278)
103 mm	.519	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28199

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD3967	201707D
		07/25/2019
0.080	SD3968	201707E
		07/25/2019
0.200	SD3969	201707C
		07/24/2019
0.080 DGS	N/A	AG805702
		02/26/2020

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By DERR

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☑ Other Forms load/ Changed pass

Temperature Checks Performed By DERR

- ☑ Lab Temp °C 22.24

External Digital Therm. ID#: 300504

- ☑ 34°C +-.2 Serial #: SD3967
- ☑ 34°C +-.2 Serial #: SD3968
- ☑ 34°C +-.2 Serial #: SD3969

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ *Post Calibration Adjustment Stability Checks*

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DERR

Barometric Pressure ID# 68639

Gauge 1025 Instrument 1023

Mouth Alcohol Solution Lot # 2017-B

Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	SD3963
Interferent	SD3965
0.050	SD3967
0.080	SD3968
0.200	SD3969

Attachments

- ☑ Form 41
- ☑ Stability Checks
- ☑ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: E-mailed ☑ APPROVED

Approved date 3/7/2019

DERR Digitally signed by DERR Date: 2020.05.20 13:10:29 -04'00'

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

DERR 3/15/19 Tech Review / Date

[signature] 3/15/19 Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI BEACH PD
Time of Inspection: 12:46 Date of Inspection: 03/07/2019
Serial Number: 80-003224
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.081

Standard Deviations	0.0000	0.0003	0.0005	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

PJM

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAVID E REYES-RIVERA
Signature and Printed Name

03/07/2019
Date

3/15/19
DR

FDLE/ATP Form 41 –Revised August 2005

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003224	Miami Beach Police Department	03/07/2019	

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.09C	SN: SD3968 Temp: 34.06C	SN: SD3969 Temp:34.11C	Lot AG805702
0.047 to 0.053 ☑	0.077 to 0.083 ☑	0.194 to 0.206 ☑	0.077 to 0.083 ☑

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.050	10:26
Air Blank	0.000	10:26
Control Test	0.049	10:27
Air Blank	0.000	10:28
Control Test	0.050	10:29
Air Blank	0.000	10:29

Control Test Stats

Average	0.0497
Std Dev	0.0006
Rel Std Dev(%)	1.1625

DCLR
Operator's Signature

3/15/19 DP

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.080	10:31
Air Blank	0.000	10:31
Control Test	0.081	10:32
Air Blank	0.000	10:33
Control Test	0.081	10:33
Air Blank	0.000	10:34

Control Test Stats

Average	0.0807
Std Dev	0.0006
Rel Std Dev(%)	0.7157

DCLR
Operator's Signature

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.199	10:35
Air Blank	0.000	10:36
Control Test	0.199	10:37
Air Blank	0.000	10:37
Control Test	0.199	10:38
Air Blank	0.000	10:39

Control Test Stats

Average	0.1990
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DCLR
Operator's Signature

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:39
Control Test	0.079	10:40
Air Blank	0.000	10:40
Control Test	0.079	10:41
Air Blank	0.000	10:41
Control Test	0.079	10:41
Air Blank	0.000	10:42

Control Test Stats

Average	0.0790
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DCLR
Operator's Signature

Room

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003224 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-003224
Owning Agency:	MIAMI BEACH PD
Calibration Date:	03/07/2019
Calibration Time:	12:46

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/07/2019
Date

DAVID E REYES-RIVERA,
Department Inspector

3/5/19

Form

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Miami Beach Police Department S/N 80-003224

Date In 03/07/2019 DI Completion Date 03/07/2019 ☐Ship ☐P/U ☑H/D ☐CMI ☐EE

Intake Performed By DCLL

☑ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☐ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date
FDLE
MAR 15 2019
Alcohol Testing Program

Quality Checks Performed By DCLL

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 223
☑ Flow Verification (L/s)
Flow Column # ATP 101
32 mm .152 (.139 - .169)
36 mm .171 (.156 - .190)
53 mm .238 (.228 - .278)
103 mm .519 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 28199
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD3967	201707D 07/25/2019
0.080	SD3968	201707E 07/25/2019
0.200	SD3969	201707C 07/24/2019
0.080 DGS	N/A	AG805702 02/26/2020

Flow Calibration Performed By _____

Flow Column # _____
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value _____
☐ Post Calibration Verification (L/s)
Flow Column # _____
32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance Performed By DCLL

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☑ Other Forms load/ Changed pass

Temperature Checks Performed By DCLL

☑ Lab Temp °C 22.24
External Digital Therm. ID#: 300504
☑ 34°C +-.2 Serial #: SD3967
☑ 34°C +-.2 Serial #: SD3968
☑ 34°C +-.2 Serial #: SD3969

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DCLL

Barometric Pressure ID# 68639
Gauge 1025 Instrument 1023
Mouth Alcohol Solution Lot # 2017-B
Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	SD3963
Interferent	SD3965
0.050	SD3967
0.080	SD3968
0.200	SD3969

Attachments

☑ Form 41	☐ Post-Stability Checks
☑ Stability Checks	☐ Flow Calibration
☑ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☐ Other _____

Notes/Suggested Service: E-mailed ✓ APPROVED

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☑ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
☑ **Conduct an Agency Inspection Before Evidentiary Use**

DCLL 3/15/19 Tech Review / Date
[signature] 3/15/19 Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI BEACH PD
Time of Inspection: 12:46 Date of Inspection: 03/07/2019
Serial Number: 80-003224
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.081

Standard Deviations	0.0000	0.0003	0.0005	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

PJM

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAVID E REYES-RIVERA
Signature and Printed Name

03/07/2019
Date

3/15/19
DR

FDLE/ATP Form 41 –Revised August 2005

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003224	Miami Beach Police Department	03/07/2019	DCR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.09C	SN: SD3968 Temp: 34.06C	SN: SD3969 Temp:34.11C	Lot AG805702
0.047 to 0.053 ☑	0.077 to 0.083 ☑	0.194 to 0.206 ☑	0.077 to 0.083 ☑

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.050	10:26
Air Blank	0.000	10:26
Control Test	0.049	10:27
Air Blank	0.000	10:28
Control Test	0.050	10:29
Air Blank	0.000	10:29

Control Test Stats

Average	0.0497
Std Dev	0.0006
Rel Std Dev(%)	1.1625

DCR
Operator's Signature

3/15/19 DD

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.080	10:31
Air Blank	0.000	10:31
Control Test	0.081	10:32
Air Blank	0.000	10:33
Control Test	0.081	10:33
Air Blank	0.000	10:34

Control Test Stats

Average	0.0807
Std Dev	0.0006
Rel Std Dev(%)	0.7157

DCR
Operator's Signature

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.199	10:35
Air Blank	0.000	10:36
Control Test	0.199	10:37
Air Blank	0.000	10:37
Control Test	0.199	10:38
Air Blank	0.000	10:39

Control Test Stats

Average	0.1990
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DCR
Operator's Signature

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003224
03/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:39
Control Test	0.079	10:40
Air Blank	0.000	10:40
Control Test	0.079	10:41
Air Blank	0.000	10:41
Control Test	0.079	10:41
Air Blank	0.000	10:42

Control Test Stats

Average	0.0790
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DCR
Operator's Signature

Copy

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003224 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-003224
Owning Agency:	MIAMI BEACH PD
Calibration Date:	03/07/2019
Calibration Time:	12:46

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/07/2019
Date

DAVID E REYES-RIVERA,
Department Inspector

3/5/19 DP

Form