



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP MIAMI  
Time of Inspection: 14:46

Date of Inspection: 01/07/2019

Serial Number: 80-003348  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.200	0.079
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.082	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.082	0.200	0.078
0.000	0.050	0.082	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.051	0.082	0.200	0.079
0.000	0.050	0.082	0.200	0.079

Standard Deviations	0.0005	0.0005	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

*pgm*

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David Reyes Rivera* Signature and Printed Name DAVID E REYES-RIVERA

01/07/2019  
Date

*1/15/19  
JD*



TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003348	Florida Highway Patrol Miami	01/07/2019	Deer

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.07c <b>0.047 to 0.053</b> <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.07c <b>0.077 to 0.083</b> <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.07c <b>0.194 to 0.206</b> <input checked="" type="checkbox"/>	Lot AG805701 <b>0.077 to 0.083</b> <input checked="" type="checkbox"/>																																																																																																																																																
FHP MIAMI Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-003348 01/07/2019 Software: 8100.27	FHP MIAMI Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-003348 01/07/2019 Software: 8100.27	FHP MIAMI Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-003348 01/07/2019 Software: 8100.27	FHP MIAMI Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-003348 01/07/2019 Software: 8100.27																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:09	Control Test	0.050	12:10	Air Blank	0.000	12:10	Control Test	0.051	12:11	Air Blank	0.000	12:11	Control Test	0.050	12:12	Air Blank	0.000	12:13	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.082</td><td>12:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test</td><td>0.082</td><td>12:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0817</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:14	Control Test	0.081	12:14	Air Blank	0.000	12:15	Control Test	0.082	12:16	Air Blank	0.000	12:16	Control Test	0.082	12:17	Air Blank	0.000	12:17	Control Test Stats			Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7070		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr> <tr><td>Control Test</td><td>0.201</td><td>12:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2003</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2882</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:19	Control Test	0.200	12:20	Air Blank	0.000	12:21	Control Test	0.200	12:21	Air Blank	0.000	12:22	Control Test	0.201	12:22	Air Blank	0.000	12:23	Control Test Stats			Average	0.2003		Std Dev	0.0006		Rel Std Dev(%)	0.2882		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:25	Control Test	0.080	12:25	Air Blank	0.000	12:26	Control Test	0.080	12:26	Air Blank	0.000	12:27	Control Test	0.080	12:27	Air Blank	0.000	12:28	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:09																																																																																																																																																	
Control Test	0.050	12:10																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test	0.051	12:11																																																																																																																																																	
Air Blank	0.000	12:11																																																																																																																																																	
Control Test	0.050	12:12																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.081	12:14																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.082	12:16																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test	0.082	12:17																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0817																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7070																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:19																																																																																																																																																	
Control Test	0.200	12:20																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.200	12:21																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.201	12:22																																																																																																																																																	
Air Blank	0.000	12:23																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2882																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.080	12:25																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test	0.080	12:26																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test	0.080	12:27																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Operator's Signature <i>Deer</i> 11/5/19	Operator's Signature <i>Deer</i>	Operator's Signature <i>Deer</i>	Operator's Signature <i>Deer</i>																																																																																																																																																

WBB



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003348, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003348</u>	UNCERTAINTY* ±
Owning Agency:	<u>FHP MIAMI</u>	0.050 g/ 210 L      0.004
Calibration Date:	<u>01/07/2019</u>	0.080 g/ 210 L      0.004
Calibration Time:	<u>14:46</u>	0.200 g/ 210 L      0.007
		0.080 g/ 210 L Dry Gas Control    0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/07/2019      *David Rojas Aguirre*      DAVID E REYES-RIVERA,  
Date      Department Inspector

FDLE/ATP Form 69 July 2018  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*1/15/19*  
*DA*

*WBR*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Flow calibration	80-003348	Florida Highway Patrol Miami	01/07/2019	JELL

**Calibration 1**

FHP MIAMI  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003348  
 01/07/2019  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
 SORT(Diff) ) = 7.211  
 2: Rate (Liters/min) = 15  
 SORT(Diff) ) = 12.082  
 3: Rate (Liters/min) = 30  
 SORT(Diff) ) = 21.141

Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 694  
 Rounded Intercept = -728567  
 Correlation = 0.99838

**Calibration 2**

FHP MIAMI  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003348  
 01/07/2019  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
 SORT(Diff) ) = 6.855  
 2: Rate (Liters/min) = 15  
 SORT(Diff) ) = 11.617  
 3: Rate (Liters/min) = 30  
 SORT(Diff) ) = 21.188

Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 671  
 Rounded Intercept = -605680  
 Correlation = 0.99708

**Done**

*WJD*

*1/15/19  
J*